

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2769

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	989	1000	99	95 - 105	P	08/11/2025	16:44	LB136779
	Iron	3930	4000	98	95 - 105	P	08/11/2025	16:44	LB136779
	Lead	3970	4000	99	95 - 105	P	08/11/2025	16:44	LB136779

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2769

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.3	20.0	107	80 - 120	P	08/11/2025	16:49	LB136779
	Iron	84.3	100	84	80 - 120	P	08/11/2025	16:49	LB136779
	Lead	12.5	12.0	104	80 - 120	P	08/11/2025	16:49	LB136779

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2769

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1240	1250	99	90 - 110	P	08/11/2025	17:19	LB136779
	Iron	4850	5000	97	90 - 110	P	08/11/2025	17:19	LB136779
	Lead	4840	5000	97	90 - 110	P	08/11/2025	17:19	LB136779
CCV02	Copper	1250	1250	100	90 - 110	P	08/11/2025	17:44	LB136779
	Iron	4960	5000	99	90 - 110	P	08/11/2025	17:44	LB136779
	Lead	4900	5000	98	90 - 110	P	08/11/2025	17:44	LB136779
CCV03	Copper	1240	1250	99	90 - 110	P	08/11/2025	18:26	LB136779
	Iron	4910	5000	98	90 - 110	P	08/11/2025	18:26	LB136779
	Lead	4850	5000	97	90 - 110	P	08/11/2025	18:26	LB136779
CCV04	Copper	1230	1250	98	90 - 110	P	08/11/2025	18:59	LB136779
	Iron	4970	5000	99	90 - 110	P	08/11/2025	18:59	LB136779
	Lead	4810	5000	96	90 - 110	P	08/11/2025	18:59	LB136779

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2769

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	999	1000	100	95 - 105	P	08/12/2025	11:34	LB136789
	Iron	3930	4000	98	95 - 105	P	08/12/2025	11:34	LB136789
	Lead	4050	4000	101	95 - 105	P	08/12/2025	11:34	LB136789

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2769

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.2	20.0	106	80 - 120	P	08/12/2025	11:43	LB136789
	Iron	102	100	102	80 - 120	P	08/12/2025	11:43	LB136789
	Lead	10.9	12.0	91	80 - 120	P	08/12/2025	11:43	LB136789

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2769

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1220	1250	98	90 - 110	P	08/12/2025	12:50	LB136789
	Iron	5030	5000	101	90 - 110	P	08/12/2025	12:50	LB136789
	Lead	4880	5000	98	90 - 110	P	08/12/2025	12:50	LB136789
CCV02	Copper	1270	1250	102	90 - 110	P	08/12/2025	14:27	LB136789
	Iron	5300	5000	106	90 - 110	P	08/12/2025	14:27	LB136789
	Lead	4880	5000	98	90 - 110	P	08/12/2025	14:27	LB136789
CCV03	Copper	1260	1250	101	90 - 110	P	08/12/2025	16:07	LB136789
	Iron	5310	5000	106	90 - 110	P	08/12/2025	16:07	LB136789
	Lead	4820	5000	96	90 - 110	P	08/12/2025	16:07	LB136789
CCV04	Copper	1280	1250	103	90 - 110	P	08/12/2025	16:30	LB136789
	Iron	5400	5000	108	90 - 110	P	08/12/2025	16:30	LB136789
	Lead	4870	5000	97	90 - 110	P	08/12/2025	16:30	LB136789