



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Environmental, Inc

SDG No.: Q2769

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Copper	20.7	20.0	103	65 - 135	P	08/11/2025	16:58	LB136779
	Iron	83.7	100	84	65 - 135	P	08/11/2025	16:58	LB136779
	Lead	11.4	12.0	95	65 - 135	P	08/11/2025	16:58	LB136779
CRI01	Copper	20.3	20.0	102	65 - 135	P	08/12/2025	11:56	LB136789
	Iron	98.1	100	98	65 - 135	P	08/12/2025	11:56	LB136789
	Lead	11.1	12.0	93	65 - 135	P	08/12/2025	11:56	LB136789