



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q2769

Contract: TULL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	4.60	+/-10	U	20.0	P	08/11/2025	16:54	LB136779
	Iron	23.4	+/-50	U	100	P	08/11/2025	16:54	LB136779
	Lead	2.30	+/-6	U	12.0	P	08/11/2025	16:54	LB136779

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q2769

Contract: TULL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	4.60	+/-10	U	20.0	P	08/11/2025	17:23	LB136779
	Iron	23.4	+/-50	U	100	P	08/11/2025	17:23	LB136779
	Lead	2.30	+/-6	U	12.0	P	08/11/2025	17:23	LB136779
CCB02	Copper	4.60	+/-10	U	20.0	P	08/11/2025	17:48	LB136779
	Iron	23.4	+/-50	U	100	P	08/11/2025	17:48	LB136779
	Lead	2.30	+/-6	U	12.0	P	08/11/2025	17:48	LB136779
CCB03	Copper	4.60	+/-10	U	20.0	P	08/11/2025	18:30	LB136779
	Iron	23.4	+/-50	U	100	P	08/11/2025	18:30	LB136779
	Lead	2.30	+/-6	U	12.0	P	08/11/2025	18:30	LB136779
CCB04	Copper	4.60	+/-10	U	20.0	P	08/11/2025	19:03	LB136779
	Iron	23.4	+/-50	U	100	P	08/11/2025	19:03	LB136779
	Lead	2.30	+/-6	U	12.0	P	08/11/2025	19:03	LB136779

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q2769

Contract: TULL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	4.60	+/-10	U	20.0	P	08/12/2025	11:48	LB136789
	Iron	23.4	+/-50	U	100	P	08/12/2025	11:48	LB136789
	Lead	2.30	+/-6	U	12.0	P	08/12/2025	11:48	LB136789

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q2769

Contract: TULL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	4.60	+/-10	U	20.0	P	08/12/2025	12:54	LB136789
	Iron	23.4	+/-50	U	100	P	08/12/2025	12:54	LB136789
	Lead	2.30	+/-6	U	12.0	P	08/12/2025	12:54	LB136789
CCB02	Copper	4.60	+/-10	U	20.0	P	08/12/2025	14:31	LB136789
	Iron	23.4	+/-50	U	100	P	08/12/2025	14:31	LB136789
	Lead	2.30	+/-6	U	12.0	P	08/12/2025	14:31	LB136789
CCB03	Copper	4.60	+/-10	U	20.0	P	08/12/2025	16:11	LB136789
	Iron	23.4	+/-50	U	100	P	08/12/2025	16:11	LB136789
	Lead	2.30	+/-6	U	12.0	P	08/12/2025	16:11	LB136789
CCB04	Copper	4.60	+/-10	U	20.0	P	08/12/2025	16:34	LB136789
	Iron	23.4	+/-50	U	100	P	08/12/2025	16:34	LB136789
	Lead	2.30	+/-6	U	12.0	P	08/12/2025	16:34	LB136789