

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q2769</u>
contract: <u>TULL01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2811-02</u>	client id: <u>EFF-WASTE WATERMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2811-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	223		84.6		150	92		P
Iron	ug/L	75 - 125	1820		201		1500	108		P
Lead	ug/L	75 - 125	430		1.76	J	500	86		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q2769</u>
contract: <u>TULL01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2811-02</u>	client id: <u>EFF-WASTE WATERMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2811-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	243		84.6		150	106		P
Iron	ug/L	75 - 125	1970		201		1500	118		P
Lead	ug/L	75 - 125	465		1.76	J	500	93		P