

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TULL01
 Lab Code: ACE SDG NO.: Q2769
 Lab File ID: VN087468.D Date Analyzed: 08/06/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:25
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	66799	7.80	375782	9.08	344046	11.85
UPPER LIMIT	133598	8.301	751564	9.583	688092	12.347
LOWER LIMIT	33399.5	7.301	187891	8.583	172023	11.347
EPA SAMPLE NO.						
001-WILLETS-PT-BLVD (AUG)	56870	7.80	314501	9.08	279409	11.85
001-WILLETS-PT-BLVD (AUG)MS	43168	7.80	229415	9.08	212461	11.85
001-WILLETS-PT-BLVD (AUG)MSD	62979	7.80	347287	9.08	312549	11.85
002-35TH-AVE (AUG)	76097	7.80	421418	9.08	373245	11.85
VN0806WBL01	60753	7.80	332450	9.08	296861	11.84
VN0806WBS02	63036	7.79	350968	9.08	308794	11.85

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.