

Hit Summary Sheet
SW-846

SDG No.: Q2771
Client: Tully Environmental, Inc

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	001-WILLETS-PT-BLVD(JUNE)							
Q2771-01	001-WILLETS-PT-BLVD Water	Toluene		25.7		0.46	5.00	ug/L
		Total Voc :		25.7				
		Total Concentration:		25.7				
Client ID:	002-35TH-AVE(JUNE)							
Q2771-04	002-35TH-AVE(JUNE) Water	Toluene		20.7		0.46	5.00	ug/L
		Total Voc :		20.7				
		Total Concentration:		20.7				



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	Tully Environmental, Inc		Date Collected:	08/04/25	
Project:	Transfer Station-SPDES		Date Received:	08/05/25	
Client Sample ID:	001-WILLETS-PT-BLVD(JUNE)		SDG No.:	Q2771	
Lab Sample ID:	Q2771-01		Matrix:	Water	
Analytical Method:	E624.1		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-BTEX	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087463.D	1	08/05/25 18:14	VN080525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
71-43-2	Benzene	0.45	U	0.45	5.00	ug/L
108-88-3	Toluene	25.7		0.46	5.00	ug/L
100-41-4	Ethyl Benzene	0.56	U	0.56	5.00	ug/L
179601-23-1	m/p-Xylenes	1.30	U	1.30	10.0	ug/L
95-47-6	o-Xylene	0.67	U	0.67	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	31.5		91 - 110	105%	SPK: 30
2037-26-5	Toluene-d8	30.1		91 - 112	100%	SPK: 30
460-00-4	4-Bromofluorobenzene	28.1		63 - 112	94%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	54800	7.794			
540-36-3	1,4-Difluorobenzene	307000	9.083			
3114-55-4	Chlorobenzene-d5	270000	11.847			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tully Environmental, Inc		Date Collected:	08/04/25	
Project:	Transfer Station-SPDES		Date Received:	08/05/25	
Client Sample ID:	002-35TH-AVE(JUNE)		SDG No.:	Q2771	
Lab Sample ID:	Q2771-04		Matrix:	Water	
Analytical Method:	E624.1		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-BTEX	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087462.D	1	08/05/25 17:53	VN080525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
71-43-2	Benzene	0.45	U	0.45	5.00	ug/L
108-88-3	Toluene	20.7		0.46	5.00	ug/L
100-41-4	Ethyl Benzene	0.56	U	0.56	5.00	ug/L
179601-23-1	m/p-Xylenes	1.30	U	1.30	10.0	ug/L
95-47-6	o-Xylene	0.67	U	0.67	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	31.9		91 - 110	106%	SPK: 30
2037-26-5	Toluene-d8	30.2		91 - 112	101%	SPK: 30
460-00-4	4-Bromofluorobenzene	28.1		63 - 112	94%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	63500	7.8			
540-36-3	1,4-Difluorobenzene	369000	9.083			
3114-55-4	Chlorobenzene-d5	325000	11.847			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q2771	OrderDate:	8/5/2025 11:44:00 AM
Client:	Tully Environmental, Inc	Project:	Transfer Station-SPDES
Contact:	Dean Devoe	Location:	D31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2771-01	001-WILLETS-PT-BLV D(JUNE)	Water			08/04/25			08/05/25
			VOC-BTEX	624.1			08/05/25	
Q2771-04	002-35TH-AVE(JUNE)	Water			08/04/25			08/05/25
			VOC-BTEX	624.1			08/05/25	



SAMPLE DATA

A

B

C

Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	08/04/25 11:30
Project:	Transfer Station-SPDES	Date Received:	08/05/25
Client Sample ID:	001-WILLETS-PT-BLVD(JUNE)	SDG No.:	Q2771
Lab Sample ID:	Q2771-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
TSS	9.10		1	1.00	4.00	mg/L		08/06/25 10:30	SM 2540 D-20

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	08/04/25 11:30
Project:	Transfer Station-SPDES	Date Received:	08/05/25
Client Sample ID:	002-35TH-AVE(JUNE)	SDG No.:	Q2771
Lab Sample ID:	Q2771-04	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
TSS	19.2		1	1.00	4.00	mg/L		08/06/25 10:30	SM 2540 D-20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements
 H = Sample Analysis Out Of Hold Time

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

LAB CHRONICLE

OrderID:	Q2771	OrderDate:	8/5/2025 11:44:00 AM
Client:	Tully Environmental, Inc	Project:	Transfer Station-SPDES
Contact:	Dean Devoe	Location:	D31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2771-01	001-WILLETS-PT-BLV D(JUNE)	WATER	TSS	SM2540 D	08/04/25 11:30		08/06/25 10:30	08/05/25
Q2771-04	002-35TH-AVE(JUNE)	WATER	TSS	SM2540 D	08/04/25 11:30		08/06/25 10:30	08/05/25