

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q2771

Project: Transfer Station-SPDES

Run No.: LB136718

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136718BS							
TSS	mg/L	550	531		96	1	90-110	08/06/2025