

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TULL01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2771</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>08/05/2025 13:58</u>
Lab File ID:	<u>VN087453.D</u>	Init. Calib. Date(s):	<u>08/05/2025 08/05/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>13:37 15:01</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.545	1.540		-0.32	
Toluene	1.852	1.905		2.86	
Ethyl Benzene	2.064	2.115		2.47	
m/p-Xylenes	0.758	0.768		1.32	
o-Xylene	0.745	0.754		1.21	
1,2-Dichloroethane-d4	2.767	2.838		2.57	
Toluene-d8	1.380	1.382		0.14	
4-Bromofluorobenzene	0.516	0.520		0.77	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.