

Cover Page

Order ID : Q2771

Project ID : Transfer Station-SPDES

Client : Tully Environmental, Inc

Lab Sample Number

Q2771-01
Q2771-02
Q2771-03
Q2771-04

Client Sample Number

001-WILLETS-PT-BLVD(JUNE)
Q2771-01MS
Q2771-01MSD
002-35TH-AVE(JUNE)

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 8/7/2025

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following "Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2771

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: KETAN PATEL

Date: 08/07/2025

LAB CHRONICLE

OrderID:	Q2771	OrderDate:	8/5/2025 11:44:00 AM
Client:	Tully Environmental, Inc	Project:	Transfer Station-SPDES
Contact:	Dean Devoe	Location:	D31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2771-01	001-WILLETS-PT-BLV D(JUNE)	Water			08/04/25			08/05/25
			VOC-BTEX	624.1			08/05/25	
Q2771-04	002-35TH-AVE(JUNE)	Water			08/04/25			08/05/25
			VOC-BTEX	624.1			08/05/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2771
Client: Tully Environmental, Inc

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID: Q2771-01	001-WILLETS-PT-BLVD(JUNE) 001-WILLETS-PT-BLVD Water		Toluene	25.7		0.46	5.00	ug/L
			Total Voc :	25.7				
			Total Concentration:	25.7				
Client ID: Q2771-04	002-35TH-AVE(JUNE) 002-35TH-AVE(JUNE) Water		Toluene	20.7		0.46	5.00	ug/L
			Total Voc :	20.7				
			Total Concentration:	20.7				



QC SUMMARY

Surrogate Summary

SDG No.: Q2771

Client: Tully Environmental, Inc

Analytical Method: SW624.1

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2771-01	001-WILLETS-PT-BLVD(JUNE)	1,2-Dichloroethane-d4	30	31.4	105		91	110
		Toluene-d8	30	30.1	100		91	112
		4-Bromofluorobenzene	30	28.1	94		63	112
Q2771-02MS	001-WILLETS-PT-BLVD(JUNE)MS	1,2-Dichloroethane-d4	30	30.9	103		91	110
		Toluene-d8	30	29.9	100		91	112
		4-Bromofluorobenzene	30	29.4	98		63	112
Q2771-03MSD	001-WILLETS-PT-BLVD(JUNE)MSD	1,2-Dichloroethane-d4	30	31.6	106		91	110
		Toluene-d8	30	30.3	101		91	112
		4-Bromofluorobenzene	30	29.7	99		63	112
Q2771-04	002-35TH-AVE(JUNE)	1,2-Dichloroethane-d4	30	31.9	106		91	110
		Toluene-d8	30	30.2	101		91	112
		4-Bromofluorobenzene	30	28.1	94		63	112
VN0805WBL02	VN0805WBL02	1,2-Dichloroethane-d4	30	30.6	102		91	110
		Toluene-d8	30	30.0	100		91	112
		4-Bromofluorobenzene	30	28.3	94		63	112
VN0805WBS02	VN0805WBS02	1,2-Dichloroethane-d4	30	30.0	100		91	110
		Toluene-d8	30	30.7	102		91	112
		4-Bromofluorobenzene	30	30.9	103		63	112

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2771

Analytical Method: SW624.1

Client: Tully Environmental, Inc

Datafile : VN087464.D

Parameter	Spike	Sample	Result	Units	Rec			RPD	Limits			
		Result			Rec	Qual	RPD	Qual	Low	High	RPD	
Lab Sample ID :	Q2771-02MS	Client Sample ID :	001-WILLETS-PT-BLVD(JUNE)MS									
Benzene	20	0	20.5	ug/L	103				37	151		
Toluene	20	25.7	26.1	ug/L	2	*			47	150		
Ethyl Benzene	20	0	20.4	ug/L	102				37	162		
m/p-Xylenes	40	0	40.7	ug/L	102				77	116		
o-Xylene	20	0	20.3	ug/L	102				83	113		

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2771

Analytical Method: SW624.1

Client: Tully Environmental, Inc

Datafile : VN087465.D

Parameter	Spike	Sample	Result	Units	Rec			RPD	Limits			
		Result			Qual	RPD	Qual	Low	High	RPD		
Lab Sample ID :	Q2771-03MSD	Client Sample ID :	001-WILLETS-PT-BLVD(JUNE)MSI									
Benzene	20	0	20.3	ug/L	102				37	151	20	
Toluene	20	25.7	26.0	ug/L	1	*			47	150	20	
Ethyl Benzene	20	0	20.6	ug/L	103				37	162	20	
m/p-Xylenes	40	0	40.9	ug/L	102				77	116	20	
o-Xylene	20	0	20.2	ug/L	101				83	113	20	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2771

Analytical Method: SW624.1

Client: Tully Environmental, Inc

Datafile : VN087459.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VN0805WBS02	Benzene	20	18.5	ug/L	93			65	135	
	Toluene	20	18.6	ug/L	93			70	130	
	Ethyl Benzene	20	18.8	ug/L	94			60	140	
	m/p-Xylenes	40	37.4	ug/L	94			87	111	
	o-Xylene	20	18.6	ug/L	93			87	111	

VOLATILE METHOD BLANK SUMMARY

Client ID

VN0805WBL02

Lab Name: Alliance

Contract: TULL01

Lab Code: ACE

SDG NO.: Q2771

Lab File ID: VN087461.D

Lab Sample ID: VN0805WBL02

Date Analyzed: 08/05/2025

Time Analyzed: 17:32

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VN0805WBS02	VN0805WBS02	VN087459.D	08/05/2025
002-35TH-AVE (JUNE)	Q2771-04	VN087462.D	08/05/2025
001-WILLETS-PT-BLVD (JUNE)	Q2771-01	VN087463.D	08/05/2025
001-WILLETS-PT-BLVD (JUNE) MS	Q2771-02MS	VN087464.D	08/05/2025
001-WILLETS-PT-BLVD (JUNE) MSD	Q2771-03MSD	VN087465.D	08/05/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: TULL01

Lab Code: ACE

SDG NO.: Q2771

Lab File ID: VN087447.D

BFB Injection Date: 08/05/2025

Instrument ID: MSVOA_N

BFB Injection Time: 09:27

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	24.6
75	30.0 - 60.0% of mass 95	58.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	1 (1.5) 1
174	50.0 - 100.0% of mass 95	66.8
175	5.0 - 9.0% of mass 174	4.8 (7.2) 1
176	95.0 - 101.0% of mass 174	65.2 (97.7) 1
177	5.0 - 9.0% of mass 176	4.8 (7.3) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VN087452.D	08/05/2025	13:37
VSTDIC020	VSTDIC020	VN087453.D	08/05/2025	13:58
VSTDIC050	VSTDIC050	VN087454.D	08/05/2025	14:19
VSTDIC100	VSTDIC100	VN087455.D	08/05/2025	14:40
VSTDIC150	VSTDIC150	VN087456.D	08/05/2025	15:01
VN0805WBS02	VN0805WBS02	VN087459.D	08/05/2025	16:17
VN0805WBL02	VN0805WBL02	VN087461.D	08/05/2025	17:32
002-35TH-AVE (JUNE)	Q2771-04	VN087462.D	08/05/2025	17:53
001-WILLETS-PT-BLVD (JUNE)	Q2771-01	VN087463.D	08/05/2025	18:14
001-WILLETS-PT-BLVD (JUNE) MS	Q2771-02MS	VN087464.D	08/05/2025	18:35
001-WILLETS-PT-BLVD (JUNE) MSD	Q2771-03MSD	VN087465.D	08/05/2025	18:56

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TULL01
Lab Code: ACE SDG NO.: Q2771
Lab File ID: VN087453.D Date Analyzed: 08/05/2025
Instrument ID: MSVOA_N Time Analyzed: 13:58
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	64218	7.80	367115	9.08	327914	11.85
UPPER LIMIT	128436	8.3	734230	9.582	655828	12.347
LOWER LIMIT	32109	7.3	183558	8.582	163957	11.347
EPA SAMPLE NO.						
001-WILLETS-PT-BLVD (JUNE)	54767	7.79	306638	9.08	270091	11.85
001-WILLETS-PT-BLVD (JUNE)MS	57858	7.80	327425	9.08	292747	11.85
001-WILLETS-PT-BLVD (JUNE)MSD	63371	7.79	362030	9.08	321655	11.85
002-35TH-AVE (JUNE)	63498	7.80	368676	9.08	324794	11.85
VN0805WBL02	68205	7.80	376206	9.08	334367	11.85
VN0805WBS02	69933	7.80	387635	9.08	345816	11.85

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TULL01
 Lab Code: ACE SDG NO.: Q2771
 Lab File ID: VN087453.D Date Analyzed: 08/05/2025
 Instrument ID: MSVOA_N Time Analyzed: 13:58
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
001-WILLETS-PT-BLVD (JUNE)	0	0.00				
001-WILLETS-PT-BLVD (JUNE) MS	0	0.00				
001-WILLETS-PT-BLVD (JUNE) MSD	0	0.00				
002-35TH-AVE (JUNE)	0	0.00				
VN0805WBL02	0	0.00				
VN0805WBS02	0	0.00				

IS4 =

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client:	Tully Environmental, Inc			Date Collected:	08/04/25	
Project:	Transfer Station-SPDES			Date Received:	08/05/25	
Client Sample ID:	001-WILLETTS-PT-BLVD(JUNE)			SDG No.:	Q2771	
Lab Sample ID:	Q2771-01			Matrix:	Water	
Analytical Method:	E624.1			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOC-BTEX	
GC Column:	RXI-624	ID :	0.25	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087463.D	1	08/05/25 18:14	VN080525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
71-43-2	Benzene	0.45	U	0.45	5.00	ug/L
108-88-3	Toluene	25.7		0.46	5.00	ug/L
100-41-4	Ethyl Benzene	0.56	U	0.56	5.00	ug/L
179601-23-1	m/p-Xylenes	1.30	U	1.30	10.0	ug/L
95-47-6	o-Xylene	0.67	U	0.67	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	31.5		91 - 110	105%	SPK: 30
2037-26-5	Toluene-d8	30.1		91 - 112	100%	SPK: 30
460-00-4	4-Bromofluorobenzene	28.1		63 - 112	94%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	54800	7.794			
540-36-3	1,4-Difluorobenzene	307000	9.083			
3114-55-4	Chlorobenzene-d5	270000	11.847			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087463.D
 Acq On : 05 Aug 2025 18:14
 Operator : JC\MD
 Sample : Q2771-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-WILLETS-PT-BLVD(JUNE)

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 02:44:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.794	128	54767	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.083	114	306638	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	270091	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.559	65	158888	31.454	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery = 104.833%			
60) 4-Bromofluorobenzene	12.829	95	130789	28.138	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery = 93.800%			
63) Toluene-d8	10.547	98	374556	30.139	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery = 100.467%			
Target Compounds						
15) Acetone	4.424	58	18833m	27.270	ug/l	Qvalue
25) Chloroform	7.953	83	266418	35.452	ug/l	98
30) 2-Butanone	7.483	43	24308	6.594	ug/l #	80
41) Bromodichloromethane	9.865	83	21982	3.483	ug/l #	83
46) Methyl methacrylate	9.665	41	13676	2.400	ug/l	96
62) Toluene	10.612	91	429185	25.746	ug/l	100
82) p-Isopropyltoluene	13.706	119	9452	0.635	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
Data File : VN087463.D
Acq On : 05 Aug 2025 18:14
Operator : JC\MD
Sample : Q2771-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
001-WILLETS-PT-BLVD(JUNE)

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/06/2025
Supervised By :Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 02:44:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 06 02:01:21 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 7.794 min Scan# 994
Delta R.T. -0.006 min
Lab File: VN087463.D
Acq: 05 Aug 2025 18:14

Instrument :
MSVOA_N
ClientSampleId :
001-WILLETS-PT-BLVD(JUNE)

Tgt Ion	Ratio	Lower	Upper
128	100		
49	202.0	0.0	494.8
130	128.3	0.0	321.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/06/2025
Supervised By :Mahesh Dadoda 08/06/2025



#15
Acetone
Concen: 27.270 ug/l m
RT: 4.424 min Scan# 420
Delta R.T. 0.006 min
Lab File: VN087463.D
Acq: 05 Aug 2025 18:14

Tgt Ion	Ratio	Lower	Upper
58	100		
43	322.9	268.9	403.3





#25

Chloroform

Concen: 35.452 ug/l

RT: 7.953 min Scan# 1019

Delta R.T. 0.006 min

Lab File: VN087463.D

Acq: 05 Aug 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

001-WILLETS-PT-BLVD(JUNE)

Tgt Ion: 83 Resp: 266418

Ion Ratio Lower Upper

83 100





85 64.3 52.6 78.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/06/2025

Supervised By :Mahesh Dadoda 08/06/2025



#27

1,2-Dichloroethane-d4

Concen: 31.454 ug/l

RT: 8.559 min Scan# 1123

Delta R.T. 0.000 min

Lab File: VN087463.D




Acq: 05 Aug 2025 18:14


Tgt Ion: 65 Resp: 158888

Ion Ratio Lower Upper

65 100

67 51.0 40.8 61.2





#28

1,4-Difluorobenzene

Concen: 30.000 ug/l

RT: 9.083 min Scan# 114

Delta R.T. 0.000 min

Lab File: VN087463.D

Acq: 05 Aug 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

001-WILLETS-PT-BLVD(JUNE)

Tgt Ion:114 Resp: 306638

Ion Ratio Lower Upper

114 100

63 23.7 18.9 28.3





88 16.0 13.1 19.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/06/2025

Supervised By :Mahesh Dadoda 08/06/2025



#30

2-Butanone

Concen: 6.594 ug/l

RT: 7.483 min Scan# 940

Delta R.T. 0.012 min

Lab File: VN087463.D

Acq: 05 Aug 2025 18:14




Tgt Ion: 43 Resp: 24308


Ion Ratio Lower Upper

43 100

57 6.6 6.6 9.8


72 12.0 19.7 29.5#





#41
Bromodichloromethane
Concen: 3.483 ug/l
RT: 9.865 min Scan# 1346
Delta R.T. -0.006 min
Lab File: VN087463.D
Acq: 05 Aug 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
001-WILLETS-PT-BLVD(JUNE)



Tgt Ion: 83 Resp: 2198
Ion Ratio Lower Upper
83 100
85 51.2 52.9 79.3
127 5.7 5.8 8.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/06/2025
Supervised By :Mahesh Dadoda 08/06/2025



#46
Methyl methacrylate
Concen: 2.400 ug/l
RT: 9.665 min Scan# 1311
Delta R.T. 0.000 min
Lab File: VN087463.D
Acq: 05 Aug 2025 18:14


Tgt Ion: 41 Resp: 13676
Ion Ratio Lower Upper
41 100
69 75.3 37.1 111.5
39 55.7 30.7 92.1





#57
Chlorobenzene-d5
Concen: 30.000 ug/l
RT: 11.847 min Scan# 1682
Delta R.T. 0.000 min
Lab File: VN087463.D
Acq: 05 Aug 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
001-WILLETS-PT-BLVD(JUNE)



Tgt Ion: 117 Resp: 270091
Ion Ratio Lower Upper
117 100
82 61.1 49.2 73.8
119 31.9 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/06/2025
Supervised By :Mahesh Dadoda 08/06/2025



#60
4-Bromofluorobenzene
Concen: 28.138 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087463.D
Acq: 05 Aug 2025 18:14


Tgt Ion: 95 Resp: 130789
Ion Ratio Lower Upper
95 100
174 67.0 52.3 78.5
176 63.3 49.8 74.8





#62
Toluene
Concen: 25.746 ug/l
RT: 10.612 min Scan# 1472
Delta R.T. 0.000 min
Lab File: VN087463.D
Acq: 05 Aug 2025 18:14




Instrument :
MSVOA_N
ClientSampleId :
001-WILLETS-PT-BLVD(JUNE)



Tgt Ion: 91 Resp: 429185
Ion Ratio Lower Upper
91 100
92 56.9 45.8 68.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/06/2025
Supervised By :Mahesh Dadoda 08/06/2025



#63
Toluene-d8
Concen: 30.139 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. 0.000 min
Lab File: VN087463.D
Acq: 05 Aug 2025 18:14

Tgt Ion: 98 Resp: 374556
Ion Ratio Lower Upper
98 100
100 63.9 50.5 75.7





#82

p-Isopropyltoluene

Concen: 0.635 ug/l

RT: 13.706 min Scan# 1998

Delta R.T. 0.000 min

Lab File: VN087463.D


Acq: 05 Aug 2025 18:14

Instrument :

MSVOA_N

ClientSampleId :

001-WILLETS-PT-BLVD(JUNE)



Tgt Ion:119 Resp: 945

Ion Ratio Lower Upper

119 100

134 23.3 0.0 51.4



91 13.5 0.0 52.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/06/2025

Supervised By :Mahesh Dadoda 08/06/2025



Report of Analysis

Client:	Tully Environmental, Inc			Date Collected:	08/04/25	
Project:	Transfer Station-SPDES			Date Received:	08/05/25	
Client Sample ID:	002-35TH-AVE(JUNE)			SDG No.:	Q2771	
Lab Sample ID:	Q2771-04			Matrix:	Water	
Analytical Method:	E624.1			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOC-BTEX	
GC Column:	RXI-624	ID :	0.25	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087462.D	1	08/05/25 17:53	VN080525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
71-43-2	Benzene	0.45	U	0.45	5.00	ug/L
108-88-3	Toluene	20.7		0.46	5.00	ug/L
100-41-4	Ethyl Benzene	0.56	U	0.56	5.00	ug/L
179601-23-1	m/p-Xylenes	1.30	U	1.30	10.0	ug/L
95-47-6	o-Xylene	0.67	U	0.67	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	31.9		91 - 110	106%	SPK: 30
2037-26-5	Toluene-d8	30.2		91 - 112	101%	SPK: 30
460-00-4	4-Bromofluorobenzene	28.1		63 - 112	94%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	63500	7.8			
540-36-3	1,4-Difluorobenzene	369000	9.083			
3114-55-4	Chlorobenzene-d5	325000	11.847			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087462.D
 Acq On : 05 Aug 2025 17:53
 Operator : JC\MD
 Sample : Q2771-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 002-35TH-AVE(JUNE)

Quant Time: Aug 06 02:43:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Bromochloromethane	7.800	128	63498	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.083	114	368676	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	324794	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.565	65	187070	31.941	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	= 106.467%	
60) 4-Bromofluorobenzene	12.829	95	157159	28.117	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	= 93.733%	
63) Toluene-d8	10.547	98	451558	30.216	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	= 100.733%	
Target Compounds						Qvalue
15) Acetone	4.418	58	19149	23.915	ug/l	94
25) Chloroform	7.953	83	266729	30.613	ug/l	99
30) 2-Butanone	7.483	43	23874	5.386	ug/l #	93
41) Bromodichloromethane	9.871	83	22253	2.933	ug/l #	92
46) Methyl methacrylate	9.665	41	13479	1.967	ug/l	88
62) Toluene	10.612	91	415924	20.748	ug/l	100
82) p-Isopropyltoluene	13.712	119	10229	0.572	ug/l	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
Data File : VN087462.D
Acq On : 05 Aug 2025 17:53
Operator : JC\MD
Sample : Q2771-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
002-35TH-AVE(JUNE)

Quant Time: Aug 06 02:43:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 06 02:01:21 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 7.800 min Scan# 994
Delta R.T. 0.000 min
Lab File: VN087462.D
Acq: 05 Aug 2025 17:53




Instrument :
MSVOA_N
ClientSampleId :
002-35TH-AVE(JUNE)


Tgt Ion	Ratio	Lower	Upper
128	100		
49	206.1	0.0	494.8
130	131.7	0.0	321.5



#15
Acetone
Concen: 23.915 ug/l
RT: 4.418 min Scan# 419
Delta R.T. 0.000 min
Lab File: VN087462.D
Acq: 05 Aug 2025 17:53

Tgt Ion	Ratio	Lower	Upper
58	100		
43	349.8	268.9	403.3









#25
Chloroform
Concen: 30.613 ug/l
RT: 7.953 min Scan# 1020
Delta R.T. 0.006 min
Lab File: VN087462.D
Acq: 05 Aug 2025 17:53




Instrument :
MSVOA_N
ClientSampleId :
002-35TH-AVE(JUNE)


Tgt Ion: 83 Resp: 266729
Ion Ratio Lower Upper
83 100
85 66.2 52.6 78.8



#27
1,2-Dichloroethane-d4
Concen: 31.941 ug/l
RT: 8.565 min Scan# 1124
Delta R.T. 0.006 min
Lab File: VN087462.D
Acq: 05 Aug 2025 17:53

Tgt Ion: 65 Resp: 187070
Ion Ratio Lower Upper
65 100
67 50.0 40.8 61.2





#28

1,4-Difluorobenzene

Concen: 30.000 ug/l

RT: 9.083 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN087462.D

Acq: 05 Aug 2025 17:53

Instrument :

MSVOA_N

ClientSampleId :

002-35TH-AVE(JUNE)





Tgt Ion:114 Resp: 368676

Ion Ratio Lower Upper

114 100

63 24.0 18.9 28.3

88 16.7 13.1 19.7



#30

2-Butanone

Concen: 5.386 ug/l

RT: 7.483 min Scan# 940

Delta R.T. 0.012 min

Lab File: VN087462.D

Acq: 05 Aug 2025 17:53




Tgt Ion: 43 Resp: 23874


Ion Ratio Lower Upper

43 100

57 2.8 6.6 9.8#

72 22.4 19.7 29.5









#41
Bromodichloromethane
Concen: 2.933 ug/l
RT: 9.871 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN087462.D
Acq: 05 Aug 2025 17:53




Instrument :
MSVOA_N
ClientSampleId :
002-35TH-AVE(JUNE)


Tgt Ion: 83 Resp: 22253
Ion Ratio Lower Upper
83 100
85 72.1 52.9 79.3
127 10.8 5.8 8.6#



#46
Methyl methacrylate
Concen: 1.967 ug/l
RT: 9.665 min Scan# 1311
Delta R.T. 0.000 min
Lab File: VN087462.D
Acq: 05 Aug 2025 17:53

Tgt Ion: 41 Resp: 13479
Ion Ratio Lower Upper
41 100
69 66.7 37.1 111.5
39 73.4 30.7 92.1









#57
Chlorobenzene-d5
Concen: 30.000 ug/l
RT: 11.847 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN087462.D
Acq: 05 Aug 2025 17:53




Instrument :
MSVOA_N
ClientSampleId :
002-35TH-AVE(JUNE)


Tgt Ion:	117	Resp:	324794
Ion	Ratio	Lower	Upper
117	100		
82	61.2	49.2	73.8
119	31.6	25.7	38.5



#60
4-Bromofluorobenzene
Concen: 28.117 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087462.D
Acq: 05 Aug 2025 17:53

Tgt Ion:	95	Resp:	157159
Ion	Ratio	Lower	Upper
95	100		
174	66.7	52.3	78.5
176	64.9	49.8	74.8









#62
Toluene
Concen: 20.748 ug/l
RT: 10.612 min Scan# 1472
Delta R.T. 0.000 min
Lab File: VN087462.D
Acq: 05 Aug 2025 17:53




Instrument :
MSVOA_N
ClientSampleId :
002-35TH-AVE(JUNE)


Tgt Ion: 91 Resp: 415924
Ion Ratio Lower Upper
91 100
92 57.3 45.8 68.6



#63
Toluene-d8
Concen: 30.216 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. 0.000 min
Lab File: VN087462.D
Acq: 05 Aug 2025 17:53

Tgt Ion: 98 Resp: 451558
Ion Ratio Lower Upper
98 100
100 64.2 50.5 75.7








#82
p-Isopropyltoluene
Concen: 0.572 ug/l
RT: 13.712 min Scan# 1999
Delta R.T. 0.006 min
Lab File: VN087462.D
Acq: 05 Aug 2025 17:53

Instrument :
MSVOA_N
ClientSampleId :
002-35TH-AVE(JUNE)

Tgt Ion	Ratio	Lower	Upper
119	100		
134	26.8	0.0	51.4
91	32.2	0.0	52.0





CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: TULL01
 Lab Code: ACE SDG No.: Q2771
 Instrument ID: MSVOA_N Calibration Date(s): 08/05/2025 08/05/2025
 Heated Purge: (Y/N) N Calibration Time(s): 13:37 15:01
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VN087452.D		RRF020 = VN087453.D		RRF050 = VN087454.D		
		RRF100 = VN087455.D		RRF150 = VN087456.D		RRF =		
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	RRF	% RSD
Benzene	1.608	1.540	1.410	1.580	1.584		1.545	5.1
Toluene	1.975	1.905	1.689	1.854	1.835		1.852	5.7
Ethyl Benzene	2.154	2.115	1.892	2.094	2.064		2.064	4.9
m/p-Xylenes	0.796	0.768	0.702	0.768	0.753		0.758	4.6
o-Xylene	0.791	0.754	0.681	0.756	0.745		0.745	5.4
1,2-Dichloroethane-d4	2.772	2.838	2.776	2.704	2.746		2.767	1.8
Toluene-d8	1.413	1.382	1.386	1.386	1.336		1.380	2
4-Bromofluorobenzene	0.512	0.520	0.522	0.520	0.508		0.516	1.1

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\
 Method File : 624N080525W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Wed Aug 06 02:01:21 2025
 Response Via : Initial Calibration

Calibration Files

5 =VN087452.D 20 =VN087453.D 50 =VN087454.D 100 =VN087455.D 150 =VN087456.D

Compound		5	20	50	100	150	Avg	%RSD

1) I	Bromochloromethane	-----ISTD-----						
2) M	Dichlorodifluoro...	1.779	2.065	1.842	2.012	2.094	1.959	7.15
3) M	Chloromethane	1.994	2.165	1.805	1.983	2.101	2.010	6.82
4) M	Vinyl Chloride	2.092	2.289	1.905	2.161	2.210	2.131	6.82
5) M	Bromomethane	1.027	1.160	1.070	1.220	1.291	1.154	9.32
6) M	Chloroethane	1.413	1.441	1.211	1.381	1.427	1.374	6.83
7) M	Trichlorofluorom...	3.231	3.208	2.715	3.043	3.141	3.067	6.85
8) T	Diethyl Ether	1.339	1.321	1.158	1.256	1.316	1.278	5.81
9)	1,1,2-Trichlorot...	1.716	1.591	1.388	1.528	1.598	1.564	7.66
10) M	1,1-Dichloroethene	1.704	1.730	1.440	1.616	1.725	1.643	7.46
11)	Methyl Iodide	0.778	1.076	1.217	1.615	1.782	1.294	31.42
12)	Methyl Acetate	2.942	3.094	2.763	3.128	3.234	3.032	6.05
13) M	Acrolein	0.460	0.463	0.413	0.467	0.513	0.463	7.62
14) M	Acrylonitrile	1.353	1.358	1.193	1.335	1.382	1.324	5.68
15) M	Acetone	0.421	0.418	0.335	0.356	0.361	0.378	10.28
16) M	Carbon Disulfide	5.413	5.176	4.534	5.095	5.266	5.097	6.60
17)	Allyl chloride	3.346	3.377	3.088	3.403	3.531	3.349	4.83
18) M	Methylene Chloride	2.274	2.168	1.887	2.100	2.148	2.115	6.75
19) M	trans-1,2-Dichlo...	2.081	2.014	1.729	1.927	2.019	1.954	7.02
20) T	Diisopropyl ether	7.680	7.823	6.906	7.782	7.955	7.629	5.46
21) M	1,1-Dichloroethane	3.979	4.147	3.567	3.949	4.076	3.944	5.70
22) M	cis-1,2-Dichloro...	2.477	2.499	2.149	2.403	2.493	2.404	6.14
23) M	tert-Butyl Alcohol	0.606	0.569	0.509	0.558	0.571	0.563	6.23
24) M	Methyl tert-Buty...	7.611	7.656	6.765	7.529	7.874	7.487	5.65
25) M	Chloroform	4.236	4.223	3.738	4.105	4.280	4.116	5.37
26)	Cyclohexane	3.153	3.329	2.842	3.205	3.258	3.157	5.95
27) s	1,2-Dichloroetha...	2.772	2.838	2.776	2.704	2.746	2.767	1.76
28) I	1,4-Difluorobenzene	-----ISTD-----						
29)	1,1-Dichloropropene	0.518	0.490	0.445	0.495	0.498	0.489	5.50
30) M	2-Butanone	0.377	0.365	0.334	0.365	0.363	0.361	4.46
31)	2,2-Dichloropropane	0.672	0.638	0.576	0.636	0.632	0.631	5.48
32) M	1,1,1-Trichloroe...	0.651	0.618	0.571	0.636	0.645	0.624	5.19
33) M	Carbon Tetrachlo...	0.536	0.520	0.474	0.527	0.535	0.518	4.94
34) M	Benzene	1.608	1.540	1.410	1.580	1.584	1.545	5.12
35)	Methacrylonitrile	0.409	0.399	0.373	0.382	0.383	0.389	3.77
36) M	1,2-Dichloroethane	0.661	0.636	0.564	0.626	0.634	0.624	5.77
37) M	Trichloroethene	0.355	0.345	0.311	0.344	0.346	0.340	4.97
38)	Methylcyclohexane	0.603	0.572	0.512	0.576	0.575	0.568	5.94
39) M	1,2-Dichloropropane	0.422	0.401	0.358	0.399	0.400	0.396	5.95
40)	Dibromomethane	0.317	0.300	0.269	0.298	0.302	0.297	5.83
41) M	Bromodichloromet...	0.662	0.606	0.561	0.623	0.634	0.617	6.04
42) M	Vinyl Acetate	1.222	1.212	1.088	1.205	1.212	1.188	4.74
43)	Ethyl Acetate	0.745	0.699	0.626	0.699	0.716	0.697	6.31
44)	Isopropyl Acetate	1.228	1.191	1.082	1.201	1.204	1.181	4.82
45) T	1,4-Dioxane	0.008	0.008	0.007	0.007	0.008	0.008	4.20
46)	Methyl methacrylate	0.576	0.558	0.503	0.565	0.585	0.558	5.76
47)	n-amyl Acetate	0.537	0.566	0.572	0.765	0.823	0.652	20.12
48) M	t-1,3-Dichloropr...	0.676	0.663	0.618	0.700	0.702	0.672	5.10
49) T	cis-1,3-Dichloro...	0.687	0.664	0.629	0.696	0.704	0.676	4.48
50) M	1,1,2-Trichloroe...	0.385	0.375	0.344	0.386	0.388	0.375	4.88
51)	Ethyl methacrylate	0.599	0.652	0.624	0.714	0.716	0.661	8.02
52)	1,3-Dichloropropane	0.699	0.678	0.621	0.688	0.696	0.676	4.70
53) M	Dibromochloromet...	0.427	0.424	0.402	0.455	0.460	0.434	5.49
54) M	1,2-Dibromoethane	0.401	0.400	0.364	0.414	0.413	0.399	5.12
55) M	2-Chloroethyl vi...	0.373	0.349	0.321	0.345	0.353	0.348	5.40
56) M	Bromoform	0.268	0.276	0.268	0.310	0.312	0.287	7.82

Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\
 Method File : 624N080525W.M

		-----ISTD-----							
57) I	Chlorobenzene-d5								
58) M	4-Methyl-2-Penta...	0.813	0.802	0.713	0.780	0.751	0.772		5.27
59) M	2-Hexanone	0.484	0.527	0.504	0.561	0.545	0.524		5.95
60) S	4-Bromofluoroben...	0.512	0.520	0.522	0.520	0.508	0.516		1.14
61) M	Tetrachloroethene	0.320	0.292	0.270	0.290	0.289	0.292		6.10
62) M	Toluene	1.975	1.905	1.689	1.854	1.835	1.852		5.70
63) S	Toluene-d8	1.413	1.382	1.386	1.386	1.336	1.380		2.02
64) M	Chlorobenzene	1.218	1.156	1.030	1.149	1.135	1.138		5.97
65)	1,1,1,2-Tetrachl...	0.431	0.404	0.364	0.405	0.396	0.400		5.99
66) M	Ethyl Benzene	2.154	2.115	1.892	2.094	2.064	2.064		4.92
67) M	m/p-Xylenes	0.796	0.768	0.702	0.768	0.753	0.758		4.55
68) M	o-Xylene	0.791	0.754	0.681	0.756	0.745	0.745		5.38
69) M	Styrene	1.257	1.295	1.182	1.323	1.307	1.273		4.43
70)	Isopropylbenzene	1.938	1.936	1.729	1.947	1.916	1.893		4.88
71) M	1,1,2,2-Tetrachl...	0.692	0.679	0.590	0.654	0.628	0.649		6.28
72)	1,2,3-Trichlorop...	0.631	0.567	0.582	0.629	0.613	0.604		4.77
73)	Bromobenzene	0.468	0.440	0.398	0.451	0.438	0.439		5.88
74)	n-propylbenzene	2.495	2.410	2.165	2.415	2.338	2.365		5.27
75)	2-Chlorotoluene	1.570	1.484	1.328	1.466	1.424	1.455		6.07
76)	1,3,5-Trimethylb...	1.705	1.661	1.495	1.646	1.599	1.621		4.93
77)	t-1,4-Dichloro-2...	0.192	0.216	0.220	0.261	0.256	0.229	12.71	
78)	4-Chlorotoluene	1.549	1.474	1.362	1.513	1.472	1.474		4.77
79)	tert-butylbenzene	1.443	1.398	1.266	1.395	1.332	1.367		5.03
80)	1,2,4-Trimethylb...	1.672	1.639	1.511	1.667	1.611	1.620		4.05
81)	sec-Butylbenzene	2.093	2.047	1.849	2.015	1.944	1.990		4.80
82)	p-Isopropyltoluene	1.761	1.692	1.528	1.670	1.610	1.652		5.32
83) M	1,3-Dichlorobenzene	0.874	0.853	0.777	0.847	0.828	0.836		4.37
84) M	1,4-Dichlorobenzene	0.884	0.870	0.790	0.858	0.835	0.847		4.35
85)	n-Butylbenzene	1.647	1.662	1.513	1.658	1.602	1.616		3.89
86) T	Hexachloroethane	0.327	0.326	0.300	0.334	0.327	0.323		4.07
87) M	1,2-Dichlorobenzene	0.812	0.819	0.750	0.814	0.794	0.798		3.56
88)	1,2-Dibromo-3-Ch...	0.173	0.170	0.149	0.166	0.164	0.164		5.70
89)	1,2,4-Trichlorob...	0.495	0.478	0.435	0.486	0.490	0.477		5.05
90)	Hexachlorobutadiene	0.222	0.189	0.165	0.177	0.174	0.185	11.79	
91) M	Naphthalene	1.807	1.840	1.675	1.885	1.883	1.818		4.73
92)	1,2,3-Trichlorob...	0.495	0.478	0.435	0.486	0.490	0.477		5.05

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087452.D
 Acq On : 05 Aug 2025 13:37
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/06/2025

Supervised By :Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 01:26:38 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M

Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS

QLast Update : Wed Aug 06 01:24:35 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.794	128	66791	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.082	114	367765	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	324158	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.565	65	185120	30.050	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery = 100.167%			
60) 4-Bromofluorobenzene	12.829	95	166050	29.766	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery = 99.233%			
63) Toluene-d8	10.547	98	457922	30.702	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery = 102.333%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.142	85	19809	4.543	ug/l	97
3) Chloromethane	2.383	50	22193	4.960	ug/l	98
4) Vinyl Chloride	2.542	62	23292	4.908	ug/l	93
5) Bromomethane	2.977	94	11433	4.451	ug/l	96
6) Chloroethane	3.136	64	15730	5.141	ug/l	96
7) Trichlorofluoromethane	3.500	101	35964	5.266	ug/l #	81
8) Diethyl Ether	3.965	74	14911	5.241	ug/l	95
9) 1,1,2-Trichlorotrifluo...	4.359	101	19105	5.486	ug/l	63
10) 1,1-Dichloroethene	4.342	96	18969	5.185	ug/l	96
11) Methyl Iodide	4.583	142	8659m	3.021	ug/l	
12) Methyl Acetate	5.024	43	32749	4.851	ug/l #	91
13) Acrolein	4.183	56	25602	24.838	ug/l	92
14) Acrylonitrile	5.706	53	75304	25.546	ug/l	99
15) Acetone	4.424	58	23455	27.848	ug/l	88
16) Carbon Disulfide	4.706	76	60257	5.310	ug/l	100
17) Allyl chloride	5.018	41	37245	4.995	ug/l	91
18) Methylene Chloride	5.265	84	25316	5.377	ug/l	87
19) trans-1,2-Dichloroethene	5.777	96	23167m	5.326	ug/l	
20) Diisopropyl ether	6.659	45	85498	5.036	ug/l #	98
21) 1,1-Dichloroethane	6.553	63	44299	5.045	ug/l	99
22) cis-1,2-Dichloroethene	7.465	96	27569	5.151	ug/l #	88
23) tert-Butyl Alcohol	5.524	59	33736	26.933	ug/l #	100
24) Methyl tert-Butyl Ether	5.783	73	84726	5.083	ug/l #	69
25) Chloroform	7.947	83	47159	5.146	ug/l	98
26) Cyclohexane	8.241	56	35104	4.994	ug/l #	93
29) 1,1-Dichloropropene	8.353	75	31744	5.292	ug/l	92
30) 2-Butanone	7.471	43	115688	26.165	ug/l #	83
31) 2,2-Dichloropropane	7.471	77	41219	5.328	ug/l #	61
32) 1,1,1-Trichloroethane	8.153	97	39911	5.217	ug/l	99
33) Carbon Tetrachloride	8.341	117	32877	5.173	ug/l	96
34) Benzene	8.582	78	98586	5.207	ug/l #	81
35) Methacrylonitrile	7.771	41	25099	5.260	ug/l	93
36) 1,2-Dichloroethane	8.647	62	40495	5.291	ug/l	97
37) Trichloroethene	9.335	130	21742	5.213	ug/l	92
38) Methylcyclohexane	9.577	83	36974	5.313	ug/l	98
39) 1,2-Dichloropropane	9.600	63	25889	5.333	ug/l	91
40) Dibromomethane	9.688	93	19402	5.327	ug/l	99
41) Bromodichloromethane	9.871	83	40582	5.362	ug/l #	95
42) Vinyl Acetate	6.594	43	374548	25.723	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087452.D
 Acq On : 05 Aug 2025 13:37
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 01:26:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 01:24:35 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	7.553	43	45639m	5.423	ug/l	
44) Isopropyl Acetate	8.677	43	75250	5.197	ug/l	98
45) 1,4-Dioxane	9.682	88	9597	103.972	ug/l #	85
46) Methyl methacrylate	9.665	41	35322	5.168	ug/l	98
47) n-amyl Acetate	12.682	43	32939m	4.524	ug/l	
48) t-1,3-Dichloropropene	10.818	75	41412	5.029	ug/l	95
49) cis-1,3-Dichloropropene	10.294	75	42124	5.083	ug/l #	94
50) 1,1,2-Trichloroethane	10.994	97	23578	5.123	ug/l	93
51) Ethyl methacrylate	10.859	69	36705	4.529	ug/l	97
52) 1,3-Dichloropropane	11.141	76	42828	5.166	ug/l	97
53) Dibromochloromethane	11.341	129	26196	4.929	ug/l	95
54) 1,2-Dibromoethane	11.447	107	24555	5.026	ug/l	97
55) 2-Chloroethyl vinyl ether	10.141	63	114298	26.788	ug/l	99
56) Bromoform	12.559	173	16434	4.675	ug/l #	78
58) 4-Methyl-2-Pentanone	10.424	43	219603	26.335	ug/l	99
59) 2-Hexanone	11.182	43	130616	23.053	ug/l	99
61) Tetrachloroethene	11.082	164	17278	5.471	ug/l	92
62) Toluene	10.612	91	106690	5.333	ug/l	100
64) Chlorobenzene	11.871	112	65821	5.354	ug/l	100
65) 1,1,1,2-Tetrachloroethane	11.941	131	23260	5.381	ug/l	98
66) Ethyl Benzene	11.947	91	116379	5.219	ug/l	97
67) m/p-Xylenes	12.053	106	85967	10.503	ug/l	96
68) o-Xylene	12.376	106	42724	5.306	ug/l	95
69) Styrene	12.394	104	67893	4.938	ug/l	97
70) Isopropylbenzene	12.676	105	104723	5.119	ug/l	98
71) 1,1,2,2-Tetrachloroethane	12.918	83	37401	5.336	ug/l	97
72) 1,2,3-Trichloropropane	12.976	75	34102m	4.733	ug/l	
73) Bromobenzene	12.959	156	25284	5.331	ug/l	94
74) n-propylbenzene	13.018	91	134786	5.275	ug/l	99
75) 2-Chlorotoluene	13.106	91	84827	5.397	ug/l	95
76) 1,3,5-Trimethylbenzene	13.153	105	92128	5.259	ug/l	100
77) t-1,4-Dichloro-2-butene	12.718	75	10355	4.189	ug/l	79
78) 4-Chlorotoluene	13.200	91	83697	5.255	ug/l	98
79) tert-butylbenzene	13.417	119	77933	5.277	ug/l	99
80) 1,2,4-Trimethylbenzene	13.459	105	90309	5.159	ug/l	100
81) sec-Butylbenzene	13.594	105	113073	5.259	ug/l	97
82) p-Isopropyltoluene	13.706	119	95127	5.329	ug/l	99
83) 1,3-Dichlorobenzene	13.717	146	47197	5.226	ug/l	96
84) 1,4-Dichlorobenzene	13.788	146	47756	5.216	ug/l	97
85) n-Butylbenzene	14.035	91	89004	5.096	ug/l	99
86) Hexachloroethane	14.312	117	17680	5.072	ug/l	97
87) 1,2-Dichlorobenzene	14.082	146	43857	5.087	ug/l	96
88) 1,2-Dibromo-3-Chloropr...	14.694	75	9336	5.257	ug/l	98
89) 1,2,4-Trichlorobenzene	15.811	180	26729	5.188	ug/l	98
90) Hexachlorobutadiene	15.470	225	11972	5.976	ug/l	80
91) Naphthalene	15.617	128	97646	4.971	ug/l	98
92) 1,2,3-Trichlorobenzene	15.811	180	26729	5.188	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087452.D
 Acq On : 05 Aug 2025 13:37
 Operator : JC\MD
 Sample : VSTDIC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/06/2025

Supervised By :Mahesh Dadoda 08/06/2025


Quant Time: Aug 06 01:26:38 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M

Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS

QLast Update : Wed Aug 06 01:24:35 2025

Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087453.D
 Acq On : 05 Aug 2025 13:58
 Operator : JC\MD
 Sample : VSTDICCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 01:27:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 01:24:35 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.800	128	64218	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.082	114	367115	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	327914	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.559	65	182238	30.767	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery = 102.567%			
60) 4-Bromofluorobenzene	12.829	95	170429	30.201	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery = 100.667%			
63) Toluene-d8	10.547	98	453121	30.032	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery = 100.100%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.142	85	88427	21.090	ug/l	100
3) Chloromethane	2.383	50	92684	21.546	ug/l	100
4) Vinyl Chloride	2.542	62	97984	21.476	ug/l	100
5) Bromomethane	2.983	94	49674	20.114	ug/l	100
6) Chloroethane	3.142	64	61674	20.963	ug/l	100
7) Trichlorofluoromethane	3.512	101	137336	20.915	ug/l	100
8) Diethyl Ether	3.965	74	56555	20.674	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.371	101	68121	20.344	ug/l	100
10) 1,1-Dichloroethene	4.336	96	74078	21.062	ug/l	100
11) Methyl Iodide	4.589	142	46084m	16.723	ug/l	
12) Methyl Acetate	5.018	43	132440	20.405	ug/l	100
13) Acrolein	4.171	56	99154	100.048	ug/l	99
14) Acrylonitrile	5.712	53	290627	102.544	ug/l	100
15) Acetone	4.418	58	89398	110.396	ug/l	100
16) Carbon Disulfide	4.700	76	221586	20.310	ug/l	100
17) Allyl chloride	5.012	41	144593	20.170	ug/l	100
18) Methylene Chloride	5.271	84	92808m	20.501	ug/l	
19) trans-1,2-Dichloroethene	5.771	96	86216	20.613	ug/l	100
20) Diisopropyl ether	6.665	45	334922	20.517	ug/l	100
21) 1,1-Dichloroethane	6.553	63	177532	21.030	ug/l	100
22) cis-1,2-Dichloroethene	7.471	96	106998	20.791	ug/l	100
23) tert-Butyl Alcohol	5.518	59	121824	101.156	ug/l #	100
24) Methyl tert-Butyl Ether	5.789	73	327775	20.452	ug/l	100
25) Chloroform	7.947	83	180797	20.518	ug/l	100
26) Cyclohexane	8.241	56	142519	21.086	ug/l	100
29) 1,1-Dichloropropene	8.353	75	120007	20.042	ug/l	100
30) 2-Butanone	7.471	43	446235	101.104	ug/l	100
31) 2,2-Dichloropropane	7.471	77	156255	20.233	ug/l	100
32) 1,1,1-Trichloroethane	8.147	97	151297	19.812	ug/l	100
33) Carbon Tetrachloride	8.341	117	127325	20.070	ug/l	100
34) Benzene	8.588	78	377003	19.946	ug/l	100
35) Methacrylonitrile	7.765	41	97561	20.484	ug/l	100
36) 1,2-Dichloroethane	8.653	62	155656	20.373	ug/l	100
37) Trichloroethene	9.335	130	84369	20.264	ug/l	100
38) Methylcyclohexane	9.582	83	139951	20.147	ug/l	100
39) 1,2-Dichloropropane	9.600	63	98038	20.230	ug/l	100
40) Dibromomethane	9.688	93	73356	20.175	ug/l	100
41) Bromodichloromethane	9.871	83	148435	19.646	ug/l	100
42) Vinyl Acetate	6.588	43	1483194	102.044	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087453.D
 Acq On : 05 Aug 2025 13:58
 Operator : JC\MD
 Sample : VSTDICCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 01:27:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 01:24:35 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	7.553	43	170967m	20.349	ug/l	
44) Isopropyl Acetate	8.677	43	291509	20.169	ug/l	100
45) 1,4-Dioxane	9.677	88	37910	411.437	ug/l	100
46) Methyl methacrylate	9.665	41	136612	20.024	ug/l	100
47) n-amyl Acetate	12.529	43	138459m	19.052	ug/l	
48) t-1,3-Dichloropropene	10.818	75	162191	19.732	ug/l	100
49) cis-1,3-Dichloropropene	10.294	75	162479	19.641	ug/l	100
50) 1,1,2-Trichloroethane	10.994	97	91845	19.993	ug/l	100
51) Ethyl methacrylate	10.853	69	159690	19.738	ug/l	100
52) 1,3-Dichloropropane	11.141	76	165895	20.044	ug/l	100
53) Dibromochloromethane	11.341	129	103790	19.563	ug/l	100
54) 1,2-Dibromoethane	11.447	107	97894	20.073	ug/l	100
55) 2-Chloroethyl vinyl ether	10.141	63	426876	100.225	ug/l	100
56) Bromoform	12.559	173	67550	19.250	ug/l	100
58) 4-Methyl-2-Pentanone	10.424	43	876419	103.897	ug/l	100
59) 2-Hexanone	11.182	43	576264	100.543	ug/l	100
61) Tetrachloroethene	11.082	164	63867	19.992	ug/l	100
62) Toluene	10.612	91	416382	20.573	ug/l	100
64) Chlorobenzene	11.871	112	252716	20.321	ug/l	100
65) 1,1,1,2-Tetrachloroethane	11.941	131	88390	20.216	ug/l	100
66) Ethyl Benzene	11.941	91	462372	20.497	ug/l	100
67) m/p-Xylenes	12.047	106	335952	40.573	ug/l	100
68) o-Xylene	12.376	106	164778	20.228	ug/l	100
69) Styrene	12.394	104	283102	20.353	ug/l	100
70) Isopropylbenzene	12.670	105	423292	20.453	ug/l	100
71) 1,1,2,2-Tetrachloroethane	12.918	83	148347	20.920	ug/l	100
72) 1,2,3-Trichloropropane	12.970	75	123856m	16.992	ug/l	
73) Bromobenzene	12.959	156	96161	20.043	ug/l	100
74) n-propylbenzene	13.012	91	526811	20.382	ug/l	100
75) 2-Chlorotoluene	13.106	91	324478	20.409	ug/l	100
76) 1,3,5-Trimethylbenzene	13.153	105	363042	20.488	ug/l	100
77) t-1,4-Dichloro-2-butene	12.718	75	47119	18.843	ug/l	100
78) 4-Chlorotoluene	13.200	91	322316	20.005	ug/l	100
79) tert-butylbenzene	13.417	119	305702	20.462	ug/l	100
80) 1,2,4-Trimethylbenzene	13.459	105	358349	20.238	ug/l	100
81) sec-Butylbenzene	13.594	105	447541	20.577	ug/l	100
82) p-Isopropyltoluene	13.706	119	369955	20.486	ug/l	100
83) 1,3-Dichlorobenzene	13.712	146	186482	20.412	ug/l	100
84) 1,4-Dichlorobenzene	13.788	146	190232	20.538	ug/l	100
85) n-Butylbenzene	14.035	91	363428	20.569	ug/l	100
86) Hexachloroethane	14.312	117	71198	20.189	ug/l	100
87) 1,2-Dichlorobenzene	14.082	146	179117	20.539	ug/l	100
88) 1,2-Dibromo-3-Chloropr...	14.694	75	37135	20.672	ug/l	100
89) 1,2,4-Trichlorobenzene	15.811	180	104426	20.037	ug/l	100
90) Hexachlorobutadiene	15.476	225	41238	20.349	ug/l	100
91) Naphthalene	15.611	128	402240	20.241	ug/l	100
92) 1,2,3-Trichlorobenzene	15.811	180	104426	20.037	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020

Manual Integrations APPROVED

Reviewed By :John Carlone 08/06/2025
Supervised By :Mahesh Dadoda 08/06/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087454.D
 Acq On : 05 Aug 2025 14:19
 Operator : JC\MD
 Sample : VSTDICC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC050

Quant Time: Aug 06 01:28:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 01:24:35 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :John Carlone 08/06/2025
 Supervised By :Mahesh Dadoda 08/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.800	128	76249	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.083	114	418905	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	385657	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.559	65	211654	30.095	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery = 100.333%			
60) 4-Bromofluorobenzene	12.829	95	201264	30.325	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery = 101.067%			
63) Toluene-d8	10.547	98	534559	30.124	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery = 100.400%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.142	85	234078	47.020	ug/l	97
3) Chloromethane	2.383	50	229410	44.915	ug/l	98
4) Vinyl Chloride	2.542	62	242082	44.687	ug/l	93
5) Bromomethane	2.977	94	136035	46.393	ug/l	84
6) Chloroethane	3.136	64	153939	44.067	ug/l	98
7) Trichlorofluoromethane	3.506	101	344998	44.251	ug/l	96
8) Diethyl Ether	3.959	74	147110	45.292	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.365	101	176341	44.354	ug/l	88
10) 1,1-Dichloroethene	4.336	96	182951	43.809	ug/l	94
11) Methyl Iodide	4.577	142	154678	47.274	ug/l	99
12) Methyl Acetate	5.012	43	351071	45.556	ug/l	98
13) Acrolein	4.177	56	262498	223.074	ug/l	96
14) Acrylonitrile	5.706	53	757874	225.213	ug/l	100
15) Acetone	4.424	58	212947	221.472	ug/l	95
16) Carbon Disulfide	4.700	76	576164	44.477	ug/l	98
17) Allyl chloride	5.012	41	392451	46.106	ug/l	97
18) Methylene Chloride	5.265	84	239778	44.610	ug/l	97
19) trans-1,2-Dichloroethene	5.771	96	219724	44.245	ug/l	98
20) Diisopropyl ether	6.659	45	877578	45.278	ug/l	96
21) 1,1-Dichloroethane	6.553	63	453264	45.221	ug/l	99
22) cis-1,2-Dichloroethene	7.471	96	273116	44.697	ug/l	99
23) tert-Butyl Alcohol	5.518	59	323387	226.154	ug/l #	100
24) Methyl tert-Butyl Ether	5.783	73	859723	45.179	ug/l	99
25) Chloroform	7.947	83	475045	45.405	ug/l	98
26) Cyclohexane	8.241	56	361178	45.006	ug/l #	97
29) 1,1-Dichloropropene	8.353	75	310679	45.472	ug/l	97
30) 2-Butanone	7.471	43	1165660	231.454	ug/l	99
31) 2,2-Dichloropropane	7.471	77	402451	45.671	ug/l	100
32) 1,1,1-Trichloroethane	8.153	97	398348	45.714	ug/l	97
33) Carbon Tetrachloride	8.347	117	330997	45.723	ug/l	98
34) Benzene	8.588	78	984335	45.639	ug/l	99
35) Methacrylonitrile	7.765	41	260104	47.860	ug/l	99
36) 1,2-Dichloroethane	8.653	62	393795	45.169	ug/l	98
37) Trichloroethene	9.335	130	217082	45.693	ug/l	95
38) Methylcyclohexane	9.582	83	357253	45.072	ug/l	99
39) 1,2-Dichloropropane	9.600	63	249627	45.142	ug/l	97
40) Dibromomethane	9.688	93	187857	45.278	ug/l	97
41) Bromodichloromethane	9.871	83	391953	45.463	ug/l	98
42) Vinyl Acetate	6.589	43	3796699	228.920	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087454.D
 Acq On : 05 Aug 2025 14:19
 Operator : JC\MD
 Sample : VSTDICC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 01:28:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 01:24:35 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	7.547	43	436753	45.558	ug/l	96
44) Isopropyl Acetate	8.677	43	755567	45.813	ug/l	99
45) 1,4-Dioxane	9.682	88	98035	932.433	ug/l	97
46) Methyl methacrylate	9.665	41	351323	45.128	ug/l	98
47) n-amyl Acetate	12.494	43	399157m	48.133	ug/l	
48) t-1,3-Dichloropropene	10.818	75	431587	46.014	ug/l	98
49) cis-1,3-Dichloropropene	10.294	75	439121	46.521	ug/l	96
50) 1,1,2-Trichloroethane	10.994	97	240010	45.786	ug/l	98
51) Ethyl methacrylate	10.853	69	435415	47.164	ug/l	97
52) 1,3-Dichloropropane	11.141	76	433869	45.941	ug/l	98
53) Dibromochloromethane	11.335	129	280542	46.341	ug/l	100
54) 1,2-Dibromoethane	11.447	107	254243	45.688	ug/l	98
55) 2-Chloroethyl vinyl ether	10.141	63	1118997	230.244	ug/l	100
56) Bromoform	12.559	173	186820	46.656	ug/l #	99
58) 4-Methyl-2-Pentanone	10.424	43	2290380	230.864	ug/l	99
59) 2-Hexanone	11.177	43	1620747	240.439	ug/l	100
61) Tetrachloroethene	11.082	164	173500	46.177	ug/l	93
62) Toluene	10.612	91	1085878	45.620	ug/l	99
64) Chlorobenzene	11.871	112	662306	45.283	ug/l	98
65) 1,1,1,2-Tetrachloroethane	11.935	131	233942	45.493	ug/l	98
66) Ethyl Benzene	11.941	91	1215937	45.832	ug/l	98
67) m/p-Xylenes	12.047	106	902756	92.702	ug/l	97
68) o-Xylene	12.376	106	437515	45.667	ug/l	99
69) Styrene	12.388	104	759493	46.428	ug/l	99
70) Isopropylbenzene	12.676	105	1111500	45.666	ug/l	99
71) 1,1,2,2-Tetrachloroethane	12.918	83	379527	45.508	ug/l	99
72) 1,2,3-Trichloropropane	12.970	75	374092m	43.638	ug/l	
73) Bromobenzene	12.959	156	255773	45.330	ug/l	99
74) n-propylbenzene	13.012	91	1391658	45.781	ug/l	99
75) 2-Chlorotoluene	13.106	91	853763	45.659	ug/l	99
76) 1,3,5-Trimethylbenzene	13.153	105	961117	46.118	ug/l	100
77) t-1,4-Dichloro-2-butene	12.718	75	141496	48.111	ug/l	93
78) 4-Chlorotoluene	13.200	91	875292	46.192	ug/l	99
79) tert-butylbenzene	13.418	119	813723	46.311	ug/l	99
80) 1,2,4-Trimethylbenzene	13.459	105	971243	46.640	ug/l	100
81) sec-Butylbenzene	13.594	105	1188558	46.465	ug/l	100
82) p-Isopropyltoluene	13.706	119	982310	46.251	ug/l	100
83) 1,3-Dichlorobenzene	13.712	146	499720	46.509	ug/l	100
84) 1,4-Dichlorobenzene	13.788	146	507658	46.603	ug/l	99
85) n-Butylbenzene	14.035	91	972192	46.785	ug/l	99
86) Hexachloroethane	14.312	117	192686	46.458	ug/l	97
87) 1,2-Dichlorobenzene	14.082	146	482093	47.005	ug/l	99
88) 1,2-Dibromo-3-Chloropr...	14.694	75	95570	45.236	ug/l	95
89) 1,2,4-Trichlorobenzene	15.811	180	279770	45.643	ug/l	99
90) Hexachlorobutadiene	15.476	225	106360	44.624	ug/l	97
91) Naphthalene	15.611	128	1076915	46.078	ug/l	99
92) 1,2,3-Trichlorobenzene	15.811	180	279770	45.643	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
Data File : VN087454.D
Acq On : 05 Aug 2025 14:19
Operator : JC\MD
Sample : VSTDICC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC050

Manual Integrations APPROVED

Reviewed By :John Carlone 08/06/2025
Supervised By :Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 01:28:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 06 01:24:35 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087455.D
 Acq On : 05 Aug 2025 14:40
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 01:29:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 01:24:35 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.794	128	70635	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.082	114	389326	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	359656	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.559	65	191012	29.319	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery =	97.733%		
60) 4-Bromofluorobenzene	12.829	95	186862	30.190	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery =	100.633%		
63) Toluene-d8	10.547	98	498338	30.114	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery =	100.367%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.142	85	473762	102.730	ug/l	97
3) Chloromethane	2.383	50	466952	98.689	ug/l	99
4) Vinyl Chloride	2.536	62	508879	101.402	ug/l	97
5) Bromomethane	2.959	94	287163	105.716	ug/l	87
6) Chloroethane	3.130	64	325044	100.444	ug/l	99
7) Trichlorofluoromethane	3.506	101	716379	99.189	ug/l	99
8) Diethyl Ether	3.959	74	295673	98.265	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.365	101	359828	97.700	ug/l	88
10) 1,1-Dichloroethene	4.330	96	380598	98.381	ug/l	98
11) Methyl Iodide	4.577	142	380217	125.440	ug/l	91
12) Methyl Acetate	5.012	43	736494	103.164	ug/l	97
13) Acrolein	4.171	56	549240	503.847	ug/l	99
14) Acrylonitrile	5.706	53	1571651	504.158	ug/l	100
15) Acetone	4.424	58	418923	470.323	ug/l	93
16) Carbon Disulfide	4.700	76	1199576	99.962	ug/l	100
17) Allyl chloride	5.012	41	801152	101.602	ug/l	95
18) Methylene Chloride	5.265	84	494371	99.285	ug/l	94
19) trans-1,2-Dichloroethene	5.771	96	453658	98.611	ug/l	93
20) Diisopropyl ether	6.659	45	1832338	102.051	ug/l	98
21) 1,1-Dichloroethane	6.553	63	929854	100.142	ug/l	99
22) cis-1,2-Dichloroethene	7.477	96	565755	99.948	ug/l	99
23) tert-Butyl Alcohol	5.530	59	656761	495.796	ug/l #	100
24) Methyl tert-Butyl Ether	5.788	73	1772627	100.557	ug/l	99
25) Chloroform	7.947	83	966531	99.723	ug/l	100
26) Cyclohexane	8.241	56	754564	101.498	ug/l #	99
29) 1,1-Dichloropropene	8.353	75	642951	101.253	ug/l	98
30) 2-Butanone	7.471	43	2366356	505.563	ug/l	100
31) 2,2-Dichloropropane	7.477	77	825382	100.781	ug/l	100
32) 1,1,1-Trichloroethane	8.153	97	825110	101.882	ug/l	96
33) Carbon Tetrachloride	8.341	117	683314	101.563	ug/l	97
34) Benzene	8.588	78	2050407	102.292	ug/l	97
35) Methacrylonitrile	7.765	41	495672	98.134	ug/l	97
36) 1,2-Dichloroethane	8.653	62	812745	100.306	ug/l	100
37) Trichloroethene	9.335	130	446892	101.212	ug/l	94
38) Methylcyclohexane	9.582	83	747573	101.482	ug/l	98
39) 1,2-Dichloropropane	9.600	63	518393	100.868	ug/l	95
40) Dibromomethane	9.688	93	386489	100.231	ug/l	96
41) Bromodichloromethane	9.871	83	808690	100.928	ug/l	98
42) Vinyl Acetate	6.588	43	7820160	507.335	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087455.D
 Acq On : 05 Aug 2025 14:40
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 01:29:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 01:24:35 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	7.547	43	907628	101.868	ug/l	97
44) Isopropyl Acetate	8.677	43	1558470	101.675	ug/l	99
45) 1,4-Dioxane	9.682	88	194002	1985.385	ug/l	96
46) Methyl methacrylate	9.665	41	732745	101.273	ug/l	92
47) n-amyl Acetate	12.488	43	992295	128.748	ug/l #	60
48) t-1,3-Dichloropropene	10.818	75	908224	104.188	ug/l	100
49) cis-1,3-Dichloropropene	10.294	75	903230	102.958	ug/l	97
50) 1,1,2-Trichloroethane	11.000	97	500440	102.720	ug/l	96
51) Ethyl methacrylate	10.853	69	926992	108.040	ug/l	97
52) 1,3-Dichloropropane	11.141	76	892266	101.657	ug/l	98
53) Dibromochloromethane	11.341	129	590310	104.918	ug/l	100
54) 1,2-Dibromoethane	11.447	107	537806	103.987	ug/l	99
55) 2-Chloroethyl vinyl ether	10.141	63	2236990	495.251	ug/l	99
56) Bromoform	12.559	173	402717	108.214	ug/l #	98
58) 4-Methyl-2-Pentanone	10.429	43	4678202	505.641	ug/l	99
59) 2-Hexanone	11.176	43	3365074	535.301	ug/l	100
61) Tetrachloroethene	11.082	164	347945	99.301	ug/l	97
62) Toluene	10.612	91	2222660	100.129	ug/l	100
64) Chlorobenzene	11.870	112	1377291	100.975	ug/l	99
65) 1,1,1,2-Tetrachloroethane	11.941	131	486095	101.362	ug/l	98
66) Ethyl Benzene	11.941	91	2510073	101.451	ug/l	97
67) m/p-Xylenes	12.047	106	1841924	202.817	ug/l	99
68) o-Xylene	12.376	106	906317	101.440	ug/l	100
69) Styrene	12.388	104	1585501	103.928	ug/l	99
70) Isopropylbenzene	12.676	105	2334192	102.833	ug/l	100
71) 1,1,2,2-Tetrachloroethane	12.917	83	784278	100.840	ug/l	100
72) 1,2,3-Trichloropropane	12.970	75	753853m	94.296	ug/l	
73) Bromobenzene	12.959	156	540169	102.653	ug/l	98
74) n-propylbenzene	13.012	91	2895579	102.142	ug/l	100
75) 2-Chlorotoluene	13.106	91	1757134	100.765	ug/l	98
76) 1,3,5-Trimethylbenzene	13.153	105	1972967	101.515	ug/l	99
77) t-1,4-Dichloro-2-butene	12.717	75	312541	113.952	ug/l	92
78) 4-Chlorotoluene	13.200	91	1813669	102.633	ug/l	98
79) tert-butylbenzene	13.417	119	1672735	102.082	ug/l	98
80) 1,2,4-Trimethylbenzene	13.459	105	1998320	102.898	ug/l	100
81) sec-Butylbenzene	13.594	105	2416128	101.283	ug/l	100
82) p-Isopropyltoluene	13.706	119	2001765	101.064	ug/l	100
83) 1,3-Dichlorobenzene	13.711	146	1015919	101.387	ug/l	99
84) 1,4-Dichlorobenzene	13.788	146	1028636	101.255	ug/l	99
85) n-Butylbenzene	14.035	91	1987536	102.562	ug/l	98
86) Hexachloroethane	14.311	117	399888	103.386	ug/l	97
87) 1,2-Dichlorobenzene	14.082	146	976220	102.064	ug/l	99
88) 1,2-Dibromo-3-Chloropr...	14.700	75	199216	101.112	ug/l	95
89) 1,2,4-Trichlorobenzene	15.811	180	582952	101.981	ug/l	100
90) Hexachlorobutadiene	15.476	225	212184	95.460	ug/l	95
91) Naphthalene	15.611	128	2259551	103.669	ug/l	99
92) 1,2,3-Trichlorobenzene	15.811	180	582952	101.981	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087455.D
 Acq On : 05 Aug 2025 14:40
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Manual Integrations APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 01:29:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 01:24:35 2025
 Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087456.D
 Acq On : 05 Aug 2025 15:01
 Operator : JC\MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 01:30:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 01:24:35 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.800	128	69078	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.082	114	387583	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	364557	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.565	65	189674	29.770	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery =	99.233%		
60) 4-Bromofluorobenzene	12.829	95	185190	29.518	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery =	98.400%		
63) Toluene-d8	10.547	98	486934	29.029	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery =	96.767%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.142	85	723378	160.391	ug/l	98
3) Chloromethane	2.383	50	725628	156.816	ug/l	99
4) Vinyl Chloride	2.542	62	763252	155.517	ug/l	98
5) Bromomethane	2.948	94	445899	167.854	ug/l	88
6) Chloroethane	3.124	64	492727	155.693	ug/l	98
7) Trichlorofluoromethane	3.506	101	1085005	153.614	ug/l	95
8) Diethyl Ether	3.959	74	454480	154.449	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.359	101	551904	153.229	ug/l	90
10) 1,1-Dichloroethene	4.330	96	595776	157.473	ug/l	97
11) Methyl Iodide	4.577	142	615618	207.681	ug/l	96
12) Methyl Acetate	5.012	43	1117123	160.008	ug/l	96
13) Acrolein	4.171	56	885350	830.485	ug/l	100
14) Acrylonitrile	5.706	53	2386034	782.650	ug/l	99
15) Acetone	4.424	58	624281	716.676	ug/l	93
16) Carbon Disulfide	4.700	76	1818928	154.990	ug/l	99
17) Allyl chloride	5.006	41	1219550	158.149	ug/l	96
18) Methylene Chloride	5.259	84	741791	152.333	ug/l	96
19) trans-1,2-Dichloroethene	5.771	96	697274	154.982	ug/l	97
20) Diisopropyl ether	6.659	45	2747446	156.466	ug/l	98
21) 1,1-Dichloroethane	6.553	63	1407817	155.035	ug/l	99
22) cis-1,2-Dichloroethene	7.471	96	860976	155.531	ug/l	99
23) tert-Butyl Alcohol	5.530	59	986070	761.173	ug/l #	100
24) Methyl tert-Butyl Ether	5.789	73	2719481	157.748	ug/l	98
25) Chloroform	7.947	83	1478128	155.945	ug/l	97
26) Cyclohexane	8.235	56	1125305	154.779	ug/l #	98
29) 1,1-Dichloropropene	8.353	75	964808	152.623	ug/l	98
30) 2-Butanone	7.471	43	3513639	754.051	ug/l	99
31) 2,2-Dichloropropane	7.471	77	1224801	150.224	ug/l	100
32) 1,1,1-Trichloroethane	8.153	97	1249167	154.937	ug/l	96
33) Carbon Tetrachloride	8.347	117	1036608	154.768	ug/l	94
34) Benzene	8.588	78	3070072	153.850	ug/l	98
35) Methacrylonitrile	7.765	41	743083	147.778	ug/l	97
36) 1,2-Dichloroethane	8.653	62	1229451	152.417	ug/l	99
37) Trichloroethene	9.335	130	671378	152.738	ug/l	95
38) Methylcyclohexane	9.582	83	1115098	152.053	ug/l	98
39) 1,2-Dichloropropane	9.600	63	775438	151.562	ug/l	98
40) Dibromomethane	9.688	93	586203	152.708	ug/l	97
41) Bromodichloromethane	9.871	83	1228566	154.020	ug/l	96
42) Vinyl Acetate	6.588	43	11742175	765.203	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087456.D
 Acq On : 05 Aug 2025 15:01
 Operator : JC\MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Quant Time: Aug 06 01:30:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 01:24:35 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	7.547	43	1387131	156.385	ug/l	98
44) Isopropyl Acetate	8.671	43	2332673	152.869	ug/l	99
45) 1,4-Dioxane	9.682	88	293747	3019.679	ug/l #	93
46) Methyl methacrylate	9.665	41	1134360	157.486	ug/l	96
47) n-amyl Acetate	12.482	43	1594700	207.839	ug/l #	59
48) t-1,3-Dichloropropene	10.818	75	1360863	156.815	ug/l	99
49) cis-1,3-Dichloropropene	10.294	75	1364130	156.195	ug/l	96
50) 1,1,2-Trichloroethane	10.994	97	751367	154.919	ug/l	97
51) Ethyl methacrylate	10.853	69	1388488	162.554	ug/l	99
52) 1,3-Dichloropropane	11.141	76	1349103	154.396	ug/l	99
53) Dibromochloromethane	11.341	129	890655	159.011	ug/l	99
54) 1,2-Dibromoethane	11.447	107	801234	155.619	ug/l	98
55) 2-Chloroethyl vinyl ether	10.141	63	3422234	761.062	ug/l	99
56) Bromoform	12.559	173	604236	163.094	ug/l #	99
58) 4-Methyl-2-Pentanone	10.429	43	6842960	729.674	ug/l	98
59) 2-Hexanone	11.176	43	4970555	780.064	ug/l	100
61) Tetrachloroethene	11.082	164	527237	148.447	ug/l	97
62) Toluene	10.612	91	3345117	148.669	ug/l	100
64) Chlorobenzene	11.870	112	2069135	149.658	ug/l	98
65) 1,1,1,2-Tetrachloroethane	11.941	131	721459	148.418	ug/l	99
66) Ethyl Benzene	11.941	91	3762645	150.033	ug/l	98
67) m/p-Xylenes	12.053	106	2745929	298.294	ug/l	98
68) o-Xylene	12.376	106	1358102	149.962	ug/l	99
69) Styrene	12.388	104	2382097	154.045	ug/l	99
70) Isopropylbenzene	12.676	105	3492470	151.792	ug/l	99
71) 1,1,2,2-Tetrachloroethane	12.917	83	1145045	145.247	ug/l	98
72) 1,2,3-Trichloropropane	12.970	75	1117688m	137.926	ug/l	
73) Bromobenzene	12.959	156	798843	149.771	ug/l	99
74) n-propylbenzene	13.012	91	4261831	148.316	ug/l	100
75) 2-Chlorotoluene	13.106	91	2596357	146.890	ug/l	99
76) 1,3,5-Trimethylbenzene	13.153	105	2914301	147.934	ug/l	99
77) t-1,4-Dichloro-2-butene	12.717	75	466370	167.752	ug/l	89
78) 4-Chlorotoluene	13.200	91	2683008	149.787	ug/l	99
79) tert-butylbenzene	13.417	119	2427930	146.177	ug/l	96
80) 1,2,4-Trimethylbenzene	13.459	105	2936296	149.164	ug/l	99
81) sec-Butylbenzene	13.594	105	3544404	146.583	ug/l	99
82) p-Isopropyltoluene	13.706	119	2934141	146.146	ug/l	99
83) 1,3-Dichlorobenzene	13.712	146	1508514	148.523	ug/l	99
84) 1,4-Dichlorobenzene	13.788	146	1521926	147.799	ug/l	100
85) n-Butylbenzene	14.035	91	2920119	148.660	ug/l	99
86) Hexachloroethane	14.311	117	595863	151.982	ug/l	97
87) 1,2-Dichlorobenzene	14.082	146	1446749	149.224	ug/l	100
88) 1,2-Dibromo-3-Chloropr...	14.700	75	299287	149.861	ug/l	96
89) 1,2,4-Trichlorobenzene	15.811	180	893366	154.184	ug/l	99
90) Hexachlorobutadiene	15.476	225	317783	141.046	ug/l	96
91) Naphthalene	15.611	128	3431814	155.335	ug/l	100
92) 1,2,3-Trichlorobenzene	15.811	180	893366	154.184	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\  
Data File : VN087456.D  
Acq On    : 05 Aug 2025 15:01  
Operator  : JC\MD  
Sample    : VSTDICC150  
Misc      : 5.0mL/MSVOA_N/WATER  
ALS Vial  : 10 Sample Multiplier: 1
```

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED

Reviewed By :John Carlone 08/06/2025
Supervised By :Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 01:30:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 06 01:24:35 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087458.D
 Acq On : 05 Aug 2025 15:43
 Operator : JC\MD
 Sample : VSTDICV020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN080525

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 02:12:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.800	128	71823	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.082	114	404390	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	367577	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.559	65	202347	30.545	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery = 101.800%			
60) 4-Bromofluorobenzene	12.829	95	192410	30.417	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery = 101.400%			
63) Toluene-d8	10.547	98	513267	30.347	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery = 101.167%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.142	85	87746	18.712	ug/l	98
3) Chloromethane	2.383	50	85826	17.839	ug/l	97
4) Vinyl Chloride	2.542	62	97923	19.190	ug/l	92
5) Bromomethane	2.983	94	53929	19.525	ug/l	85
6) Chloroethane	3.142	64	59300	18.022	ug/l	96
7) Trichlorofluoromethane	3.506	101	133795	18.219	ug/l	100
8) Diethyl Ether	3.959	74	56065	18.325	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.365	101	68759	18.360	ug/l	88
10) 1,1-Dichloroethene	4.336	96	71378	18.145	ug/l	90
11) Methyl Iodide	4.583	142	48013	15.501	ug/l	96
12) Methyl Acetate	5.018	43	133791	18.431	ug/l	95
13) Acrolein	4.177	56	114938	103.668	ug/l	98
14) Acrylonitrile	5.712	53	287956	90.843	ug/l	98
15) Acetone	4.424	58	88879	98.134	ug/l	96
16) Carbon Disulfide	4.706	76	219977	18.028	ug/l	99
17) Allyl chloride	5.018	41	149799	18.683	ug/l	88
18) Methylene Chloride	5.265	84	93115	18.387	ug/l	96
19) trans-1,2-Dichloroethene	5.765	96	83760	17.906	ug/l	96
20) Diisopropyl ether	6.659	45	327254	17.917	ug/l	99
21) 1,1-Dichloroethane	6.559	63	171210	18.134	ug/l	98
22) cis-1,2-Dichloroethene	7.471	96	103737	18.023	ug/l	100
23) tert-Butyl Alcohol	5.530	59	128000	95.030	ug/l #	100
24) Methyl tert-Butyl Ether	5.789	73	324724	18.116	ug/l	98
25) Chloroform	7.953	83	177598	18.021	ug/l	94
26) Cyclohexane	8.241	56	135890	17.977	ug/l #	99
29) 1,1-Dichloropropene	8.353	75	117565	17.825	ug/l	98
30) 2-Butanone	7.471	43	435716	89.621	ug/l	99
31) 2,2-Dichloropropane	7.477	77	158483	18.630	ug/l	98
32) 1,1,1-Trichloroethane	8.147	97	151313	17.988	ug/l	96
33) Carbon Tetrachloride	8.347	117	124231	17.777	ug/l	93
34) Benzene	8.588	78	376019	18.060	ug/l	95
35) Methacrylonitrile	7.765	41	95290	18.163	ug/l	96
36) 1,2-Dichloroethane	8.653	62	150784	17.918	ug/l	99
37) Trichloroethene	9.335	130	82742	18.041	ug/l	97
38) Methylcyclohexane	9.582	83	136830	17.882	ug/l	98
39) 1,2-Dichloropropane	9.600	63	94981	17.793	ug/l	96
40) Dibromomethane	9.688	93	69241	17.288	ug/l	93
41) Bromodichloromethane	9.871	83	146339	17.583	ug/l	98
42) Vinyl Acetate	6.589	43	1423556	88.913	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087458.D
 Acq On : 05 Aug 2025 15:43
 Operator : JC\MD
 Sample : VSTDICV020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN080525

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 02:12:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	7.553	43	171684	18.279	ug/l	99
44) Isopropyl Acetate	8.677	43	278498	17.492	ug/l	100
45) 1,4-Dioxane	9.677	88	38322	377.572	ug/l	98
46) Methyl methacrylate	9.665	41	127548	16.972	ug/l	92
47) n-amyl Acetate	12.523	43	146094m	16.611	ug/l	
48) t-1,3-Dichloropropene	10.818	75	157686	17.415	ug/l	96
49) cis-1,3-Dichloropropene	10.294	75	162308	17.812	ug/l	99
50) 1,1,2-Trichloroethane	10.994	97	90030	17.791	ug/l	93
51) Ethyl methacrylate	10.859	69	153822	17.260	ug/l	95
52) 1,3-Dichloropropane	11.141	76	161040	17.664	ug/l	99
53) Dibromochloromethane	11.335	129	101996	17.453	ug/l	98
54) 1,2-Dibromoethane	11.453	107	93933	17.486	ug/l	99
55) 2-Chloroethyl vinyl ether	10.141	63	443163	94.458	ug/l	99
56) Bromoform	12.559	173	65682	16.992	ug/l #	99
58) 4-Methyl-2-Pentanone	10.429	43	848509	89.734	ug/l	100
59) 2-Hexanone	11.182	43	570046	88.726	ug/l	100
61) Tetrachloroethene	11.082	164	64788	18.092	ug/l	96
62) Toluene	10.612	91	405628	17.879	ug/l	100
64) Chlorobenzene	11.871	112	248284	17.811	ug/l	99
65) 1,1,1,2-Tetrachloroethane	11.941	131	85480	17.440	ug/l	98
66) Ethyl Benzene	11.941	91	452435	17.892	ug/l	100
67) m/p-Xylenes	12.053	106	323616	34.866	ug/l	100
68) o-Xylene	12.376	106	159060	17.419	ug/l	100
69) Styrene	12.394	104	269120	17.260	ug/l	98
70) Isopropylbenzene	12.676	105	412806	17.794	ug/l	98
71) 1,1,2,2-Tetrachloroethane	12.918	83	140546	17.682	ug/l	100
72) 1,2,3-Trichloropropane	12.976	75	120488m	16.271	ug/l	
73) Bromobenzene	12.959	156	91029	16.926	ug/l	96
74) n-propylbenzene	13.012	91	521263	17.991	ug/l	99
75) 2-Chlorotoluene	13.106	91	315743	17.717	ug/l	99
76) 1,3,5-Trimethylbenzene	13.153	105	356748	17.960	ug/l	100
77) t-1,4-Dichloro-2-butene	12.723	75	47247	16.855	ug/l	90
78) 4-Chlorotoluene	13.200	91	327647	18.142	ug/l	99
79) tert-butylbenzene	13.417	119	299129	17.861	ug/l	97
80) 1,2,4-Trimethylbenzene	13.459	105	358425	18.058	ug/l	98
81) sec-Butylbenzene	13.594	105	447713	18.364	ug/l	99
82) p-Isopropyltoluene	13.706	119	367898	18.174	ug/l	99
83) 1,3-Dichlorobenzene	13.712	146	187078	18.268	ug/l	99
84) 1,4-Dichlorobenzene	13.788	146	188698	18.174	ug/l	99
85) n-Butylbenzene	14.035	91	370033	18.683	ug/l	99
86) Hexachloroethane	14.312	117	69621	17.612	ug/l	95
87) 1,2-Dichlorobenzene	14.082	146	178646	18.275	ug/l	99
88) 1,2-Dibromo-3-Chloropr...	14.694	75	35995	17.876	ug/l	99
89) 1,2,4-Trichlorobenzene	15.811	180	107490	18.399	ug/l	99
90) Hexachlorobutadiene	15.476	225	42365	18.649	ug/l	95
91) Naphthalene	15.611	128	404839	18.174	ug/l	99
92) 1,2,3-Trichlorobenzene	15.811	180	107490	18.399	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\  
Data File : VN087458.D  
Acq On    : 05 Aug 2025   15:43  
Operator  : JC\MD  
Sample    : VSTDICV020  
Misc      : 5.0mL/MSVOA_N/WATER  
ALS Vial  : 12   Sample Multiplier: 1
```

Instrument :
MSVOA_N
ClientSampleId :
ICVVN080525

Manual Integrations APPROVED

Reviewed By :John Carlone 08/06/2025
Supervised By :Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 02:12:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 06 02:01:21 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087458.D
 Acq On : 05 Aug 2025 15:43
 Operator : JC\MD
 Sample : VSTDICV020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN080525

Quant Time: Aug 06 02:12:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	112	0.00
2 M	Dichlorodifluoromethane	1.959	1.833	6.4	99	0.00
3 M	Chloromethane	2.010	1.792	10.8	93	0.00
4 M	Vinyl Chloride	2.131	2.045	4.0	100	0.00
5 M	Bromomethane	1.154	1.126	2.4	109	0.00
6 M	Chloroethane	1.374	1.238	9.9	96	0.00
7 M	Trichlorofluoromethane	3.067	2.794	8.9	97	0.00
8 T	Diethyl Ether	1.278	1.171	8.4	99	0.00
9	1,1,2-Trichlorotrifluoroeth	1.564	1.436	8.2	101	0.00
10 M	1,1-Dichloroethene	1.643	1.491	9.3	96	0.00
11	Methyl Iodide	1.294	1.003	22.5	104	0.00
12	Methyl Acetate	3.032	2.794	7.8	101	0.00
13 M	Acrolein	0.463	0.480	-3.7	116	0.00
14 M	Acrylonitrile	1.324	1.203	9.1	99	0.00
15 M	Acetone	0.378	0.371	1.9	99	0.00
16 M	Carbon Disulfide	5.097	4.594	9.9	99	0.00
17	Allyl chloride	3.349	3.129	6.6	104	0.00
18 M	Methylene Chloride	2.115	1.945	8.0	100	0.00
19 M	trans-1,2-Dichloroethene	1.954	1.749	10.5	97	0.00
20 T	Diisopropyl ether	7.629	6.835	10.4	98	0.00
21 M	1,1-Dichloroethane	3.944	3.576	9.3	96	0.00
22 M	cis-1,2-Dichloroethene	2.404	2.167	9.9	97	0.00
23 M	tert-Butyl Alcohol	0.563	0.535	5.0	105	0.01
24 M	Methyl tert-Butyl Ether	7.487	6.782	9.4	99	0.00
25 M	Chloroform	4.116	3.709	9.9	98	0.00
26	Cyclohexane	3.157	2.838	10.1	95	0.00
27 s	1,2-Dichloroethane-d4	2.767	2.817	-1.8	111	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	110	0.00
29	1,1-Dichloropropene	0.489	0.436	10.8	98	0.00
30 M	2-Butanone	0.361	0.323	10.5	98	0.00
31	2,2-Dichloropropane	0.631	0.588	6.8	101	0.00
32 M	1,1,1-Trichloroethane	0.624	0.561	10.1	100	0.00
33 M	Carbon Tetrachloride	0.518	0.461	11.0	98	0.00
34 M	Benzene	1.545	1.395	9.7	100	0.00
35	Methacrylonitrile	0.389	0.353	9.3	98	0.00
36 M	1,2-Dichloroethane	0.624	0.559	10.4	97	0.00
37 M	Trichloroethene	0.340	0.307	9.7	98	0.00
38	Methylcyclohexane	0.568	0.508	10.6	98	0.00
39 M	1,2-Dichloropropane	0.396	0.352	11.1	97	0.00
40	Dibromomethane	0.297	0.257	13.5	94	0.00
41 M	Bromodichloromethane	0.617	0.543	12.0	99	0.00
42 M	Vinyl Acetate	1.188	1.056	11.1	96	0.00
43	Ethyl Acetate	0.697	0.637	8.6	100	0.00
44	Isopropyl Acetate	1.181	1.033	12.5	96	0.00
45 T	1,4-Dioxane	0.008	0.007	12.5	101	0.00
46	Methyl methacrylate	0.558	0.473	15.2	93	0.00
47	n-amyl Acetate	0.652	0.542	16.9	106	0.00
48 M	t-1,3-Dichloropropene	0.672	0.585	12.9	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087458.D
 Acq On : 05 Aug 2025 15:43
 Operator : JC\MD
 Sample : VSTDICV020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN080525

Quant Time: Aug 06 02:12:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.676	0.602	10.9	100	0.00
50 M	1,1,2-Trichloroethane	0.375	0.334	10.9	98	0.00
51	Ethyl methacrylate	0.661	0.571	13.6	96	0.00
52	1,3-Dichloropropane	0.676	0.597	11.7	97	0.00
53 M	Dibromochloromethane	0.434	0.378	12.9	98	0.00
54 M	1,2-Dibromoethane	0.399	0.348	12.8	96	0.00
55 M	2-Chloroethyl vinyl ether	0.348	0.329	5.5	104	0.00
56 M	Bromoform	0.287	0.244	15.0	97	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	112	0.00
58 M	4-Methyl-2-Pentanone	0.772	0.693	10.2	97	0.00
59 M	2-Hexanone	0.524	0.465	11.3	99	0.00
60 S	4-Bromofluorobenzene	0.516	0.523	-1.4	113	0.00
61 M	Tetrachloroethene	0.292	0.264	9.6	101	0.00
62 M	Toluene	1.852	1.655	10.6	97	0.00
63 S	Toluene-d8	1.380	1.396	-1.2	113	0.00
64 M	Chlorobenzene	1.138	1.013	11.0	98	0.00
65	1,1,1,2-Tetrachloroethane	0.400	0.349	12.8	97	0.00
66 M	Ethyl Benzene	2.064	1.846	10.6	98	0.00
67 M	m/p-Xylenes	0.758	0.660	12.9	96	0.00
68 M	o-Xylene	0.745	0.649	12.9	97	0.00
69 M	Styrene	1.273	1.098	13.7	95	0.00
70	Isopropylbenzene	1.893	1.685	11.0	98	0.00
71 M	1,1,2,2-Tetrachloroethane	0.649	0.574	11.6	95	0.00
72	1,2,3-Trichloropropane	0.604	0.492	18.5	97	0.00
73	Bromobenzene	0.439	0.371	15.5	95	0.00
74	n-propylbenzene	2.365	2.127	10.1	99	0.00
75	2-Chlorotoluene	1.455	1.288	11.5	97	0.00
76	1,3,5-Trimethylbenzene	1.621	1.456	10.2	98	0.00
77	t-1,4-Dichloro-2-butene	0.229	0.193	15.7	100	0.00
78	4-Chlorotoluene	1.474	1.337	9.3	102	0.00
79	tert-butylbenzene	1.367	1.221	10.7	98	0.00
80	1,2,4-Trimethylbenzene	1.620	1.463	9.7	100	0.00
81	sec-Butylbenzene	1.990	1.827	8.2	100	0.00
82	p-Isopropyltoluene	1.652	1.501	9.1	99	0.00
83 M	1,3-Dichlorobenzene	0.836	0.763	8.7	100	0.00
84 M	1,4-Dichlorobenzene	0.847	0.770	9.1	99	0.00
85	n-Butylbenzene	1.616	1.510	6.6	102	0.00
86 T	Hexachloroethane	0.323	0.284	12.1	98	0.00
87 M	1,2-Dichlorobenzene	0.798	0.729	8.6	100	0.00
88	1,2-Dibromo-3-Chloropropane	0.164	0.147	10.4	97	0.00
89	1,2,4-Trichlorobenzene	0.477	0.439	8.0	103	0.00
90	Hexachlorobutadiene	0.185	0.173	6.5	103	0.00
91 M	Naphthalene	1.818	1.652	9.1	101	0.00
92	1,2,3-Trichlorobenzene	0.477	0.439	8.0	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087458.D
 Acq On : 05 Aug 2025 15:43
 Operator : JC\MD
 Sample : VSTDICV020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN080525

Quant Time: Aug 06 02:12:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	30.000	30.000	0.0	112	0.00
2 M	Dichlorodifluoromethane	20.000	18.712	6.4	99	0.00
3 M	Chloromethane	20.000	17.839	10.8	93	0.00
4 M	Vinyl Chloride	20.000	19.190	4.0	100	0.00
5 M	Bromomethane	20.000	19.525	2.4	109	0.00
6 M	Chloroethane	20.000	18.022	9.9	96	0.00
7 M	Trichlorofluoromethane	20.000	18.219	8.9	97	0.00
8 T	Diethyl Ether	20.000	18.325	8.4	99	0.00
9	1,1,2-Trichlorotrifluoroeth	20.000	18.360	8.2	101	0.00
10 M	1,1-Dichloroethene	20.000	18.145	9.3	96	0.00
11	Methyl Iodide	20.000	15.501	22.5	104	0.00
12	Methyl Acetate	20.000	18.431	7.8	101	0.00
13 M	Acrolein	100.000	103.668	-3.7	116	0.00
14 M	Acrylonitrile	100.000	90.843	9.2	99	0.00
15 M	Acetone	100.000	98.134	1.9	99	0.00
16 M	Carbon Disulfide	20.000	18.028	9.9	99	0.00
17	Allyl chloride	20.000	18.683	6.6	104	0.00
18 M	Methylene Chloride	20.000	18.387	8.1	100	0.00
19 M	trans-1,2-Dichloroethene	20.000	17.906	10.5	97	0.00
20 T	Diisopropyl ether	20.000	17.917	10.4	98	0.00
21 M	1,1-Dichloroethane	20.000	18.134	9.3	96	0.00
22 M	cis-1,2-Dichloroethene	20.000	18.023	9.9	97	0.00
23 M	tert-Butyl Alcohol	100.000	95.030	5.0	105	0.01
24 M	Methyl tert-Butyl Ether	20.000	18.116	9.4	99	0.00
25 M	Chloroform	20.000	18.021	9.9	98	0.00
26	Cyclohexane	20.000	17.977	10.1	95	0.00
27 s	1,2-Dichloroethane-d4	30.000	30.545	-1.8	111	0.00
28 I	1,4-Difluorobenzene	30.000	30.000	0.0	110	0.00
29	1,1-Dichloropropene	20.000	17.825	10.9	98	0.00
30 M	2-Butanone	100.000	89.621	10.4	98	0.00
31	2,2-Dichloropropane	20.000	18.630	6.9	101	0.00
32 M	1,1,1-Trichloroethane	20.000	17.988	10.1	100	0.00
33 M	Carbon Tetrachloride	20.000	17.777	11.1	98	0.00
34 M	Benzene	20.000	18.060	9.7	100	0.00
35	Methacrylonitrile	20.000	18.163	9.2	98	0.00
36 M	1,2-Dichloroethane	20.000	17.918	10.4	97	0.00
37 M	Trichloroethene	20.000	18.041	9.8	98	0.00
38	Methylcyclohexane	20.000	17.882	10.6	98	0.00
39 M	1,2-Dichloropropane	20.000	17.793	11.0	97	0.00
40	Dibromomethane	20.000	17.288	13.6	94	0.00
41 M	Bromodichloromethane	20.000	17.583	12.1	99	0.00
42 M	Vinyl Acetate	100.000	88.913	11.1	96	0.00
43	Ethyl Acetate	20.000	18.279	8.6	100	0.00
44	Isopropyl Acetate	20.000	17.492	12.5	96	0.00
45 T	1,4-Dioxane	400.000	377.572	5.6	101	0.00
46	Methyl methacrylate	20.000	16.972	15.1	93	0.00
47	n-amyl Acetate	20.000	16.611	16.9	106	0.00
48 M	t-1,3-Dichloropropene	20.000	17.415	12.9	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087458.D
 Acq On : 05 Aug 2025 15:43
 Operator : JC\MD
 Sample : VSTDICV020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN080525

Quant Time: Aug 06 02:12:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	20.000	17.812	10.9	100	0.00
50 M	1,1,2-Trichloroethane	20.000	17.791	11.0	98	0.00
51	Ethyl methacrylate	20.000	17.260	13.7	96	0.00
52	1,3-Dichloropropane	20.000	17.664	11.7	97	0.00
53 M	Dibromochloromethane	20.000	17.453	12.7	98	0.00
54 M	1,2-Dibromoethane	20.000	17.486	12.6	96	0.00
55 M	2-Chloroethyl vinyl ether	100.000	94.458	5.5	104	0.00
56 M	Bromoform	20.000	16.992	15.0	97	0.00
57 I	Chlorobenzene-d5	30.000	30.000	0.0	112	0.00
58 M	4-Methyl-2-Pentanone	100.000	89.734	10.3	97	0.00
59 M	2-Hexanone	100.000	88.726	11.3	99	0.00
60 S	4-Bromofluorobenzene	30.000	30.417	-1.4	113	0.00
61 M	Tetrachloroethene	20.000	18.092	9.5	101	0.00
62 M	Toluene	20.000	17.879	10.6	97	0.00
63 S	Toluene-d8	30.000	30.347	-1.2	113	0.00
64 M	Chlorobenzene	20.000	17.811	10.9	98	0.00
65	1,1,1,2-Tetrachloroethane	20.000	17.440	12.8	97	0.00
66 M	Ethyl Benzene	20.000	17.892	10.5	98	0.00
67 M	m/p-Xylenes	40.000	34.866	12.8	96	0.00
68 M	o-Xylene	20.000	17.419	12.9	97	0.00
69 M	Styrene	20.000	17.260	13.7	95	0.00
70	Isopropylbenzene	20.000	17.794	11.0	98	0.00
71 M	1,1,2,2-Tetrachloroethane	20.000	17.682	11.6	95	0.00
72	1,2,3-Trichloropropane	20.000	16.271	18.6	97	0.00
73	Bromobenzene	20.000	16.926	15.4	95	0.00
74	n-propylbenzene	20.000	17.991	10.0	99	0.00
75	2-Chlorotoluene	20.000	17.717	11.4	97	0.00
76	1,3,5-Trimethylbenzene	20.000	17.960	10.2	98	0.00
77	t-1,4-Dichloro-2-butene	20.000	16.855	15.7	100	0.00
78	4-Chlorotoluene	20.000	18.142	9.3	102	0.00
79	tert-butylbenzene	20.000	17.861	10.7	98	0.00
80	1,2,4-Trimethylbenzene	20.000	18.058	9.7	100	0.00
81	sec-Butylbenzene	20.000	18.364	8.2	100	0.00
82	p-Isopropyltoluene	20.000	18.174	9.1	99	0.00
83 M	1,3-Dichlorobenzene	20.000	18.268	8.7	100	0.00
84 M	1,4-Dichlorobenzene	20.000	18.174	9.1	99	0.00
85	n-Butylbenzene	20.000	18.683	6.6	102	0.00
86 T	Hexachloroethane	20.000	17.612	11.9	98	0.00
87 M	1,2-Dichlorobenzene	20.000	18.275	8.6	100	0.00
88	1,2-Dibromo-3-Chloropropane	20.000	17.876	10.6	97	0.00
89	1,2,4-Trichlorobenzene	20.000	18.399	8.0	103	0.00
90	Hexachlorobutadiene	20.000	18.649	6.8	103	0.00
91 M	Naphthalene	20.000	18.174	9.1	101	0.00
92	1,2,3-Trichlorobenzene	20.000	18.399	8.0	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TULL01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2771</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>08/05/2025 13:58</u>
Lab File ID:	<u>VN087453.D</u>	Init. Calib. Date(s):	<u>08/05/2025 08/05/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>13:37 15:01</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.545	1.540		-0.32	
Toluene	1.852	1.905		2.86	
Ethyl Benzene	2.064	2.115		2.47	
m/p-Xylenes	0.758	0.768		1.32	
o-Xylene	0.745	0.754		1.21	
1,2-Dichloroethane-d4	2.767	2.838		2.57	
Toluene-d8	1.380	1.382		0.14	
4-Bromofluorobenzene	0.516	0.520		0.77	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087453.D
 Acq On : 05 Aug 2025 13:58
 Operator : JC\MD
 Sample : VSTDICCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 01:27:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 01:24:35 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.800	128	64218	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.082	114	367115	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	327914	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.559	65	182238	30.767	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery = 102.567%			
60) 4-Bromofluorobenzene	12.829	95	170429	30.201	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery = 100.667%			
63) Toluene-d8	10.547	98	453121	30.032	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery = 100.100%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.142	85	88427	21.090	ug/l	100
3) Chloromethane	2.383	50	92684	21.546	ug/l	100
4) Vinyl Chloride	2.542	62	97984	21.476	ug/l	100
5) Bromomethane	2.983	94	49674	20.114	ug/l	100
6) Chloroethane	3.142	64	61674	20.963	ug/l	100
7) Trichlorofluoromethane	3.512	101	137336	20.915	ug/l	100
8) Diethyl Ether	3.965	74	56555	20.674	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.371	101	68121	20.344	ug/l	100
10) 1,1-Dichloroethene	4.336	96	74078	21.062	ug/l	100
11) Methyl Iodide	4.589	142	46084m	16.723	ug/l	
12) Methyl Acetate	5.018	43	132440	20.405	ug/l	100
13) Acrolein	4.171	56	99154	100.048	ug/l	99
14) Acrylonitrile	5.712	53	290627	102.544	ug/l	100
15) Acetone	4.418	58	89398	110.396	ug/l	100
16) Carbon Disulfide	4.700	76	221586	20.310	ug/l	100
17) Allyl chloride	5.012	41	144593	20.170	ug/l	100
18) Methylene Chloride	5.271	84	92808m	20.501	ug/l	
19) trans-1,2-Dichloroethene	5.771	96	86216	20.613	ug/l	100
20) Diisopropyl ether	6.665	45	334922	20.517	ug/l	100
21) 1,1-Dichloroethane	6.553	63	177532	21.030	ug/l	100
22) cis-1,2-Dichloroethene	7.471	96	106998	20.791	ug/l	100
23) tert-Butyl Alcohol	5.518	59	121824	101.156	ug/l #	100
24) Methyl tert-Butyl Ether	5.789	73	327775	20.452	ug/l	100
25) Chloroform	7.947	83	180797	20.518	ug/l	100
26) Cyclohexane	8.241	56	142519	21.086	ug/l	100
29) 1,1-Dichloropropene	8.353	75	120007	20.042	ug/l	100
30) 2-Butanone	7.471	43	446235	101.104	ug/l	100
31) 2,2-Dichloropropane	7.471	77	156255	20.233	ug/l	100
32) 1,1,1-Trichloroethane	8.147	97	151297	19.812	ug/l	100
33) Carbon Tetrachloride	8.341	117	127325	20.070	ug/l	100
34) Benzene	8.588	78	377003	19.946	ug/l	100
35) Methacrylonitrile	7.765	41	97561	20.484	ug/l	100
36) 1,2-Dichloroethane	8.653	62	155656	20.373	ug/l	100
37) Trichloroethene	9.335	130	84369	20.264	ug/l	100
38) Methylcyclohexane	9.582	83	139951	20.147	ug/l	100
39) 1,2-Dichloropropane	9.600	63	98038	20.230	ug/l	100
40) Dibromomethane	9.688	93	73356	20.175	ug/l	100
41) Bromodichloromethane	9.871	83	148435	19.646	ug/l	100
42) Vinyl Acetate	6.588	43	1483194	102.044	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087453.D
 Acq On : 05 Aug 2025 13:58
 Operator : JC\MD
 Sample : VSTDICCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 01:27:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 01:24:35 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	7.553	43	170967m	20.349	ug/l	
44) Isopropyl Acetate	8.677	43	291509	20.169	ug/l	100
45) 1,4-Dioxane	9.677	88	37910	411.437	ug/l	100
46) Methyl methacrylate	9.665	41	136612	20.024	ug/l	100
47) n-amyl Acetate	12.529	43	138459m	19.052	ug/l	
48) t-1,3-Dichloropropene	10.818	75	162191	19.732	ug/l	100
49) cis-1,3-Dichloropropene	10.294	75	162479	19.641	ug/l	100
50) 1,1,2-Trichloroethane	10.994	97	91845	19.993	ug/l	100
51) Ethyl methacrylate	10.853	69	159690	19.738	ug/l	100
52) 1,3-Dichloropropane	11.141	76	165895	20.044	ug/l	100
53) Dibromochloromethane	11.341	129	103790	19.563	ug/l	100
54) 1,2-Dibromoethane	11.447	107	97894	20.073	ug/l	100
55) 2-Chloroethyl vinyl ether	10.141	63	426876	100.225	ug/l	100
56) Bromoform	12.559	173	67550	19.250	ug/l	100
58) 4-Methyl-2-Pentanone	10.424	43	876419	103.897	ug/l	100
59) 2-Hexanone	11.182	43	576264	100.543	ug/l	100
61) Tetrachloroethene	11.082	164	63867	19.992	ug/l	100
62) Toluene	10.612	91	416382	20.573	ug/l	100
64) Chlorobenzene	11.871	112	252716	20.321	ug/l	100
65) 1,1,1,2-Tetrachloroethane	11.941	131	88390	20.216	ug/l	100
66) Ethyl Benzene	11.941	91	462372	20.497	ug/l	100
67) m/p-Xylenes	12.047	106	335952	40.573	ug/l	100
68) o-Xylene	12.376	106	164778	20.228	ug/l	100
69) Styrene	12.394	104	283102	20.353	ug/l	100
70) Isopropylbenzene	12.670	105	423292	20.453	ug/l	100
71) 1,1,2,2-Tetrachloroethane	12.918	83	148347	20.920	ug/l	100
72) 1,2,3-Trichloropropane	12.970	75	123856m	16.992	ug/l	
73) Bromobenzene	12.959	156	96161	20.043	ug/l	100
74) n-propylbenzene	13.012	91	526811	20.382	ug/l	100
75) 2-Chlorotoluene	13.106	91	324478	20.409	ug/l	100
76) 1,3,5-Trimethylbenzene	13.153	105	363042	20.488	ug/l	100
77) t-1,4-Dichloro-2-butene	12.718	75	47119	18.843	ug/l	100
78) 4-Chlorotoluene	13.200	91	322316	20.005	ug/l	100
79) tert-butylbenzene	13.417	119	305702	20.462	ug/l	100
80) 1,2,4-Trimethylbenzene	13.459	105	358349	20.238	ug/l	100
81) sec-Butylbenzene	13.594	105	447541	20.577	ug/l	100
82) p-Isopropyltoluene	13.706	119	369955	20.486	ug/l	100
83) 1,3-Dichlorobenzene	13.712	146	186482	20.412	ug/l	100
84) 1,4-Dichlorobenzene	13.788	146	190232	20.538	ug/l	100
85) n-Butylbenzene	14.035	91	363428	20.569	ug/l	100
86) Hexachloroethane	14.312	117	71198	20.189	ug/l	100
87) 1,2-Dichlorobenzene	14.082	146	179117	20.539	ug/l	100
88) 1,2-Dibromo-3-Chloropr...	14.694	75	37135	20.672	ug/l	100
89) 1,2,4-Trichlorobenzene	15.811	180	104426	20.037	ug/l	100
90) Hexachlorobutadiene	15.476	225	41238	20.349	ug/l	100
91) Naphthalene	15.611	128	402240	20.241	ug/l	100
92) 1,2,3-Trichlorobenzene	15.811	180	104426	20.037	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
Data File : VN087453.D
Acq On : 05 Aug 2025 13:58
Operator : JC\MD
Sample : VSTDICCC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020

Manual Integrations APPROVED

Reviewed By :John Carlone 08/06/2025
Supervised By :Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 01:27:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 06 01:24:35 2025
Response via : Initial Calibration







QC SAMPLE DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
Data File : VN087447.D
Acq On : 05 Aug 2025 09:27
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Wed Aug 06 02:01:21 2025



AutoFind: Averaged scan 1841 to 1856; Bkg corrected with scan 1840

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	24.6	6530	PASS
75	95	30	60	58.7	15571	PASS
95	95	100	100	100.0	26522	PASS
96	95	5	9	6.7	1774	PASS
173	174	0.00	2	1.5	263	PASS
174	95	50	100	66.8	17704	PASS
175	174	5	9	7.2	1276	PASS
176	174	95	101	97.7	17301	PASS
177	176	5	9	7.3	1269	PASS

Report of Analysis

Client:	Tully Environmental, Inc			Date Collected:	
Project:	Transfer Station-SPDES			Date Received:	
Client Sample ID:	VN0805WBL02			SDG No.:	Q2771
Lab Sample ID:	VN0805WBL02			Matrix:	Water
Analytical Method:	E624.1			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOC-BTEX
GC Column:	RXI-624	ID :	0.25	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087461.D	1	08/05/25 17:32	VN080525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
71-43-2	Benzene	0.45	U	0.45	5.00	ug/L
108-88-3	Toluene	0.46	U	0.46	5.00	ug/L
100-41-4	Ethyl Benzene	0.56	U	0.56	5.00	ug/L
179601-23-1	m/p-Xylenes	1.30	U	1.30	10.0	ug/L
95-47-6	o-Xylene	0.67	U	0.67	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	30.6		91 - 110	102%	SPK: 30
2037-26-5	Toluene-d8	30.0		91 - 112	100%	SPK: 30
460-00-4	4-Bromofluorobenzene	28.2		63 - 112	94%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	68200	7.8			
540-36-3	1,4-Difluorobenzene	376000	9.082			
3114-55-4	Chlorobenzene-d5	334000	11.847			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of OC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087461.D
 Acq On : 05 Aug 2025 17:32
 Operator : JC\MD
 Sample : VN0805WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0805WBL02

Quant Time: Aug 06 02:43:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.800	128	68205	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.082	114	376206	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	334367	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.559	65	192416	30.586	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.967%
60) 4-Bromofluorobenzene	12.829	95	162550	28.249	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	94.167%
63) Toluene-d8	10.547	98	462100	30.036	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.133%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
Data File : VN087461.D
Acq On : 05 Aug 2025 17:32
Operator : JC\MD
Sample : VN0805WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0805WBL02

Quant Time: Aug 06 02:43:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 06 02:01:21 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 7.800 min Scan# 994
Delta R.T. 0.000 min
Lab File: VN087461.D
Acq: 05 Aug 2025 17:32




Instrument :
MSVOA_N
ClientSampleId :
VN0805WBL02


Tgt Ion	Ratio	Lower	Upper
128	100		
49	189.8	0.0	494.8
130	127.1	0.0	321.5



#27
1,2-Dichloroethane-d4
Concen: 30.586 ug/l
RT: 8.559 min Scan# 1123
Delta R.T. 0.000 min
Lab File: VN087461.D
Acq: 05 Aug 2025 17:32

Tgt Ion	Ratio	Lower	Upper
65	100		
67	50.8	40.8	61.2





#28

1,4-Difluorobenzene

Concen: 30.000 ug/l

RT: 9.082 min Scan# 1212

Delta R.T. 0.000 min

Lab File: VN087461.D

Acq: 05 Aug 2025 17:32

Instrument :

MSVOA_N

ClientSampleId :

VN0805WBL02





Tgt Ion:114 Resp: 376206

Ion Ratio Lower Upper

114 100

63 24.0 18.9 28.3

88 16.7 13.1 19.7



#57

Chlorobenzene-d5

Concen: 30.000 ug/l

RT: 11.847 min Scan# 1682

Delta R.T. 0.000 min

Lab File: VN087461.D

Acq: 05 Aug 2025 17:32




Tgt Ion:117 Resp: 334367


Ion Ratio Lower Upper

117 100

82 60.0 49.2 73.8

119 30.7 25.7 38.5









#60
4-Bromofluorobenzene
Concen: 28.249 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087461.D
Acq: 05 Aug 2025 17:32




Instrument :
MSVOA_N
ClientSampleId :
VN0805WBL02

Tgt Ion: 95 Resp: 162550
Ion Ratio Lower Upper
95 100
174 66.0 52.3 78.5
176 64.4 49.8 74.8



#63
Toluene-d8
Concen: 30.036 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. 0.000 min
Lab File: VN087461.D
Acq: 05 Aug 2025 17:32

Tgt Ion: 98 Resp: 462100
Ion Ratio Lower Upper
98 100
100 64.1 50.5 75.7



Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	
Project:	Transfer Station-SPDES	Date Received:	
Client Sample ID:	VN0805WBS02	SDG No.:	Q2771
Lab Sample ID:	VN0805WBS02	Matrix:	Water
Analytical Method:	E624.1	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-BTEX
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087459.D	1	08/05/25 16:17	VN080525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
71-43-2	Benzene	18.5		0.45	5.00	ug/L
108-88-3	Toluene	18.6		0.46	5.00	ug/L
100-41-4	Ethyl Benzene	18.8		0.56	5.00	ug/L
179601-23-1	m/p-Xylenes	37.4		1.30	10.0	ug/L
95-47-6	o-Xylene	18.6		0.67	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	30.0		91 - 110	100%	SPK: 30
2037-26-5	Toluene-d8	30.7		91 - 112	102%	SPK: 30
460-00-4	4-Bromofluorobenzene	30.9		63 - 112	103%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	69900	7.8			
540-36-3	1,4-Difluorobenzene	388000	9.082			
3114-55-4	Chlorobenzene-d5	346000	11.847			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087459.D
 Acq On : 05 Aug 2025 16:17
 Operator : JC\MD
 Sample : VN0805WBS02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0805WBS02

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 02:41:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.800	128	69933	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.082	114	387635	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	345816	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.559	65	193705	30.030	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery = 100.100%			
60) 4-Bromofluorobenzene	12.829	95	183875	30.897	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery = 103.000%			
63) Toluene-d8	10.547	98	488074	30.674	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery = 102.233%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.142	85	90357	19.790	ug/l	96
3) Chloromethane	2.383	50	89767	19.162	ug/l	98
4) Vinyl Chloride	2.542	62	95110	19.142	ug/l	97
5) Bromomethane	2.983	94	54667	20.327	ug/l	89
6) Chloroethane	3.142	64	59839	18.677	ug/l	96
7) Trichlorofluoromethane	3.512	101	135051	18.887	ug/l	99
8) Diethyl Ether	3.965	74	55670	18.687	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.365	101	67747	18.579	ug/l	64
10) 1,1-Dichloroethene	4.330	96	74028	19.328	ug/l	90
11) Methyl Iodide	4.583	142	49081	16.274	ug/l	96
12) Methyl Acetate	5.018	43	130786	18.504	ug/l #	91
13) Acrolein	4.177	56	105382	97.618	ug/l	100
14) Acrylonitrile	5.712	53	293326	95.038	ug/l	97
15) Acetone	4.424	58	83888	95.126	ug/l	99
16) Carbon Disulfide	4.700	76	223354	18.799	ug/l	98
17) Allyl chloride	5.018	41	141738	18.156	ug/l	96
18) Methylene Chloride	5.265	84	89525	18.156	ug/l	94
19) trans-1,2-Dichloroethene	5.783	96	80631	17.703	ug/l	97
20) Diisopropyl ether	6.665	45	331415	18.635	ug/l	96
21) 1,1-Dichloroethane	6.559	63	169931	18.485	ug/l	97
22) cis-1,2-Dichloroethene	7.471	96	102932	18.367	ug/l	99
23) tert-Butyl Alcohol	5.518	59	125827	95.942	ug/l #	100
24) Methyl tert-Butyl Ether	5.789	73	318368	18.242	ug/l	98
25) Chloroform	7.947	83	177551	18.503	ug/l	94
26) Cyclohexane	8.236	56	140655	19.110	ug/l #	98
29) 1,1-Dichloropropene	8.353	75	116842	18.481	ug/l	99
30) 2-Butanone	7.471	43	430868	92.455	ug/l	100
31) 2,2-Dichloropropane	7.477	77	153167	18.784	ug/l	99
32) 1,1,1-Trichloroethane	8.153	97	150687	18.688	ug/l	95
33) Carbon Tetrachloride	8.347	117	122140	18.233	ug/l	96
34) Benzene	8.588	78	368280	18.453	ug/l	99
35) Methacrylonitrile	7.765	41	96374	19.163	ug/l	93
36) 1,2-Dichloroethane	8.653	62	150384	18.643	ug/l	100
37) Trichloroethene	9.335	130	81601	18.562	ug/l	93
38) Methylcyclohexane	9.582	83	141390	19.277	ug/l	99
39) 1,2-Dichloropropane	9.600	63	94473	18.463	ug/l	99
40) Dibromomethane	9.694	93	69720	18.160	ug/l	96
41) Bromodichloromethane	9.871	83	143179	17.947	ug/l	97
42) Vinyl Acetate	6.594	43	1416485	92.296	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087459.D
 Acq On : 05 Aug 2025 16:17
 Operator : JC\MD
 Sample : VN0805WBS02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0805WBS02

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 02:41:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	7.547	43	171292	19.026	ug/l	96
44) Isopropyl Acetate	8.677	43	283980	18.608	ug/l	98
45) 1,4-Dioxane	9.682	88	39804	409.125	ug/l #	68
46) Methyl methacrylate	9.665	41	125699	17.449	ug/l	91
47) n-amyl Acetate	12.523	43	140251m	16.636	ug/l	
48) t-1,3-Dichloropropene	10.818	75	157664	18.165	ug/l	100
49) cis-1,3-Dichloropropene	10.294	75	161680	18.510	ug/l	97
50) 1,1,2-Trichloroethane	11.000	97	88006	18.143	ug/l	97
51) Ethyl methacrylate	10.859	69	151713	17.759	ug/l	98
52) 1,3-Dichloropropane	11.147	76	159776	18.283	ug/l	100
53) Dibromochloromethane	11.341	129	102903	18.369	ug/l	100
54) 1,2-Dibromoethane	11.453	107	95094	18.467	ug/l	97
55) 2-Chloroethyl vinyl ether	10.141	63	445893	99.148	ug/l	99
56) Bromoform	12.565	173	63691	17.189	ug/l #	99
58) 4-Methyl-2-Pentanone	10.429	43	845319	95.022	ug/l	100
59) 2-Hexanone	11.182	43	562782	93.108	ug/l	99
61) Tetrachloroethene	11.082	164	64602	19.175	ug/l	91
62) Toluene	10.612	91	397589	18.628	ug/l	99
64) Chlorobenzene	11.871	112	244781	18.664	ug/l	94
65) 1,1,1,2-Tetrachloroethane	11.941	131	86126	18.678	ug/l	96
66) Ethyl Benzene	11.941	91	447834	18.825	ug/l	98
67) m/p-Xylenes	12.053	106	326294	37.367	ug/l	98
68) o-Xylene	12.376	106	159596	18.578	ug/l	100
69) Styrene	12.394	104	272250	18.560	ug/l	99
70) Isopropylbenzene	12.676	105	407629	18.677	ug/l	100
71) 1,1,2,2-Tetrachloroethane	12.918	83	139898	18.708	ug/l	100
72) 1,2,3-Trichloropropane	12.976	75	123625m	17.746	ug/l	
73) Bromobenzene	12.959	156	91679	18.120	ug/l	98
74) n-propylbenzene	13.018	91	514787	18.886	ug/l	99
75) 2-Chlorotoluene	13.106	91	312510	18.639	ug/l	98
76) 1,3,5-Trimethylbenzene	13.153	105	354138	18.951	ug/l	99
77) t-1,4-Dichloro-2-butene	12.718	75	46125	17.490	ug/l	90
78) 4-Chlorotoluene	13.200	91	317821	18.705	ug/l	100
79) tert-butylbenzene	13.418	119	289670	18.385	ug/l	97
80) 1,2,4-Trimethylbenzene	13.465	105	357197	19.129	ug/l	97
81) sec-Butylbenzene	13.594	105	441251	19.237	ug/l	99
82) p-Isopropyltoluene	13.706	119	363013	19.061	ug/l	99
83) 1,3-Dichlorobenzene	13.712	146	181191	18.806	ug/l	99
84) 1,4-Dichlorobenzene	13.794	146	187318	19.177	ug/l	99
85) n-Butylbenzene	14.035	91	367098	19.701	ug/l	98
86) Hexachloroethane	14.306	117	70328	18.910	ug/l	99
87) 1,2-Dichlorobenzene	14.082	146	176099	19.148	ug/l	99
88) 1,2-Dibromo-3-Chloropr...	14.700	75	35483	18.730	ug/l	96
89) 1,2,4-Trichlorobenzene	15.811	180	103561	18.842	ug/l	99
90) Hexachlorobutadiene	15.476	225	40240	18.828	ug/l	97
91) Naphthalene	15.617	128	396205	18.905	ug/l	99
92) 1,2,3-Trichlorobenzene	15.811	180	103561	18.842	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087459.D
 Acq On : 05 Aug 2025 16:17
 Operator : JC\MD
 Sample : VN0805WBS02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0805WBS02

Manual Integrations APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 02:41:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration



Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	08/04/25
Project:	Transfer Station-SPDES	Date Received:	08/05/25
Client Sample ID:	001-WILLETS-PT-BLVD(JUNE)MS	SDG No.:	Q2771
Lab Sample ID:	Q2771-02MS	Matrix:	Water
Analytical Method:	E624.1	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-BTEX
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087464.D	1	08/05/25 18:35	VN080525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
71-43-2	Benzene	20.5		0.45	5.00	ug/L
108-88-3	Toluene	26.1		0.46	5.00	ug/L
100-41-4	Ethyl Benzene	20.4		0.56	5.00	ug/L
179601-23-1	m/p-Xylenes	40.7		1.30	10.0	ug/L
95-47-6	o-Xylene	20.3		0.67	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	30.9		91 - 110	103%	SPK: 30
2037-26-5	Toluene-d8	29.9		91 - 112	100%	SPK: 30
460-00-4	4-Bromofluorobenzene	29.4		63 - 112	98%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	57900	7.8			
540-36-3	1,4-Difluorobenzene	327000	9.082			
3114-55-4	Chlorobenzene-d5	293000	11.847			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087464.D
 Acq On : 05 Aug 2025 18:35
 Operator : JC\MD
 Sample : Q2771-02MS
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-WILLETS-PT-BLVD(JUNE)MS

Quant Time: Aug 06 02:44:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.800	128	57858	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.082	114	327425	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	292747	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.559	65	165120	30.941	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery = 103.133%			
60) 4-Bromofluorobenzene	12.829	95	147984	29.374	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery = 97.900%			
63) Toluene-d8	10.547	98	402285	29.865	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery = 99.567%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.142	85	86069	22.784	ug/l	91
3) Chloromethane	2.383	50	84289	21.748	ug/l	94
4) Vinyl Chloride	2.542	62	89963	21.885	ug/l	99
5) Bromomethane	2.983	94	40650	18.270	ug/l	93
6) Chloroethane	3.136	64	57248	21.597	ug/l	99
7) Trichlorofluoromethane	3.512	101	128573	21.733	ug/l	97
8) Diethyl Ether	3.965	74	52589	21.337	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.371	101	65737	21.790	ug/l	88
10) 1,1-Dichloroethene	4.336	96	68672	21.671	ug/l	91
11) Methyl Iodide	4.577	142	39694	15.909	ug/l	95
12) Methyl Acetate	5.018	43	121879	20.842	ug/l	94
13) Acrolein	4.177	56	83429	93.411	ug/l	99
14) Acrylonitrile	5.712	53	270925	106.100	ug/l	98
15) Acetone	4.424	58	78660	107.813	ug/l	93
16) Carbon Disulfide	4.700	76	206068	20.964	ug/l	99
17) Allyl chloride	5.006	41	129694	20.080	ug/l	92
18) Methylene Chloride	5.265	84	85179	20.880	ug/l	93
19) trans-1,2-Dichloroethene	5.771	96	75693	20.087	ug/l	95
20) Diisopropyl ether	6.665	45	299448	20.352	ug/l	97
21) 1,1-Dichloroethane	6.553	63	156845	20.622	ug/l	99
22) cis-1,2-Dichloroethene	7.471	96	96912	20.902	ug/l	98
23) tert-Butyl Alcohol	5.518	59	115151	106.126	ug/l #	100
24) Methyl tert-Butyl Ether	5.794	73	302253	20.933	ug/l	99
25) Chloroform	7.947	83	296504	37.348	ug/l	95
26) Cyclohexane	8.235	56	132657	21.785	ug/l #	98
29) 1,1-Dichloropropene	8.353	75	110179	20.631	ug/l	99
30) 2-Butanone	7.471	43	408485	103.770	ug/l	100
31) 2,2-Dichloropropane	7.477	77	135864	19.726	ug/l	99
32) 1,1,1-Trichloroethane	8.153	97	143139	21.016	ug/l	97
33) Carbon Tetrachloride	8.347	117	120006	21.209	ug/l	100
34) Benzene	8.588	78	346136	20.533	ug/l	99
35) Methacrylonitrile	7.759	41	88788	20.902	ug/l	95
36) 1,2-Dichloroethane	8.653	62	137652	20.203	ug/l	100
37) Trichloroethene	9.335	130	76827	20.689	ug/l	96
38) Methylcyclohexane	9.582	83	130586	21.078	ug/l	98
39) 1,2-Dichloropropane	9.606	63	89042	20.601	ug/l	99
40) Dibromomethane	9.694	93	65247	20.120	ug/l	97
41) Bromodichloromethane	9.871	83	145959	21.660	ug/l	97
42) Vinyl Acetate	6.594	43	1300901	100.352	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087464.D
 Acq On : 05 Aug 2025 18:35
 Operator : JC\MD
 Sample : Q2771-02MS
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-WILLETS-PT-BLVD(JUNE)MS

Manual Integrations
 APPROVED

Reviewed By :John Carlone 08/06/2025
 Supervised By :Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 02:44:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	7.553	43	158819	20.884	ug/l	99
44) Isopropyl Acetate	8.677	43	259623	20.140	ug/l	99
45) 1,4-Dioxane	9.682	88	35383	430.561	ug/l #	89
46) Methyl methacrylate	9.665	41	134321	22.074	ug/l	94
47) n-amyl Acetate	12.523	43	159267m	22.365	ug/l	
48) t-1,3-Dichloropropene	10.818	75	146464	19.978	ug/l	96
49) cis-1,3-Dichloropropene	10.294	75	146576	19.867	ug/l	95
50) 1,1,2-Trichloroethane	11.000	97	82991	20.255	ug/l	97
51) Ethyl methacrylate	10.853	69	149180	20.674	ug/l	95
52) 1,3-Dichloropropane	11.141	76	147593	19.994	ug/l	98
53) Dibromochloromethane	11.341	129	99264	20.978	ug/l	98
54) 1,2-Dibromoethane	11.447	107	86868	19.972	ug/l	99
56) Bromoform	12.559	173	63211	20.197	ug/l #	99
58) 4-Methyl-2-Pentanone	10.424	43	811173	107.714	ug/l	100
59) 2-Hexanone	11.176	43	526710	102.937	ug/l	100
61) Tetrachloroethene	11.082	164	60085	21.067	ug/l	98
62) Toluene	10.612	91	471528	26.097	ug/l	99
64) Chlorobenzene	11.871	112	226643	20.414	ug/l	100
65) 1,1,1,2-Tetrachloroethane	11.941	131	78931	20.221	ug/l	99
66) Ethyl Benzene	11.941	91	410709	20.394	ug/l	100
67) m/p-Xylenes	12.053	106	301085	40.730	ug/l	99
68) o-Xylene	12.376	106	147374	20.265	ug/l	100
69) Styrene	12.388	104	246611	19.860	ug/l	98
70) Isopropylbenzene	12.670	105	384456	20.808	ug/l	100
71) 1,1,2,2-Tetrachloroethane	12.918	83	132693	20.961	ug/l	100
72) 1,2,3-Trichloropropane	12.970	75	125987m	21.363	ug/l	
73) Bromobenzene	12.959	156	87403	20.406	ug/l	99
74) n-propylbenzene	13.012	91	479069	20.762	ug/l	99
75) 2-Chlorotoluene	13.106	91	285626	20.123	ug/l	100
76) 1,3,5-Trimethylbenzene	13.153	105	323833	20.470	ug/l	100
77) t-1,4-Dichloro-2-butene	12.718	75	40864	18.304	ug/l	95
78) 4-Chlorotoluene	13.200	91	298709	20.767	ug/l	98
79) tert-butylbenzene	13.417	119	278633	20.890	ug/l	99
80) 1,2,4-Trimethylbenzene	13.459	105	322979	20.432	ug/l	100
81) sec-Butylbenzene	13.594	105	407509	20.987	ug/l	99
82) p-Isopropyltoluene	13.706	119	337514	20.935	ug/l	99
83) 1,3-Dichlorobenzene	13.712	146	165618	20.306	ug/l	97
84) 1,4-Dichlorobenzene	13.794	146	166172	20.096	ug/l	97
85) n-Butylbenzene	14.035	91	331884	21.040	ug/l	98
86) Hexachloroethane	14.312	117	61684	19.593	ug/l	98
87) 1,2-Dichlorobenzene	14.082	146	158893	20.409	ug/l	98
88) 1,2-Dibromo-3-Chloropr...	14.700	75	33483	20.878	ug/l	97
89) 1,2,4-Trichlorobenzene	15.811	180	93743	20.148	ug/l	99
90) Hexachlorobutadiene	15.476	225	39066	21.592	ug/l	97
91) Naphthalene	15.611	128	362681	20.443	ug/l	100
92) 1,2,3-Trichlorobenzene	15.811	180	93743	20.148	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087464.D
 Acq On : 05 Aug 2025 18:35
 Operator : JC\MD
 Sample : Q2771-02MS
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-WILLETS-PT-BLVD(JUNE)MS

Quant Time: Aug 06 02:44:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :John Carlone 08/06/2025
 Supervised By :Mahesh Dadoda 08/06/2025



Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	08/04/25
Project:	Transfer Station-SPDES	Date Received:	08/05/25
Client Sample ID:	001-WILLETS-PT-BLVD(JUNE)MSD	SDG No.:	Q2771
Lab Sample ID:	Q2771-03MSD	Matrix:	Water
Analytical Method:	E624.1	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-BTEX
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087465.D	1	08/05/25 18:56	VN080525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
71-43-2	Benzene	20.3		0.45	5.00	ug/L
108-88-3	Toluene	26.0		0.46	5.00	ug/L
100-41-4	Ethyl Benzene	20.6		0.56	5.00	ug/L
179601-23-1	m/p-Xylenes	40.9		1.30	10.0	ug/L
95-47-6	o-Xylene	20.2		0.67	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	31.7		91 - 110	106%	SPK: 30
2037-26-5	Toluene-d8	30.3		91 - 112	101%	SPK: 30
460-00-4	4-Bromofluorobenzene	29.7		63 - 112	99%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	63400	7.794			
540-36-3	1,4-Difluorobenzene	362000	9.082			
3114-55-4	Chlorobenzene-d5	322000	11.847			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087465.D
 Acq On : 05 Aug 2025 18:56
 Operator : JC\MD
 Sample : Q2771-03MSD
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-WILLETS-PT-BLVD(JUNE)MSD

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 02:45:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.794	128	63371	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.082	114	362030	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	321655	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.559	65	185003	31.651	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery = 105.500%			
60) 4-Bromofluorobenzene	12.829	95	164537	29.724	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery = 99.067%			
63) Toluene-d8	10.547	98	448635	30.313	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery = 101.033%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.142	85	93645	22.633	ug/l	95
3) Chloromethane	2.383	50	94056	22.157	ug/l	98
4) Vinyl Chloride	2.536	62	99195	22.032	ug/l	95
5) Bromomethane	2.983	94	49116	20.154	ug/l	86
6) Chloroethane	3.142	64	62863	21.652	ug/l	95
7) Trichlorofluoromethane	3.512	101	144212	22.256	ug/l	98
8) Diethyl Ether	3.959	74	57040	21.130	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.365	101	75321	22.795	ug/l	86
10) 1,1-Dichloroethene	4.336	96	75689	21.807	ug/l	97
11) Methyl Iodide	4.583	142	51662	18.904	ug/l	92
12) Methyl Acetate	5.018	43	132767	20.729	ug/l	93
13) Acrolein	4.183	56	95012	97.125	ug/l	97
14) Acrylonitrile	5.712	53	292494	104.582	ug/l	100
15) Acetone	4.424	58	84788	106.103	ug/l	89
16) Carbon Disulfide	4.706	76	227080	21.092	ug/l	98
17) Allyl chloride	5.018	41	149063	21.071	ug/l	88
18) Methylene Chloride	5.265	84	94609	21.174	ug/l	93
19) trans-1,2-Dichloroethene	5.771	96	85295	20.666	ug/l	95
20) Diisopropyl ether	6.659	45	335083	20.792	ug/l	98
21) 1,1-Dichloroethane	6.553	63	174555	20.954	ug/l	98
22) cis-1,2-Dichloroethene	7.477	96	104639	20.605	ug/l	97
23) tert-Butyl Alcohol	5.524	59	121535	102.265	ug/l #	100
24) Methyl tert-Butyl Ether	5.788	73	339285	21.453	ug/l	99
25) Chloroform	7.953	83	318668	36.648	ug/l	99
26) Cyclohexane	8.235	56	150258	22.528	ug/l #	100
29) 1,1-Dichloropropene	8.353	75	124594	21.101	ug/l	96
30) 2-Butanone	7.471	43	452029	103.856	ug/l #	96
31) 2,2-Dichloropropane	7.471	77	150697	19.788	ug/l	98
32) 1,1,1-Trichloroethane	8.153	97	159868	21.228	ug/l	95
33) Carbon Tetrachloride	8.347	117	132303	21.147	ug/l	95
34) Benzene	8.588	78	379279	20.348	ug/l	98
35) Methacrylonitrile	7.765	41	98835	21.043	ug/l	94
36) 1,2-Dichloroethane	8.653	62	158507	21.040	ug/l	99
37) Trichloroethene	9.329	130	83242	20.274	ug/l	95
38) Methylcyclohexane	9.582	83	144189	21.049	ug/l	98
39) 1,2-Dichloropropane	9.600	63	100476	21.024	ug/l	94
40) Dibromomethane	9.694	93	74131	20.674	ug/l	98
41) Bromodichloromethane	9.871	83	161315	21.651	ug/l	97
42) Vinyl Acetate	6.588	43	1562358	109.001	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087465.D
 Acq On : 05 Aug 2025 18:56
 Operator : JC\MD
 Sample : Q2771-03MSD
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-WILLETS-PT-BLVD(JUNE)MSD

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 02:45:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	7.547	43	171544m	20.401	ug/l	
44) Isopropyl Acetate	8.677	43	286640	20.110	ug/l	99
45) 1,4-Dioxane	9.676	88	37991	418.108	ug/l #	87
46) Methyl methacrylate	9.665	41	127427	18.940	ug/l	92
47) n-amyl Acetate	12.517	43	180598m	22.937	ug/l	
48) t-1,3-Dichloropropene	10.818	75	162195	20.009	ug/l	97
49) cis-1,3-Dichloropropene	10.294	75	160379	19.660	ug/l	96
50) 1,1,2-Trichloroethane	11.000	97	93732	20.690	ug/l	96
51) Ethyl methacrylate	10.859	69	163762	20.525	ug/l	98
52) 1,3-Dichloropropane	11.141	76	163568	20.041	ug/l	96
53) Dibromochloromethane	11.335	129	107106	20.472	ug/l	99
54) 1,2-Dibromoethane	11.447	107	96878	20.144	ug/l	99
56) Bromoform	12.559	173	68070	19.670	ug/l #	99
58) 4-Methyl-2-Pentanone	10.423	43	892848	107.904	ug/l	100
59) 2-Hexanone	11.176	43	596944	106.178	ug/l	100
61) Tetrachloroethene	11.082	164	65268	20.828	ug/l	96
62) Toluene	10.612	91	515298	25.956	ug/l	99
64) Chlorobenzene	11.870	112	250942	20.571	ug/l	99
65) 1,1,1,2-Tetrachloroethane	11.941	131	89553	20.880	ug/l	98
66) Ethyl Benzene	11.941	91	455815	20.600	ug/l	100
67) m/p-Xylenes	12.053	106	332592	40.949	ug/l	97
68) o-Xylene	12.376	106	161451	20.205	ug/l	99
69) Styrene	12.388	104	273254	20.028	ug/l	98
70) Isopropylbenzene	12.676	105	426125	20.991	ug/l	100
71) 1,1,2,2-Tetrachloroethane	12.917	83	148411	21.337	ug/l	98
72) 1,2,3-Trichloropropane	12.976	75	144354m	22.278	ug/l	
73) Bromobenzene	12.959	156	95153	20.219	ug/l	99
74) n-propylbenzene	13.012	91	523707	20.656	ug/l	100
75) 2-Chlorotoluene	13.106	91	316527	20.296	ug/l	99
76) 1,3,5-Trimethylbenzene	13.153	105	351735	20.236	ug/l	100
77) t-1,4-Dichloro-2-butene	12.717	75	42832	17.462	ug/l	95
78) 4-Chlorotoluene	13.200	91	330308	20.900	ug/l	97
79) tert-butylbenzene	13.417	119	303771	20.728	ug/l	97
80) 1,2,4-Trimethylbenzene	13.459	105	357826	20.602	ug/l	99
81) sec-Butylbenzene	13.594	105	449832	21.085	ug/l	100
82) p-Isopropyltoluene	13.706	119	365437	20.630	ug/l	99
83) 1,3-Dichlorobenzene	13.711	146	183928	20.524	ug/l	98
84) 1,4-Dichlorobenzene	13.788	146	184810	20.341	ug/l	97
85) n-Butylbenzene	14.035	91	363506	20.974	ug/l	99
86) Hexachloroethane	14.306	117	69483	20.086	ug/l	100
87) 1,2-Dichlorobenzene	14.082	146	178997	20.925	ug/l	99
88) 1,2-Dibromo-3-Chloropr...	14.694	75	36710	20.833	ug/l	99
89) 1,2,4-Trichlorobenzene	15.811	180	102214	19.994	ug/l	98
90) Hexachlorobutadiene	15.476	225	39501	19.871	ug/l	96
91) Naphthalene	15.611	128	393425	20.183	ug/l	99
92) 1,2,3-Trichlorobenzene	15.811	180	102214	19.994	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087465.D
 Acq On : 05 Aug 2025 18:56
 Operator : JC\MD
 Sample : Q2771-03MSD
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-WILLETS-PT-BLVD(JUNE)MSD

Manual Integrations APPROVED

Reviewed By : John Carlone 08/06/2025
 Supervised By : Mahesh Dadoda 08/06/2025

Quant Time: Aug 06 02:45:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration



Manual Integration Report

Sequence:	vn080525	Instrument	MSVOA_n
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC005	VN087452.D	1,2,3-Trichloropropane	JOHN	8/6/2025 8:44:05 AM	MMDadoda	8/6/2025 9:39:20 AM	Peak Integrated by Software
VSTDICC005	VN087452.D	Ethyl Acetate	JOHN	8/6/2025 8:44:05 AM	MMDadoda	8/6/2025 9:39:20 AM	Peak Integrated by Software
VSTDICC005	VN087452.D	Methyl Iodide	JOHN	8/6/2025 8:44:05 AM	MMDadoda	8/6/2025 9:39:20 AM	Peak Integrated by Software
VSTDICC005	VN087452.D	n-amyl Acetate	JOHN	8/6/2025 8:44:05 AM	MMDadoda	8/6/2025 9:39:20 AM	Peak Integrated by Software
VSTDICC005	VN087452.D	trans-1,2-Dichloroethene	JOHN	8/6/2025 8:44:05 AM	MMDadoda	8/6/2025 9:39:20 AM	Peak Integrated by Software
VSTDICCC020	VN087453.D	1,2,3-Trichloropropane	JOHN	8/6/2025 8:44:09 AM	MMDadoda	8/6/2025 9:39:23 AM	Peak Integrated by Software
VSTDICCC020	VN087453.D	Ethyl Acetate	JOHN	8/6/2025 8:44:09 AM	MMDadoda	8/6/2025 9:39:23 AM	Peak Integrated by Software
VSTDICCC020	VN087453.D	Methyl Iodide	JOHN	8/6/2025 8:44:09 AM	MMDadoda	8/6/2025 9:39:23 AM	Peak Integrated by Software
VSTDICCC020	VN087453.D	Methylene Chloride	JOHN	8/6/2025 8:44:09 AM	MMDadoda	8/6/2025 9:39:23 AM	Peak Integrated by Software
VSTDICCC020	VN087453.D	n-amyl Acetate	JOHN	8/6/2025 8:44:09 AM	MMDadoda	8/6/2025 9:39:23 AM	Peak Integrated by Software
VSTDICCC050	VN087454.D	1,2,3-Trichloropropane	JOHN	8/6/2025 8:44:15 AM	MMDadoda	8/6/2025 9:39:27 AM	Peak Integrated by Software
VSTDICCC050	VN087454.D	n-amyl Acetate	JOHN	8/6/2025 8:44:15 AM	MMDadoda	8/6/2025 9:39:27 AM	Peak Integrated by Software
VSTDICCC100	VN087455.D	1,2,3-Trichloropropane	JOHN	8/6/2025 8:44:19 AM	MMDadoda	8/6/2025 9:39:30 AM	Peak Integrated by Software

Manual Integration Report

Sequence:	vn080525	Instrument	MSVOA_n
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC150	VN087456.D	1,2,3-Trichloropropane	JOHN	8/6/2025 8:44:24 AM	MMDadoda	8/6/2025 9:39:31 AM	Peak Integrated by Software
VSTDICV020	VN087458.D	1,2,3-Trichloropropane	JOHN	8/6/2025 8:44:28 AM	MMDadoda	8/6/2025 9:39:41 AM	Peak Integrated by Software
VSTDICV020	VN087458.D	n-amyl Acetate	JOHN	8/6/2025 8:44:28 AM	MMDadoda	8/6/2025 9:39:41 AM	Peak Integrated by Software
VN0805WBS02	VN087459.D	1,2,3-Trichloropropane	JOHN	8/6/2025 8:44:35 AM	MMDadoda	8/6/2025 9:39:43 AM	Peak Integrated by Software
VN0805WBS02	VN087459.D	n-amyl Acetate	JOHN	8/6/2025 8:44:35 AM	MMDadoda	8/6/2025 9:39:43 AM	Peak Integrated by Software
Q2771-01	VN087463.D	Acetone	JOHN	8/6/2025 8:44:54 AM	MMDadoda	8/6/2025 9:39:47 AM	Peak Integrated by Software
Q2771-02MS	VN087464.D	1,2,3-Trichloropropane	JOHN	8/6/2025 8:45:00 AM	MMDadoda	8/6/2025 9:39:48 AM	Peak Integrated by Software
Q2771-02MS	VN087464.D	n-amyl Acetate	JOHN	8/6/2025 8:45:00 AM	MMDadoda	8/6/2025 9:39:48 AM	Peak Integrated by Software
Q2771-03MSD	VN087465.D	1,2,3-Trichloropropane	JOHN	8/6/2025 8:45:05 AM	MMDadoda	8/6/2025 9:39:50 AM	Peak Integrated by Software
Q2771-03MSD	VN087465.D	Ethyl Acetate	JOHN	8/6/2025 8:45:05 AM	MMDadoda	8/6/2025 9:39:50 AM	Peak Integrated by Software
Q2771-03MSD	VN087465.D	n-amyl Acetate	JOHN	8/6/2025 8:45:05 AM	MMDadoda	8/6/2025 9:39:50 AM	Peak Integrated by Software
VSTDCCC020	VN087466.D	1,2,3-Trichloropropane	JOHN	8/6/2025 8:45:09 AM	MMDadoda	8/6/2025 9:39:51 AM	Peak Integrated by Software
VSTDCCC020	VN087466.D	Ethyl Acetate	JOHN	8/6/2025 8:45:09 AM	MMDadoda	8/6/2025 9:39:51 AM	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	vn080525	Instrument	MSVOA_n
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC020	VN087466.D	n-amyl Acetate	JOHN	8/6/2025 8:45:09 AM	MMDadoda	8/6/2025 9:39:51 AM	Peak Integrated by Software

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN080525

Review By	John Carlone	Review On	8/6/2025 8:45:40 AM
Supervise By	Mahesh Dadoda	Supervise On	8/6/2025 9:40:10 AM
SubDirectory	VN080525	HP Acquire Method	HP Processing Method 624N080525W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP134999		
Initial Calibration Stds	VP135004,VP135005,VP135006,VP135007,VP135008		
CCC	VP135003,VP135010		
Internal Standard/PEM			
ICV/I.BLK	VP135009		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN087447.D	05 Aug 2025 09:27	JC\MD	Ok
2	VSTDCCC020	VN087448.D	05 Aug 2025 10:06	JC\MD	Not Ok
3	VN0805WBS01	VN087449.D	05 Aug 2025 11:21	JC\MD	Not Ok
4	VN0805WBSD01	VN087450.D	05 Aug 2025 12:02	JC\MD	Not Ok
5	VN0805WBL01	VN087451.D	05 Aug 2025 12:58	JC\MD	Not Ok
6	VSTDICC005	VN087452.D	05 Aug 2025 13:37	JC\MD	Ok,M
7	VSTDICCC020	VN087453.D	05 Aug 2025 13:58	JC\MD	Ok,M
8	VSTDICC050	VN087454.D	05 Aug 2025 14:19	JC\MD	Ok,M
9	VSTDICC100	VN087455.D	05 Aug 2025 14:40	JC\MD	Ok,M
10	VSTDICC150	VN087456.D	05 Aug 2025 15:01	JC\MD	Ok,M
11	IBLK	VN087457.D	05 Aug 2025 15:22	JC\MD	Ok
12	VSTDICV020	VN087458.D	05 Aug 2025 15:43	JC\MD	Ok,M
13	VN0805WBS02	VN087459.D	05 Aug 2025 16:17	JC\MD	Ok,M
14	VN0805WBSD02	VN087460.D	05 Aug 2025 16:50	JC\MD	Ok,M
15	VN0805WBL02	VN087461.D	05 Aug 2025 17:32	JC\MD	Ok
16	Q2771-04	VN087462.D	05 Aug 2025 17:53	JC\MD	Ok
17	Q2771-01	VN087463.D	05 Aug 2025 18:14	JC\MD	Ok,M
18	Q2771-02MS	VN087464.D	05 Aug 2025 18:35	JC\MD	Ok,M
19	Q2771-03MSD	VN087465.D	05 Aug 2025 18:56	JC\MD	Ok,M
20	VSTDCCC020	VN087466.D	05 Aug 2025 19:17	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN080525

Review By	John Carlone	Review On	8/6/2025 8:45:40 AM
Supervise By	Maresh Dadoda	Supervise On	8/6/2025 9:40:10 AM
SubDirectory	VN080525	HP Acquire Method	HP Processing Method 624N080525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134999
Initial Calibration Stds	VP135004,VP135005,VP135006,VP135007,VP135008
CCC	VP135003,VP135010
Internal Standard/PEM	
ICV/I.BLK	VP135009
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087447.D	05 Aug 2025 09:27		JC\MD	Ok
2	VSTDCCC020	VSTDCCC020	VN087448.D	05 Aug 2025 10:06	Need ICAL	JC\MD	Not Ok
3	VN0805WBS01	VN0805WBS01	VN087449.D	05 Aug 2025 11:21		JC\MD	Not Ok
4	VN0805WBSD01	VN0805WBSD01	VN087450.D	05 Aug 2025 12:02		JC\MD	Not Ok
5	VN0805WBL01	VN0805WBL01	VN087451.D	05 Aug 2025 12:58		JC\MD	Not Ok
6	VSTDICC005	VSTDICC005	VN087452.D	05 Aug 2025 13:37		JC\MD	Ok,M
7	VSTDICCC020	VSTDICCC020	VN087453.D	05 Aug 2025 13:58		JC\MD	Ok,M
8	VSTDICC050	VSTDICC050	VN087454.D	05 Aug 2025 14:19		JC\MD	Ok,M
9	VSTDICC100	VSTDICC100	VN087455.D	05 Aug 2025 14:40		JC\MD	Ok,M
10	VSTDICC150	VSTDICC150	VN087456.D	05 Aug 2025 15:01		JC\MD	Ok,M
11	IBLK	IBLK	VN087457.D	05 Aug 2025 15:22		JC\MD	Ok
12	VSTDICV020	ICVVN080525	VN087458.D	05 Aug 2025 15:43	pH#Lot#V12668	JC\MD	Ok,M
13	VN0805WBS02	VN0805WBS02	VN087459.D	05 Aug 2025 16:17		JC\MD	Ok,M
14	VN0805WBSD02	VN0805WBSD02	VN087460.D	05 Aug 2025 16:50		JC\MD	Ok,M
15	VN0805WBL02	VN0805WBL02	VN087461.D	05 Aug 2025 17:32		JC\MD	Ok
16	Q2771-04	002-35TH-AVE(JUNE)	VN087462.D	05 Aug 2025 17:53	vial A pH<2	JC\MD	Ok
17	Q2771-01	001-WILLETS-PT-BLVD	VN087463.D	05 Aug 2025 18:14	vial A pH<2	JC\MD	Ok,M
18	Q2771-02MS	001-WILLETS-PT-BLVD	VN087464.D	05 Aug 2025 18:35	vial A pH<2	JC\MD	Ok,M

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN080525

Review By	John Carlone	Review On	8/6/2025 8:45:40 AM
Supervise By	Maresh Dadoda	Supervise On	8/6/2025 9:40:10 AM
SubDirectory	VN080525	HP Acquire Method	HP Processing Method 624N080525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134999
Initial Calibration Stds	VP135004,VP135005,VP135006,VP135007,VP135008
CCC	VP135003,VP135010
Internal Standard/PEM	
ICV/I.BLK	VP135009
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	Q2771-03MSD	001-WILLETS-PT-BLVD	VN087465.D	05 Aug 2025 18:56	vial A pH<2	JC\MD	Ok,M
20	VSTDCCC020	VSTDCCC020EC	VN087466.D	05 Aug 2025 19:17		JC\MD	Ok,M

M : Manual Integration



SHIPPING DOCUMENTS



284 Sheffield Street, Mountainside, NJ 07092

(908) 789-8900 Fax: (908) 788-9222

www.chemtech.net

CHAIN OF CUSTODY RECORD

Alliance Project Number:

Q2771

COC Number:

CLIENT INFORMATION

COMPANY: Tully Environmental Inc.

ADDRESS: 57 Seaview Blvd

CITY: Pt Washington STATE: NY ZIP: 11050

ATTENTION: Dean Devoe

PHONE: 718 446 7000

FAX:

PROJECT INFORMATION

PROJECT NAME: Transfer Station SPDES

PROJECT #: 252113

LOCATION:

PROJECT MANAGER:

E-MAIL:

PHONE:

FAX:

BILLING INFORMATION

BILL TO: Same

PO#

ADDRESS:

CITY:

STATE: ZIP:

ATTENTION:

PHONE:

DATA TURNAROUND INFORMATION

FAX: _____ DAYS*

HARD COPY: _____ DAYS*

EDD _____ DAYS*

* TO BE APPROVED BY ALLIANCE

STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

* RESULTS ONLY

☐ RESULTS + QC☐ New Jersey REDUCED☐ New Jersey CLP☐ EDD Format _____☐ USEPA CLP☐ New York State ASP "B"☐ New York State ASP "A"☐ Other _____

ANALYSIS

BTX

TSS

1

2

3

4

5

6

7

8

9

PRESERVATIVES

COMMENTS

<- Specify Preservatives

A-HCl B-HNO3

C-H2SO4 D-NaOH

E-ICE F-Other

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles										
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	001 Willets Pt Blvd (June)	W		X	8/4/25	11:30	3	X	X								
2.	002 35th Ave (June)	W		X	8/4/25	11:30	3	X	X								
3.																	
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER	DATE/TIME Aug 4, 2025	RECEIVED BY	1.	Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant ☐ Cooler Temp <u>15.5°C</u> MeOH extraction requires an additional 4oz. Jar for percent solid Comments:	SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight ALLIANCE: ☐ Picked Up ☐ Overnight	Shipment Complete ☐ YES ☐ NO
RELINQUISHED BY	DATE/TIME 8-5-25 1130	RECEIVED BY	2.			
RELINQUISHED BY	DATE/TIME	RECEIVED FOR LAB BY	3.			

Page _____ of _____

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT

YELLOW - ALLIANCE COPY

PINK - SAMPLER COPY

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2771	TULL01	Order Date : 8/5/2025 11:44:00 AM	Project Mgr :
Client Name : Tully Environmental, Inc		Project Name : Transfer Station-SPDES	Report Type : Results Only
Client Contact : Dean Devoe		Receive DateTime : 8/5/2025 12:00:00 AM	EDD Type : EXCEL NOCLEANUP
Invoice Name : Tully Environmental, Inc		Purchase Order : 11:30:00	Hard Copy Date :
Invoice Contact : Dean Devoe			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2771-01	001-WILLETS-PT-BLVD(JUNE)	Water	08/05/2025 08/04/2025	11:30	VOC-BTEX		624.1	1 Bus. Day	
Q2771-02	Q2771-01MS	Water	08/05/2025 08/04/2025	11:30	VOC-BTEX		624.1	1 Bus. Day	
Q2771-03	Q2771-01MSD	Water	08/05/2025 08/04/2025	11:30	VOC-BTEX		624.1	1 Bus. Day	
Q2771-04	002-35TH-AVE(JUNE)	Water	08/05/2025 08/04/2025	11:30	VOC-BTEX		624.1	1 Bus. Day	

Relinquished By :

Date / Time : 8/5/25 12:00

Received By :

Date / Time : 08/05/25

Storage Area : VOA Refridgerator Room

12:22