

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TULL01
 Lab Code: ACE SDG NO.: Q2771
 Lab File ID: VN087453.D Date Analyzed: 08/05/2025
 Instrument ID: MSVOA_N Time Analyzed: 13:58
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	64218	7.80	367115	9.08	327914	11.85
UPPER LIMIT	128436	8.3	734230	9.582	655828	12.347
LOWER LIMIT	32109	7.3	183558	8.582	163957	11.347
EPA SAMPLE NO.						
001-WILLETS-PT-BLVD (JUNE)	54767	7.79	306638	9.08	270091	11.85
001-WILLETS-PT-BLVD (JUNE)MS	57858	7.80	327425	9.08	292747	11.85
001-WILLETS-PT-BLVD (JUNE)MSD	63371	7.79	362030	9.08	321655	11.85
002-35TH-AVE (JUNE)	63498	7.80	368676	9.08	324794	11.85
VN0805WBL02	68205	7.80	376206	9.08	334367	11.85
VN0805WBS02	69933	7.80	387635	9.08	345816	11.85

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.