

Initial and Continuing Calibration Verification

Client: American Wax

SDG No.: Q2778

Project: Semi Annual 2025

RunNo.: LB136758

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	50.962	50	102	95-105	05/28/2025
Sample ID: CCV1 COD	mg/L	48.931	50	98	95-105	08/08/2025
Sample ID: CCV2 COD	mg/L	50.962	50	102	95-105	08/08/2025