

Initial and Continuing Calibration Verification

Client: American Wax

SDG No.: Q2787

Project: Semi Annual 2025

RunNo.: LB136758

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: COD	ICV	mg/L	50.962	50	102	95-105	05/28/2025
Sample ID: COD	CCV1	mg/L	48.931	50	98	95-105	08/08/2025
Sample ID: COD	CCV2	mg/L	50.962	50	102	95-105	08/08/2025

Initial and Continuing Calibration Verification

Client: American Wax

SDG No.: Q2787

Project: Semi Annual 2025

RunNo.: LB136758

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------