

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2789

Project: Monthly 2025

Run No.: LB136731

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136731BS							
TPH	mg/L	20.0	17.0		85	1	78-114	08/07/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2789

Project: Monthly 2025

Run No.: LB136731

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136731BSD							
TPH	mg/L	20.0	17.2		86	1	78-114	08/07/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2789

Project: Monthly 2025

Run No.: LB136753

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136753BS							
BOD5	mg/L	198	202		102	1	84.6-115.4	08/08/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2789

Project: Monthly 2025

Run No.: LB136765

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136765BS							
TSS	mg/L	550	531		96	1	90-110	08/11/2025