

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q2789

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV11	Mercury	4.06	4.0	102	95 - 105	CV	08/12/2025	17:35	LB136791

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q2789

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV32	Mercury	4.97	5.0	99	90 - 110	CV	08/12/2025	17:42	LB136791
CCV33	Mercury	4.66	5.0	93	90 - 110	CV	08/12/2025	18:18	LB136791
CCV34	Mercury	4.75	5.0	95	90 - 110	CV	08/12/2025	18:30	LB136791

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q2789

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cadmium	1900	2000	95	95 - 105	P	08/12/2025	11:34	LB136789
	Copper	999	1000	100	95 - 105	P	08/12/2025	11:34	LB136789
	Lead	4050	4000	101	95 - 105	P	08/12/2025	11:34	LB136789
	Nickel	1950	2000	98	95 - 105	P	08/12/2025	11:34	LB136789
	Zinc	1950	2000	97	95 - 105	P	08/12/2025	11:34	LB136789

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q2789

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cadmium	5.45	6.0	91	80 - 120	P	08/12/2025	11:43	LB136789
	Copper	21.2	20.0	106	80 - 120	P	08/12/2025	11:43	LB136789
	Lead	10.9	12.0	91	80 - 120	P	08/12/2025	11:43	LB136789
	Nickel	39.0	40.0	97	80 - 120	P	08/12/2025	11:43	LB136789
	Zinc	40.5	40.0	101	80 - 120	P	08/12/2025	11:43	LB136789

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q2789

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cadmium	2380	2500	95	90 - 110	P	08/12/2025	12:50	LB136789
	Copper	1220	1250	98	90 - 110	P	08/12/2025	12:50	LB136789
	Lead	4880	5000	98	90 - 110	P	08/12/2025	12:50	LB136789
	Nickel	2400	2500	96	90 - 110	P	08/12/2025	12:50	LB136789
	Zinc	2480	2500	99	90 - 110	P	08/12/2025	12:50	LB136789
CCV02	Cadmium	2460	2500	98	90 - 110	P	08/12/2025	14:27	LB136789
	Copper	1270	1250	102	90 - 110	P	08/12/2025	14:27	LB136789
	Lead	4880	5000	98	90 - 110	P	08/12/2025	14:27	LB136789
	Nickel	2480	2500	99	90 - 110	P	08/12/2025	14:27	LB136789
	Zinc	2560	2500	102	90 - 110	P	08/12/2025	14:27	LB136789
CCV03	Cadmium	2430	2500	97	90 - 110	P	08/12/2025	16:07	LB136789
	Copper	1260	1250	101	90 - 110	P	08/12/2025	16:07	LB136789
	Lead	4820	5000	96	90 - 110	P	08/12/2025	16:07	LB136789
	Nickel	2450	2500	98	90 - 110	P	08/12/2025	16:07	LB136789
	Zinc	2520	2500	101	90 - 110	P	08/12/2025	16:07	LB136789
CCV04	Cadmium	2460	2500	98	90 - 110	P	08/12/2025	16:30	LB136789
	Copper	1280	1250	103	90 - 110	P	08/12/2025	16:30	LB136789
	Lead	4870	5000	97	90 - 110	P	08/12/2025	16:30	LB136789
	Nickel	2490	2500	100	90 - 110	P	08/12/2025	16:30	LB136789
	Zinc	2500	2500	100	90 - 110	P	08/12/2025	16:30	LB136789