



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Aramark Uniforms

SDG No.: Q2789

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	5.34	6.0	89	65 - 135	P	08/12/2025	11:56	LB136789
	Copper	20.3	20.0	102	65 - 135	P	08/12/2025	11:56	LB136789
	Lead	11.1	12.0	93	65 - 135	P	08/12/2025	11:56	LB136789
	Nickel	38.5	40.0	96	65 - 135	P	08/12/2025	11:56	LB136789
	Zinc	40.0	40.0	100	65 - 135	P	08/12/2025	11:56	LB136789
CRA	Mercury	0.038	0.05	76	60 - 140	CV	08/12/2025	17:47	LB136791