

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Aramark Uniforms</u>	SDG No.: <u>Q2789</u>
Contract: <u>ARAM01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	1.01	1.0	101	-5	7	08/12/2025	12:00	LB136789
	Copper	4.88	2.0	244	-18	22	08/12/2025	12:00	LB136789
	Lead	-7.54			-12	12	08/12/2025	12:00	LB136789
	Nickel	5.34	2.0	267	-38	42	08/12/2025	12:00	LB136789
	Zinc	2.31			-40	40	08/12/2025	12:00	LB136789
ICSAB01	Cadmium	957	972	98	826	1120	08/12/2025	12:06	LB136789
	Copper	475	511	93	434	588	08/12/2025	12:06	LB136789
	Lead	39.5	49.0	81	37	61	08/12/2025	12:06	LB136789
	Nickel	966	954	101	810	1100	08/12/2025	12:06	LB136789
	Zinc	1010	952	106	809	1095	08/12/2025	12:06	LB136789
ICSA	Cadmium	-2.80	1.0	280	-5	7	08/12/2025	12:10	LB136789
	Copper	4.26	2.0	213	-18	22	08/12/2025	12:10	LB136789
	Lead	-5.72			-12	12	08/12/2025	12:10	LB136789
	Nickel	0.64	2.0	32	-38	42	08/12/2025	12:10	LB136789
	Zinc	3.13			-40	40	08/12/2025	12:10	LB136789
ICSAB	Cadmium	944	972	97	826	1120	08/12/2025	12:15	LB136789
	Copper	510	511	100	434	588	08/12/2025	12:15	LB136789
	Lead	17.4	49.0	36	37	61	08/12/2025	12:15	LB136789
	Nickel	962	954	101	810	1100	08/12/2025	12:15	LB136789
	Zinc	975	952	102	809	1095	08/12/2025	12:15	LB136789