

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms

SDG No.: Q2789

Contract: ARAM01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169195BS							
Cadmium	ug/L	100	92.5		92	85 - 115	P
Copper	ug/L	150	151		101	85 - 115	P
Lead	ug/L	500	463		93	85 - 115	P
Nickel	ug/L	250	240		96	85 - 115	P
Zinc	ug/L	100	99.8		100	85 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms

SDG No.: Q2789

Contract: ARAM01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169219BS Mercury	ug/L	4.0	3.91		98	85 - 115	CV