

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GARD04</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2790</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>08/07/2025 10:00</u>
Lab File ID:	<u>VN087483.D</u>	Init. Calib. Date(s):	<u>08/05/2025 08/05/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>13:37 15:01</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Acrolein	0.463	0.588		27	
Acrylonitrile	1.324	1.260		-4.83	
1,2-Dichloroethane-d4	2.767	2.737	0.01	-1.08	
Toluene-d8	1.380	1.399	0.01	1.38	
4-Bromofluorobenzene	0.516	0.509	0.2	-1.36	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.