

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GARD04
 Lab Code: ACE SDG NO.: Q2790
 Lab File ID: VX047251.D Date Analyzed: 08/07/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:15
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	180150	12.018				
UPPER LIMIT	360300	12.518				
LOWER LIMIT	90075	11.518				
EPA SAMPLE NO.						
250806078-01 VOA	195537	12.02				
250806078-01 VOADL	268161	12.02				
250806062-03 Trip blank	220877	12.02				
VX0807WBL01	198090	12.02				
VX0807WBS01	189057	12.02				
VX0807WBSD01	184985	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.