



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Aramark Uniforms

Project Name: Monthly 2025

Project # N/A

Order ID # Q2791

Test Name: Corrosivity, Ignitability, Reactive Cyanide, Reactive Sulfide

A. Number of Samples and Date of Receipt:

1 Soil sample was received on 08/07/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Corrosivity, Ignitability, Reactive Cyanide, Reactive Sulfide. This data package contains results for Corrosivity, Ignitability, Reactive Cyanide, Reactive Sulfide.

C. Analytical Techniques:

The analysis of Ignitability was based on method 1030, The analysis of Reactive Cyanide was based on method 9012B, The analysis of Reactive Sulfide was based on method 9034 and The analysis of Corrosivity was based on method 9045D.

D. QA/ QC Samples:

The Holding Times were met for all samples except for SLUDGE of Corrosivity as sample was receive out of holding time.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____