



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Aramark Uniforms

Project Name: Monthly 2025

Project # N/A

Order ID # Q2791

Test Name: TCLP ICP Metals, TCLP Mercury

A. Number of Samples and Date of Receipt:

1 Soil sample was received on 08/07/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: TCLP BNA Group1, TCLP Extraction, TCLP ICP Metals, TCLP Mercury, TCLP METALS, TCLP VOA and TCLP ZHE Extraction. This data package contains results for TCLP ICP Metals, TCLP Mercury.

C. Analytical Techniques:

The analysis of TCLP ICP Metals was based on method 6010D, digestion based on method 3010 (waters). The analysis and digestion of TCLP Mercury was based on method 7470A and TCLP extraction method was 1311.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Duplicate (VNJ-222DUP) analysis met criteria for all compounds except for Barium due to sample matrix interference. The Duplicate (VNJ-222MSD) analysis met criteria for all compounds except for Silver due to Chemical Interference during Digestion Process.

The Matrix Spike (VNJ-222MS) analysis met criteria for all compounds except for Barium due to Chemical Interference during Digestion Process.

The Matrix Spike Duplicate (VNJ-222MSD) analysis met criteria for all compounds except for Barium and Silver due to Chemical Interference during Digestion Process.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

The Post Digest Spike (VNJ-222A) analysis met criteria for all compounds except for Barium due to unknown chemical interference of matrix with the addition of spike amount after digestion and before analysis; matrix has suppression effect during addition of spike.



I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____