

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136774

Review By	sagar	Review On	8/11/2025 7:02:10 PM
Supervise By	jaswal	Supervise On	8/11/2025 11:15:28 PM

STD. NAME	STD REF.#
ICAL Standard	MP86690,MP86691,MP86692,MP86692,MP86693,MP86694,MP86695,MP86696
ICV Standard	MP86697
CCV Standard	MP86699
ICSA Standard	
CRI Standard	MP86701
LCS Standard	
Chk Standard	MP86698,MP86700,MP86702,MP86705

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	08/11/25 16:17		sagar	OK
2	S0.2	S0.2	CAL2	08/11/25 16:22		sagar	OK
3	S2.5	S2.5	CAL3	08/11/25 16:24		sagar	OK
4	S5	S5	CAL4	08/11/25 16:32		sagar	OK
5	S7.5	S7.5	CAL5	08/11/25 16:50		sagar	OK
6	S10	S10	CAL6	08/11/25 17:01		sagar	OK
7	ICV07	ICV07	ICV	08/11/25 17:07		sagar	OK
8	ICB07	ICB07	ICB	08/11/25 17:09		sagar	OK
9	CCV19	CCV19	CCV	08/11/25 17:22		sagar	OK
10	CCB19	CCB19	CCB	08/11/25 17:24		sagar	OK
11	CRA	CRA	CRDL	08/11/25 17:26		sagar	OK
12	HighStd	HighStd	HIGH STD	08/11/25 17:37		sagar	OK
13	ChkStd	ChkStd	SAM	08/11/25 17:42		sagar	OK
14	PB169166BL	PB169166BL	MB	08/11/25 17:45		sagar	OK
15	PB169166BS	PB169166BS	LCS	08/11/25 17:47		sagar	OK
16	Q2779-04	TP-4	SAM	08/11/25 17:50		sagar	OK
17	Q2779-08	TP-5	SAM	08/11/25 17:52		sagar	OK
18	Q2783-03	RT5414	SAM	08/11/25 17:54		sagar	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136774

Review By	sagar	Review On	8/11/2025 7:02:10 PM
Supervise By	jaswal	Supervise On	8/11/2025 11:15:28 PM

STD. NAME	STD REF.#
ICAL Standard	MP86690,MP86691,MP86692,MP86692,MP86693,MP86694,MP86695,MP86696
ICV Standard	MP86697
CCV Standard	MP86699
ICSA Standard	
CRI Standard	MP86701
LCS Standard	
Chk Standard	MP86698,MP86700,MP86702,MP86705

19	Q2791-01	SLUDGE	SAM	08/11/25 17:57		sagar	OK
20	Q2798-04	TP-6	SAM	08/11/25 17:59		sagar	OK
21	CCV20	CCV20	CCV	08/11/25 18:01		sagar	OK
22	CCB20	CCB20	CCB	08/11/25 18:03		sagar	OK
23	Q2800-02	VNJ-222	SAM	08/11/25 18:06		sagar	OK
24	Q2800-02DUP	VNJ-222DUP	DUP	08/11/25 18:08		sagar	OK
25	Q2800-02MS	VNJ-222MS	MS	08/11/25 18:10		sagar	OK
26	Q2800-02MSD	VNJ-222MSD	MSD	08/11/25 18:12		sagar	OK
27	Q2800-02L	VNJ-222L	SD	08/11/25 18:15		sagar	OK
28	Q2800-02A	VNJ-222A	PS	08/11/25 18:17		sagar	OK
29	PB169166TB	PB169166TB	MB	08/11/25 18:19	PB169149TB	sagar	OK
30	CCV21	CCV21	CCV	08/11/25 18:22		sagar	OK
31	CCB21	CCB21	CCB	08/11/25 18:24		sagar	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136778

Review By	Janvi	Review On	8/12/2025 7:03:55 PM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP86640,MP86656,MP86653,MP86643,MP86642,MP86641		
ICV Standard	MP86645,MP86656		
CCV Standard	MP86646		
ICSA Standard	MP86651,MP86652		
CRI Standard	MP86656		
LCS Standard			
Chk Standard	MP86647,MP86648		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	08/11/25 10:43		Jaswal	OK
2	S1	S1	CAL2	08/11/25 10:48		Jaswal	OK
3	S2	S2	CAL3	08/11/25 10:52		Jaswal	OK
4	S3	S3	CAL4	08/11/25 10:56		Jaswal	OK
5	S4	S4	CAL5	08/11/25 11:00		Jaswal	OK
6	S5	S5	CAL6	08/11/25 11:04		Jaswal	OK
7	ICV01	ICV01	ICV	08/11/25 11:09	ICV fail for B,Mo	Jaswal	OK
8	LLICV01	LLICV01	LLICV	08/11/25 11:33		Jaswal	OK
9	ICB01	ICB01	ICB	08/11/25 11:37		Jaswal	OK
10	CRI01	CRI01	CRDL	08/11/25 11:41		Jaswal	OK
11	ICSA01	ICSA01	ICSA	08/11/25 11:48		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	08/11/25 11:52	ICSAB fail for B,Mo	Jaswal	OK
13	ICSADL	ICSADL	ICSA	08/11/25 11:56		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	08/11/25 12:00		Jaswal	OK
15	CCV01	CCV01	CCV	08/11/25 12:05		Jaswal	OK
16	CCB01	CCB01	CCB	08/11/25 12:09		Jaswal	OK
17	Q2788-01	OK-01-08062025	SAM	08/11/25 12:13		Jaswal	OK
18	Q2783-03	RT5414	SAM	08/11/25 12:17		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136778

Review By	Janvi	Review On	8/12/2025 7:03:55 PM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP86640,MP86656,MP86653,MP86643,MP86642,MP86641		
ICV Standard	MP86645,MP86656		
CCV Standard	MP86646		
ICSA Standard	MP86651,MP86652		
CRI Standard	MP86656		
LCS Standard			
Chk Standard	MP86647,MP86648		

19	PB169178BL	PB169178BL	MB	08/11/25 12:22		Jaswal	OK
20	PB169178BS	PB169178BS	LCS	08/11/25 12:26		Jaswal	OK
21	PB169149TB	PB169149TB	MB	08/11/25 12:42		Jaswal	OK
22	Q2779-04	TP-4	SAM	08/11/25 12:47		Jaswal	OK
23	Q2779-08	TP-5	SAM	08/11/25 12:51		Jaswal	OK
24	Q2791-01	SLUDGE	SAM	08/11/25 12:56		Jaswal	OK
25	CCV02	CCV02	CCV	08/11/25 13:00		Jaswal	OK
26	CCB02	CCB02	CCB	08/11/25 13:04		Jaswal	OK
27	Q2798-04	TP-6	SAM	08/11/25 13:08		Jaswal	OK
28	Q2800-02	VNJ-222	SAM	08/11/25 13:13		Jaswal	OK
29	Q2800-02DUP	VNJ-222DUP	DUP	08/11/25 13:17		Jaswal	OK
30	Q2800-02L	VNJ-222L	SD	08/11/25 13:22		Jaswal	OK
31	Q2800-02MS	VNJ-222MS	MS	08/11/25 13:26		Jaswal	OK
32	Q2800-02MSD	VNJ-222MSD	MSD	08/11/25 13:30		Jaswal	OK
33	Q2800-02A	VNJ-222A	PS	08/11/25 13:34	0.1 ml of each M6004,M6017	Jaswal	OK
34	Q2779-01	TP-4	SAM	08/11/25 13:38		Jaswal	OK
35	Q2779-05	TP-5	SAM	08/11/25 13:42		Jaswal	OK
36	Q2793-01	VNJ-231	SAM	08/11/25 13:46		Jaswal	OK
37	CCV03	CCV03	CCV	08/11/25 14:54		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136778

Review By	Janvi	Review On	8/12/2025 7:03:55 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP86640,MP86656,MP86653,MP86643,MP86642,MP86641
ICV Standard	MP86645,MP86656
CCV Standard	MP86646
ICSA Standard	MP86651,MP86652
CRI Standard	MP86656
LCS Standard	
Chk Standard	MP86647,MP86648

38	CCB03	CCB03	CCB	08/11/25 14:58		Jaswal	OK
39	Q2793-01DUP	VNJ-231DUP	DUP	08/11/25 15:03		Jaswal	OK
40	Q2793-01L	VNJ-231L	SD	08/11/25 15:07		Jaswal	OK
41	Q2793-01MS	VNJ-231MS	MS	08/11/25 15:11		Jaswal	OK
42	Q2793-01MSD	VNJ-231MSD	MSD	08/11/25 15:15		Jaswal	OK
43	Q2793-01A	VNJ-231A	PS	08/11/25 15:19	0.1 ml of each M6004,M6017	Jaswal	OK
44	Q2795-01	COMP-1	SAM	08/11/25 15:23		Jaswal	OK
45	Q2795-02	COMP-2	SAM	08/11/25 15:27		Jaswal	OK
46	Q2795-03	COMP-3	SAM	08/11/25 15:31		Jaswal	OK
47	Q2759-04	LIQUID-DRUM	SAM	08/11/25 15:35		Jaswal	OK
48	Q2800-01	VNJ-222	SAM	08/11/25 15:39		Jaswal	OK
49	CCV04	CCV04	CCV	08/11/25 16:02		Jaswal	OK
50	CCB04	CCB04	CCB	08/11/25 16:06		Jaswal	OK