

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q2791

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV07	Mercury	3.76	4.0	94	90 - 110	CV	08/11/2025	17:07	LB136774

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q2791

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV19	Mercury	5.06	5.0	101	90 - 110	CV	08/11/2025	17:22	LB136774
CCV20	Mercury	4.80	5.0	96	90 - 110	CV	08/11/2025	18:01	LB136774
CCV21	Mercury	4.86	5.0	97	90 - 110	CV	08/11/2025	18:22	LB136774

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q2791

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	3700	4000	93	90 - 110	P	08/11/2025	11:09	LB136778
	Barium	7500	8000	94	90 - 110	P	08/11/2025	11:09	LB136778
	Cadmium	1840	2000	92	90 - 110	P	08/11/2025	11:09	LB136778
	Chromium	755	800	94	90 - 110	P	08/11/2025	11:09	LB136778
	Lead	3840	4000	96	90 - 110	P	08/11/2025	11:09	LB136778
	Selenium	3720	4000	93	90 - 110	P	08/11/2025	11:09	LB136778
	Silver	940	1000	94	90 - 110	P	08/11/2025	11:09	LB136778

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q2791

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	19.2	20.0	96	80 - 120	P	08/11/2025	11:33	LB136778
	Barium	97.3	100	97	80 - 120	P	08/11/2025	11:33	LB136778
	Cadmium	5.81	6.0	97	80 - 120	P	08/11/2025	11:33	LB136778
	Chromium	9.94	10.0	99	80 - 120	P	08/11/2025	11:33	LB136778
	Lead	12.4	12.0	103	80 - 120	P	08/11/2025	11:33	LB136778
	Selenium	21.6	20.0	108	80 - 120	P	08/11/2025	11:33	LB136778
	Silver	10.2	10.0	102	80 - 120	P	08/11/2025	11:33	LB136778

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q2791

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	4930	5000	99	90 - 110	P	08/11/2025	12:05	LB136778
	Barium	10000	10000	100	90 - 110	P	08/11/2025	12:05	LB136778
	Cadmium	2420	2500	97	90 - 110	P	08/11/2025	12:05	LB136778
	Chromium	985	1000	98	90 - 110	P	08/11/2025	12:05	LB136778
	Lead	4830	5000	97	90 - 110	P	08/11/2025	12:05	LB136778
	Selenium	4890	5000	98	90 - 110	P	08/11/2025	12:05	LB136778
	Silver	1240	1250	99	90 - 110	P	08/11/2025	12:05	LB136778
CCV02	Arsenic	4890	5000	98	90 - 110	P	08/11/2025	13:00	LB136778
	Barium	9690	10000	97	90 - 110	P	08/11/2025	13:00	LB136778
	Cadmium	2340	2500	94	90 - 110	P	08/11/2025	13:00	LB136778
	Chromium	961	1000	96	90 - 110	P	08/11/2025	13:00	LB136778
	Lead	4680	5000	94	90 - 110	P	08/11/2025	13:00	LB136778
	Selenium	4890	5000	98	90 - 110	P	08/11/2025	13:00	LB136778
	Silver	1220	1250	97	90 - 110	P	08/11/2025	13:00	LB136778
CCV03	Arsenic	4840	5000	97	90 - 110	P	08/11/2025	14:54	LB136778
	Barium	9090	10000	91	90 - 110	P	08/11/2025	14:54	LB136778
	Cadmium	2320	2500	93	90 - 110	P	08/11/2025	14:54	LB136778
	Chromium	913	1000	91	90 - 110	P	08/11/2025	14:54	LB136778
	Lead	4620	5000	92	90 - 110	P	08/11/2025	14:54	LB136778
	Selenium	4950	5000	99	90 - 110	P	08/11/2025	14:54	LB136778
	Silver	1170	1250	94	90 - 110	P	08/11/2025	14:54	LB136778
CCV04	Arsenic	4990	5000	100	90 - 110	P	08/11/2025	16:02	LB136778
	Barium	9100	10000	91	90 - 110	P	08/11/2025	16:02	LB136778
	Cadmium	2370	2500	95	90 - 110	P	08/11/2025	16:02	LB136778
	Chromium	921	1000	92	90 - 110	P	08/11/2025	16:02	LB136778
	Lead	4740	5000	95	90 - 110	P	08/11/2025	16:02	LB136778
	Selenium	5090	5000	102	90 - 110	P	08/11/2025	16:02	LB136778
	Silver	1180	1250	94	90 - 110	P	08/11/2025	16:02	LB136778