



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Aramark Uniforms

SDG No.: Q2791

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	18.6	20.0	93	65 - 135	P	08/11/2025	11:41	LB136778
	Barium	92.3	100	92	65 - 135	P	08/11/2025	11:41	LB136778
	Cadmium	5.33	6.0	89	65 - 135	P	08/11/2025	11:41	LB136778
	Chromium	9.43	10.0	94	65 - 135	P	08/11/2025	11:41	LB136778
	Lead	12.1	12.0	101	65 - 135	P	08/11/2025	11:41	LB136778
	Selenium	19.0	20.0	95	65 - 135	P	08/11/2025	11:41	LB136778
	Silver	9.70	10.0	97	65 - 135	P	08/11/2025	11:41	LB136778
CRA	Mercury	0.20	0.2	101	70 - 130	CV	08/11/2025	17:26	LB136774