



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Aramark Uniforms

SDG No.: Q2791

Contract: ARAM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB07	Mercury	0.076	+/-0.2	U	0.20	CV	08/11/2025	17:09	LB136774

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Aramark Uniforms

SDG No.: Q2791

Contract: ARAM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB19	Mercury	0.076	+/-0.2	U	0.20	CV	08/11/2025	17:24	LB136774
CCB20	Mercury	0.076	+/-0.2	U	0.20	CV	08/11/2025	18:03	LB136774
CCB21	Mercury	0.076	+/-0.2	U	0.20	CV	08/11/2025	18:24	LB136774

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Aramark Uniforms

SDG No.: Q2791

Contract: ARAM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	08/11/2025	11:37	LB136778
	Barium	14.6	+/-50	U	100	P	08/11/2025	11:37	LB136778
	Cadmium	0.50	+/-3	U	6.00	P	08/11/2025	11:37	LB136778
	Chromium	2.12	+/-5	U	10.0	P	08/11/2025	11:37	LB136778
	Lead	2.30	+/-6	U	12.0	P	08/11/2025	11:37	LB136778
	Selenium	9.64	+/-10	U	20.0	P	08/11/2025	11:37	LB136778
	Silver	1.62	+/-5	U	10.0	P	08/11/2025	11:37	LB136778

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Aramark Uniforms

SDG No.: Q2791

Contract: ARAM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	08/11/2025	12:09	LB136778
	Barium	14.6	+/-50	U	100	P	08/11/2025	12:09	LB136778
	Cadmium	0.90	+/-3	J	6.00	P	08/11/2025	12:09	LB136778
	Chromium	2.12	+/-5	U	10.0	P	08/11/2025	12:09	LB136778
	Lead	2.66	+/-6	J	12.0	P	08/11/2025	12:09	LB136778
	Selenium	9.64	+/-10	U	20.0	P	08/11/2025	12:09	LB136778
	Silver	1.62	+/-5	U	10.0	P	08/11/2025	12:09	LB136778
CCB02	Arsenic	5.12	+/-10	U	20.0	P	08/11/2025	13:04	LB136778
	Barium	14.6	+/-50	U	100	P	08/11/2025	13:04	LB136778
	Cadmium	0.89	+/-3	J	6.00	P	08/11/2025	13:04	LB136778
	Chromium	2.12	+/-5	U	10.0	P	08/11/2025	13:04	LB136778
	Lead	2.35	+/-6	J	12.0	P	08/11/2025	13:04	LB136778
	Selenium	9.64	+/-10	U	20.0	P	08/11/2025	13:04	LB136778
	Silver	1.62	+/-5	U	10.0	P	08/11/2025	13:04	LB136778
CCB03	Arsenic	5.12	+/-10	U	20.0	P	08/11/2025	14:58	LB136778
	Barium	14.6	+/-50	U	100	P	08/11/2025	14:58	LB136778
	Cadmium	0.50	+/-3	U	6.00	P	08/11/2025	14:58	LB136778
	Chromium	2.12	+/-5	U	10.0	P	08/11/2025	14:58	LB136778
	Lead	2.30	+/-6	U	12.0	P	08/11/2025	14:58	LB136778
	Selenium	9.64	+/-10	U	20.0	P	08/11/2025	14:58	LB136778
	Silver	1.62	+/-5	U	10.0	P	08/11/2025	14:58	LB136778
CCB04	Arsenic	5.12	+/-10	U	20.0	P	08/11/2025	16:06	LB136778
	Barium	14.6	+/-50	U	100	P	08/11/2025	16:06	LB136778
	Cadmium	0.50	+/-3	U	6.00	P	08/11/2025	16:06	LB136778
	Chromium	2.12	+/-5	U	10.0	P	08/11/2025	16:06	LB136778
	Lead	2.30	+/-6	U	12.0	P	08/11/2025	16:06	LB136778
	Selenium	9.64	+/-10	U	20.0	P	08/11/2025	16:06	LB136778
	Silver	1.62	+/-5	U	10.0	P	08/11/2025	16:06	LB136778