

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Aramark Uniforms</u>	SDG No.: <u>Q2791</u>
Contract: <u>ARAM01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	8.68			-20	20	08/11/2025	11:48	LB136778
	Barium	0.44	6.0	7	-94	106	08/11/2025	11:48	LB136778
	Cadmium	0.96	1.0	96	-5	7	08/11/2025	11:48	LB136778
	Chromium	56.6	52.0	109	42	62	08/11/2025	11:48	LB136778
	Lead	-5.21			-12	12	08/11/2025	11:48	LB136778
	Selenium	2.51			-20	20	08/11/2025	11:48	LB136778
	Silver	1.58			-10	10	08/11/2025	11:48	LB136778
ICSAB01	Arsenic	104	104	100	88.4	120	08/11/2025	11:52	LB136778
	Barium	485	537	90	437	637	08/11/2025	11:52	LB136778
	Cadmium	933	972	96	826	1120	08/11/2025	11:52	LB136778
	Chromium	530	542	98	460	624	08/11/2025	11:52	LB136778
	Lead	39.0	49.0	80	37	61	08/11/2025	11:52	LB136778
	Selenium	47.3	46.0	103	26	66	08/11/2025	11:52	LB136778
	Silver	209	201	104	170	232	08/11/2025	11:52	LB136778