

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>Q2791</u>
contract: <u>ARAM01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2800-02</u>	client id: <u>VNJ-222MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2800-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	4040		100	U	4000	101		P
Barium	ug/L	75 - 125	3140		1490		1000	165	N	P
Cadmium	ug/L	75 - 125	966		30.0	U	1000	97		P
Chromium	ug/L	75 - 125	1890		50.0	U	2000	94		P
Lead	ug/L	75 - 125	4590		31.0	J	5000	91		P
Mercury	ug/L	75 - 125	34.6		2.00	U	40.0	86		CV
Selenium	ug/L	75 - 125	10300		100	U	10000	103		P
Silver	ug/L	75 - 125	358		50.0	U	380	94		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>Q2791</u>
contract: <u>ARAM01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2800-02</u>	client id: <u>VNJ-222MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2800-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	4140		100	U	4000	104		P
Barium	ug/L	75 - 125	3030		1490		1000	154	N	P
Cadmium	ug/L	75 - 125	986		30.0	U	1000	99		P
Chromium	ug/L	75 - 125	2000		50.0	U	2000	100		P
Lead	ug/L	75 - 125	4680		31.0	J	5000	93		P
Mercury	ug/L	75 - 125	34.1		2.00	U	40.0	85		CV
Selenium	ug/L	75 - 125	10600		100	U	10000	106		P
Silver	ug/L	75 - 125	252		50.0	U	380	66	N	P