

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Aramark Uniforms</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2791</u>
Contract:	<u>ARAM01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2800-02</u>	Client ID:	<u>VNJ-222DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2800-02DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	1490		2660		56	*	P
Cadmium	ug/L	20	30.0	U	4.42	J	200.0		P
Chromium	ug/L	20	50.0	U	50.0	U			P
Lead	ug/L	20	31.0	J	50.8	J	48		P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Aramark Uniforms</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2791</u>
Contract:	<u>ARAM01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2800-02MS</u>	Client ID:	<u>VNJ-222MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2800-02MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	4040		4140		2		P
Barium	ug/L	20	3140		3030		4		P
Cadmium	ug/L	20	966		986		2		P
Chromium	ug/L	20	1890		2000		6		P
Lead	ug/L	20	4590		4680		2		P
Mercury	ug/L	20	34.6		34.1		1		CV
Selenium	ug/L	20	10300		10600		3		P
Silver	ug/L	20	358		252		35	*	P