

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms

SDG No.: Q2791

Contract: ARAM01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169166BS Mercury	ug/L	4.0	4.15		104	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms

SDG No.: Q2791

Contract: ARAM01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169178BS							
Arsenic	ug/L	4000	3870		97	80 - 120	P
Barium	ug/L	1000	950		95	80 - 120	P
Cadmium	ug/L	1000	955		96	80 - 120	P
Chromium	ug/L	2000	1980		99	80 - 120	P
Lead	ug/L	5000	4760		95	80 - 120	P
Selenium	ug/L	10000	9700		97	80 - 120	P
Silver	ug/L	380	370		97	80 - 120	P