

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2791

Analytical Method: 8270E

Client: Aramark Uniforms

DataFile: BP025364.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB169180BS	Pyridine	50	37.7	ug/L	75				29	97	
	1,4-Dichlorobenzene	50	44.6	ug/L	89				76	103	
	2-Methylphenol	50	42.9	ug/L	86				69	109	
	3+4-Methylphenols	50	41.8	ug/L	84				67	106	
	Hexachloroethane	50	45.8	ug/L	92				76	118	
	Nitrobenzene	50	46.5	ug/L	93				58	106	
	2,4,6-Trichlorophenol	50	54.1	ug/L	108				61	110	
	2,4,5-Trichlorophenol	50	52.5	ug/L	105				70	106	
	2,4-Dinitrotoluene	50	48.8	ug/L	98				60	115	
	Hexachlorobenzene	50	50.4	ug/L	101				73	106	
	Pentachlorophenol	100	110	ug/L	110				47	114	