

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2791

Client ID : SSTDCCC040

Date Analyzed: 08/11/2025

Lab File ID: BP025343.D

Time Analyzed: 11:08

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	216631	7.796	873095	10.57	558968	14.42
UPPER LIMIT	433262	8.296	1746190	11.072	1117940	14.919
LOWER LIMIT	108316	7.296	436548	10.072	279484	13.919
EPA SAMPLE NO.						
01 TP-4MS	215026	7.80	827224	10.57	489140	14.42
02 TP-4MSD	284673	7.80	1089870	10.57	635339	14.42
03 SLUDGE	334679	7.80	1364070	10.58	820532	14.43

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.