

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2791

Client ID : SSTDCCC040

Date Analyzed: 08/12/2025

Lab File ID: BP025360.D

Time Analyzed: 10:07

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	195999	7.796	806810	10.58	510213	14.43
UPPER LIMIT	391998	8.296	1613620	11.078	1020430	14.931
LOWER LIMIT	97999.5	7.296	403405	10.078	255107	13.931
EPA SAMPLE NO.						
01 PB169149TB	216198	7.80	822002	10.57	481171	14.43
02 PB169180BL	223850	7.80	834408	10.57	478110	14.42
03 PB169180BS	220840	7.80	869704	10.58	505729	14.43

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.