

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BP025359.D
Instrument ID: BNA_P

Contract: ARAM01
SDG NO.: Q2791
DFTPP Injection Date: 08/12/2025
DFTPP Injection Time: 09:26

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.6) 1
69	Mass 69 relative abundance	31.6
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	15.1
442	Greater than 50% of mass 198	97.1
443	15.0 - 24.0% of mass 442	19.1 (19.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP025360.D	08/12/2025	10:07
PB169180BL	PB169180BL	BP025363.D	08/12/2025	12:12
PB169180BS	PB169180BS	BP025364.D	08/12/2025	12:54
PB169149TB	PB169149TB	BP025368.D	08/12/2025	15:40