

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2791</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Aramark Uniforms</u>	Datafile :	<u>VX047266.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0808WBS01	Vinyl chloride	20	20.5	ug/L	103			65	117	
	1,1-Dichloroethene	20	19.1	ug/L	96			74	110	
	2-Butanone	100	99.6	ug/L	100			65	122	
	Carbon Tetrachloride	20	18.9	ug/L	95			77	113	
	Chloroform	20	19.3	ug/L	97			79	113	
	Benzene	20	19.7	ug/L	99			82	109	
	1,2-Dichloroethane	20	19.8	ug/L	99			80	115	
	Trichloroethene	20	19.2	ug/L	96			77	113	
	Tetrachloroethene	20	19.2	ug/L	96			67	123	
	Chlorobenzene	20	19.5	ug/L	98			82	109	