

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2791</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Aramark Uniforms</u>	Datafile :	<u>VX047267.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0808WBSD01	Vinyl chloride	20	19.9	ug/L	100	3		65	117	19
	1,1-Dichloroethene	20	18.7	ug/L	94	2		74	110	20
	2-Butanone	100	100	ug/L	100	0		65	122	26
	Carbon Tetrachloride	20	19.2	ug/L	96	1		77	113	15
	Chloroform	20	19.3	ug/L	97	0		79	113	20
	Benzene	20	19.6	ug/L	98	1		82	109	15
	1,2-Dichloroethane	20	20.3	ug/L	102	3		80	115	20
	Trichloroethene	20	18.9	ug/L	95	1		77	113	15
	Tetrachloroethene	20	18.8	ug/L	94	2		67	123	15
	Chlorobenzene	20	19.8	ug/L	99	1		82	109	15