

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ARAM01
 Lab Code: ACE SDG NO.: Q2791
 Lab File ID: VX047263.D Date Analyzed: 08/08/2025
 Instrument ID: MSVOA_X Time Analyzed: 11:04
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	231247	5.56	404873	6.76	364546	10.06
UPPER LIMIT	462494	6.056	809746	7.263	729092	10.555
LOWER LIMIT	115624	5.056	202437	6.263	182273	9.555
EPA SAMPLE NO.						
SLUDGE	282257	5.56	474035	6.77	393822	10.06
VX0808WBL01	368606	5.56	680757	6.76	661932	10.06
VX0808WBS01	241260	5.56	404774	6.76	361083	10.06
VX0808WBSD01	229567	5.56	385678	6.76	348369	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.