

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ARAM01
 Lab Code: ACE SDG NO.: Q2791
 Lab File ID: VX047263.D Date Analyzed: 08/08/2025
 Instrument ID: MSVOA_X Time Analyzed: 11:04
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	177002	12.018				
UPPER LIMIT	354004	12.518				
LOWER LIMIT	88501	11.518				
EPA SAMPLE NO.						
SLUDGE	211582	12.02				
VX0808WBL01	331943	12.02				
VX0808WBS01	185608	12.02				
VX0808WBSD01	177274	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.