

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081325\
 Data File : VL042849.D
 Acq On : 13 Aug 2025 12:56
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL081325

Quant Time: Aug 14 02:16:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL081325AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Thu Aug 14 02:12:54 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	1.466	1.334	9.0	91	0.00
3	Chlorodifluoromethane	1.620	1.511	6.7	92	0.00
4	Chloromethane	0.591	0.520	12.0	90	0.00
5 T	Vinyl Chloride	0.582	0.479	17.7	88	0.00
6 T	Bromomethane	0.233	0.205	12.0	93	0.00
7	Chloroethane	0.227	0.190	16.3	87	0.00
8 T	Dichlorotetrafluoroethane	1.199	1.109	7.5	91	0.00
9 T	Propene	0.700	0.624	10.9	88	0.00
10 T	Heptane	1.874	1.685	10.1	90	0.00
11 T	Trichlorofluoromethane	1.446	1.259	12.9	91	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.152	1.017	11.7	89	0.00
13	Ethanol	0.064	0.049#	23.4	97	0.00
14 T	Bromoethene	0.417	0.380	8.9	92	0.00
15 T	Acetone	1.323	0.977	26.2	95	0.00
16 T	1,3-Butadiene	0.600	0.497	17.2	91	0.00
17	tert-Butyl alcohol	1.569	1.359	13.4	91	0.00
18 T	1,1-Dichloroethene	0.507	0.460	9.3	90	0.00
19 T	Isopropyl Alcohol	0.790	0.644	18.5	91	0.00
20 T	Methylene Chloride	0.554	0.390	29.6	87	0.00
21 T	Allyl Chloride	0.935	0.837	10.5	91	0.00
22 T	trans-1,2-Dichloroethene	0.586	0.522	10.9	89	0.00
23 T	Vinyl Acetate	1.819	1.617	11.1	89	0.00
24 T	1,1-Dichloroethane	1.171	1.042	11.0	92	0.00
25 T	Ethyl Acetate	3.202	2.958	7.6	93	0.00
26 T	Hexane	1.478	1.366	7.6	90	0.00
27 T	Carbon Disulfide	1.438	1.271	11.6	89	0.00
28 T	Methyl tert-Butyl Ether	0.754	0.670	11.1	93	0.00
29 T	Chloroform	2.110	1.978	6.3	94	0.00
30 T	Cyclohexane	1.296	1.204	7.1	91	0.00
31 T	cis-1,2-Dichloroethene	1.476	1.357	8.1	93	0.00
32 T	1,1,1-Trichloroethane	2.230	2.064	7.4	95	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
34 T	2-Butanone	0.675	0.617	8.6	92	0.00
35 T	Carbon Tetrachloride	0.689	0.653	5.2	93	0.00
36 T	Benzene	0.958	0.919	4.1	91	0.00
37 T	1,2-Dichloroethane	0.500	0.500	0.0	94	0.00
38 T	Trichloroethene	0.437	0.402	8.0	91	0.00
39 T	1,2-Dichloropropane	0.350	0.330	5.7	92	0.00
40 T	1,4-Dioxane	0.166	0.148	10.8	90	0.00
41 T	Tetrahydrofuran	0.376	0.353	6.1	90	0.00
42 T	Bromodichloromethane	0.745	0.721	3.2	92	0.00
43	Methyl Methacrylate	0.394	0.364	7.6	90	0.00
44 T	2,2,4-Trimethylpentane	1.625	1.510	7.1	90	0.00
45 T	t-1,3-Dichloropropene	0.445	0.429	3.6	91	0.00
46 T	cis-1,3-Dichloropropene	0.555	0.534	3.8	90	0.00
47 T	1,1,2-Trichloroethane	0.371	0.354	4.6	94	0.00
48 T	Dibromochloromethane	0.600	0.582	3.0	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081325\
 Data File : VL042849.D
 Acq On : 13 Aug 2025 12:56
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL081325

Quant Time: Aug 14 02:16:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL081325AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Thu Aug 14 02:12:54 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.495	0.467	5.7	86	0.00
50 T	4-Methyl-2-Pentanone	0.913	0.861	5.7	88	0.00
51 T	2-Hexanone	0.718	0.703	2.1	91	0.00
52 T	Tetrachloroethene	0.346	0.296	14.5	86	0.00
53 T	Toluene	1.138	1.063	6.6	90	0.00
54 T	1,2-Dibromoethane	0.542	0.516	4.8	92	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
56	1,1,1,2-Tetrachloroethane	0.517	0.480	7.2	91	0.00
57 T	Chlorobenzene	0.951	0.892	6.2	91	0.00
58 T	Ethyl Benzene	1.765	1.661	5.9	91	0.00
59 T	m/p-Xylene	1.383	1.297	6.2	91	0.02
60 T	o-Xylene	1.391	1.286	7.5	91	0.00
61 T	Styrene	0.652	0.611	6.3	90	0.00
62	Isopropylbenzene	2.033	1.872	7.9	91	0.00
63 T	1,1,2,2-Tetrachloroethane	0.765	0.675	11.8	90	0.00
64	n-propylbenzene	0.519	0.474	8.7	90	0.00
65	tert-Butylbenzene	1.793	1.604	10.5	90	0.00
66 T	Benzyl Chloride	0.228	0.206	9.6	81	0.00
67	sec-Butylbenzene	2.516	2.287	9.1	92	0.00
68 S	1-Bromo-4-Fluorobenzene	0.750	0.756	-0.8	90	0.00
69	p-Isopropyltoluene	2.160	1.903	11.9	90	0.00
70	n-Butylbenzene	2.142	1.941	9.4	91	0.00
71	2-Chlorotoluene	1.579	1.455	7.9	92	0.00
72 T	4-Ethyltoluene	1.687	1.575	6.6	91	0.00
73 T	1,3,5-Trimethylbenzene	1.434	1.306	8.9	91	0.00
74 T	1,2,4-Trimethylbenzene	1.574	1.408	10.5	91	0.00
75 T	1,3-Dichlorobenzene	0.895	0.824	7.9	89	0.00
76 T	1,4-Dichlorobenzene	0.905	0.813	10.2	88	0.00
77 T	1,2-Dichlorobenzene	0.876	0.766	12.6	88	0.00
78 T	Hexachloro-1,3-Butadiene	0.718	0.559	22.1	88	0.00
79 T	Naphthalene	1.427	1.325	7.1	91	0.00
80 T	Naphthalene,2-methyl-	0.533	0.437	18.0	82	0.00
81 T	1,2,4-Trichlorobenzene	0.717	0.647	9.8	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0