

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: GFEL01

Lab Code: ACE

SDG No.: Q2794

Instrument ID: MSVOA_L

Calibration Date(s): 08/13/2025 08/13/2025

Heated Purge: (Y/N) N

Calibration Time(s): 08:30 12:06

GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:		RRF010 = VL042842.D		RRF002 = VL042843.D		RRF001 = VL042844.D		
		RRF0.5 = VL042845.D		RRF0.1 = VL042846.D		RRF.03 = VL042847.D		
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF.03	RRF	% RSD
Dichlorodifluoromethane	1.343	1.511	1.525	1.566			1.466	6.6
Chloromethane	0.533	0.606	0.618	0.648			0.591	8.1
Vinyl Chloride	0.502	0.550	0.559	0.593	0.639	0.728	0.582	13.7
Bromomethane	0.203	0.239	0.268	0.248			0.233	12.1
Chloroethane	0.202	0.219	0.240	0.266			0.227	11.5
Tetrahydrofuran	0.355	0.392	0.389	0.395			0.376	5.6
Trichlorofluoromethane	1.271	1.467	1.533	1.614			1.446	9.6
1,1,2-Trichlorotrifluoroethane	1.047	1.189	1.202	1.259			1.152	8
Dichlorotetrafluoroethane	1.122	1.245	1.235	1.274			1.199	6
tert-Butyl alcohol	1.371	1.557	1.660	1.920			1.569	15.1
Heptane	1.718	1.917	2.011	2.009			1.874	8
1,1-Dichloroethene	0.469	0.532	0.509	0.552			0.507	7.2
Acetone	0.950	1.476	1.475	1.708			1.323	24.9
Carbon Disulfide	1.314	1.511	1.513	1.504			1.438	6.8
Methyl tert-Butyl Ether	0.661	0.817	0.747	0.833			0.754	9.6
Methylene Chloride	0.414	0.544	0.596	0.792			0.554	27.8
trans-1,2-Dichloroethene	0.537	0.604	0.613	0.630			0.586	7.1
1,1-Dichloroethane	1.040	1.185	1.223	1.317			1.171	9.4
Cyclohexane	1.214	1.340	1.283	1.425			1.296	6.9
2-Butanone	0.610	0.700	0.702	0.753			0.675	9.3
Carbon Tetrachloride	0.636	0.692	0.700	0.682	0.738	0.736	0.689	5.8
cis-1,2-Dichloroethene	1.346	1.524	1.524	1.625			1.476	8.1
Chloroform	1.932	2.166	2.183	2.299			2.110	7.3
1,1,1-Trichloroethane	1.999	2.215	2.148	2.209	2.528	2.448	2.230	8.7
2,2,4-Trimethylpentane	1.515	1.676	1.719	1.728			1.625	7.1
Benzene	0.912	0.990	0.978	1.013			0.958	5.3
1,2-Dichloroethane	0.483	0.503	0.516	0.512			0.500	2.9
Trichloroethene	0.399	0.453	0.448	0.501	0.448	0.416	0.437	8.3
1,2-Dichloropropane	0.324	0.356	0.365	0.378			0.350	6.8
Bromodichloromethane	0.708	0.777	0.742	0.786			0.745	4.8

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: GFEL01

Lab Code: ACE

SDG No.: Q2794

Instrument ID: MSVOA_L

Calibration Date(s): 08/13/2025 08/13/2025

Heated Purge: (Y/N) N

Calibration Time(s): 08:30 12:06

GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:		RRF010 = VL042842.D		RRF002 = VL042843.D		RRF001 = VL042844.D		RRF0.5 = VL042845.D		RRF0.1 = VL042846.D		RRF.03 = VL042847.D	
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF.03	RRF	% RSD					
4-Methyl-2-Pentanone	0.881	0.946	0.917	0.960			0.913	4.7					
Toluene	1.072	1.160	1.181	1.219			1.138	6.1					
t-1,3-Dichloropropene	0.428	0.456	0.464	0.450			0.445	3.7					
cis-1,3-Dichloropropene	0.539	0.578	0.559	0.568			0.555	3.7					
1,1,2-Trichloroethane	0.341	0.386	0.382	0.395			0.371	6.5					
Dibromochloromethane	0.584	0.632	0.605	0.596			0.600	3.4					
1,2-Dibromoethane	0.506	0.552	0.532	0.549	0.599		0.542	6.2					
Tetrachloroethene	0.313	0.345	0.333	0.369	0.379	0.391	0.346	10.2					
Chlorobenzene	0.878	0.994	0.990	1.026			0.951	7.7					
Ethyl Benzene	1.636	1.838	1.866	1.854			1.765	6.8					
m/p-Xylene	1.283	1.455	1.454	1.447			1.383	6.8					
o-Xylene	1.266	1.445	1.472	1.532			1.391	9.3					
Styrene	0.610	0.660	0.677	0.711			0.652	7.1					
Bromoform	0.491	0.535	0.514	0.466			0.495	6.1					
1,1,2,2-Tetrachloroethane	0.674	0.772	0.784	0.795	0.774	0.881	0.765	9.4					
2-Chlorotoluene	1.418	1.641	1.719	1.709			1.579	9.8					
1,3,5-Trimethylbenzene	1.293	1.512	1.535	1.548			1.434	9.4					
1,2,4-Trimethylbenzene	1.394	1.698	1.706	1.710			1.574	11.3					
1,3-Dichlorobenzene	0.831	0.967	0.946	0.937			0.895	8.6					
1,4-Dichlorobenzene	0.833	0.978	0.946	0.966			0.905	8.9					
1,2-Dichlorobenzene	0.781	0.947	0.956	0.937			0.876	11					
1,2,4-Trichlorobenzene	0.661	0.855	0.797	0.623			0.717	14.2					
Hexachloro-1,3-Butadiene	0.572	0.804	0.845	0.814			0.718	19.7					
1,3-Butadiene	0.500	0.579	0.633	0.783			0.600	19.5					
Naphthalene	1.315	1.746	1.647	1.383	1.174		1.427	15.5					
4-Ethyltoluene	1.563	1.784	1.806	1.738			1.687	7.4					
1-Bromo-4-Fluorobenzene	0.755	0.744	0.760	0.755	0.746	0.736	0.750	1.1					
Hexane	1.392	1.522	1.552	1.557			1.478	6.2					
Allyl Chloride	0.850	0.945	0.902	1.122			0.935	12					
1,4-Dioxane	148.388	165.505	184.509	188.914			166.304	12.2					

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: GFEL01
 Lab Code: ACE SDG No.: Q2794
 Instrument ID: MSVOA_L Calibration Date(s): 08/13/2025 08/13/2025
 Heated Purge: (Y/N) N Calibration Time(s): 08:30 12:06
 GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:		RRF010 = VL042842.D		RRF002 = VL042843.D		RRF001 = VL042844.D		
		RRF0.5 = VL042845.D		RRF0.1 = VL042846.D		RRF.03 = VL042847.D		
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF.03	RRF	% RSD
Methyl Methacrylate	0.368	0.415	0.402	0.427			0.394	7.3

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.