

Duplicate Sample Summary

Lab Sample Id :	Q2794-01DUP	Q2794-01
Client Id :	IA1DUP	IA1
DF :	1	1
Datafile :	VL042856.D	VL042855.D
Anal Date & Time :	08/13/2025 17:43	08/13/2025 17:05

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	1.8	1.9	5.4
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	0.49	0.48	2.1
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0.15	0.15	0
1,4-Dioxane	0	0	0
2,2,4-Trimethylpentane	2.3	2.3	0
2-Butanone	0.87	0.92	5.6
2-Chlorotoluene	0	0	0
4-Ethyltoluene	0.54	0.54	0
4-Methyl-2-Pentanone	0.2	0.2	0
Acetone	32.9	32.1	2.5
Allyl Chloride	0	0	0
Benzene	1.3	1.2	8
Bromodichloromethane	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	0	0	0
Carbon Tetrachloride	0.08	0.07	13.3

Duplicate Sample Summary

Lab Sample Id :	Q2794-01DUP	Q2794-01
Client Id :	IA1DUP	IA1
DF :	1	1
Datafile :	VL042856.D	VL042855.D
Anal Date & Time :	08/13/2025 17:43	08/13/2025 17:05

Parameter	Result	Result	RPD
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	0.29	0.29	0
Chloromethane	0.56	0.56	0
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	1	1	0
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.43	0.44	2.3
Dichlorotetrafluoroethane	0	0	0
Ethyl Benzene	1.7	1.6	6.1
Heptane	1.6	1.6	0
Hexachloro-1,3-Butadiene	0	0	0
Hexane	2.4	2.3	4.3
m/p-Xylene	5.4	5.2	3.8
Methyl Methacrylate	0.57	0.57	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	0.91	0.96	5.3
Naphthalene	1.1	1	9.5
o-Xylene	2.1	2	4.9
Styrene	0	0.17	200 *
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0.25	0.24	4.1
Tetrachloroethene	0.05	0.05	0
Tetrahydrofuran	0	0	0
Toluene	11	10.9	0.91
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0
Trichlorofluoromethane	0.22	0.21	4.7
Vinyl Chloride	0	0	0