

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: Q2795

Project: Girard School - PA

Run No.: LB136743

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136743BSS							
Bromide	mg/Kg	200	203		102	1	90-110	08/07/2025
Chloride	mg/Kg	60	61.0		102	1	90-110	08/07/2025
Fluoride	mg/Kg	40	40.8		102	1	90-110	08/07/2025
Nitrite	mg/Kg	60	60.5		101	1	90-110	08/07/2025
Nitrate	mg/Kg	50	50.6		101	1	90-110	08/07/2025
Sulfate	mg/Kg	300	302		101	1	90-110	08/07/2025
Orthophosphate as P	mg/Kg	100	103		103	1	90-110	08/07/2025

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: Q2795

Project: Girard School - PA

Run No.: LB136775

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169184BS							
Hexavalent Chromium	mg/Kg	20	20.0		100	1	84-110	08/11/2025

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: Q2795

Project: Girard School - PA

Run No.: LB136784

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169198BS							
Ammonia as N	mg/Kg	50	50.4		101	1	90-110	08/12/2025