



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q2795

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB06	Mercury	0.076	+/-0.2	U	0.20	CV	08/11/2025	10:35	LB136768

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q2795

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB15	Mercury	0.076	+/-0.2	U	0.20	CV	08/11/2025	11:08	LB136768
CCB16	Mercury	0.076	+/-0.2	U	0.20	CV	08/11/2025	11:55	LB136768
CCB17	Mercury	0.076	+/-0.2	U	0.20	CV	08/11/2025	12:29	LB136768
CCB18	Mercury	0.076	+/-0.2	U	0.20	CV	08/11/2025	13:34	LB136768
CCB19	Mercury	0.076	+/-0.2	U	0.20	CV	08/11/2025	13:59	LB136768

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q2795

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	08/11/2025	11:37	LB136778
	Antimony	6.76	+/-25	U	50.0	P	08/11/2025	11:37	LB136778
	Arsenic	5.12	+/-10	U	20.0	P	08/11/2025	11:37	LB136778
	Barium	14.6	+/-50	U	100	P	08/11/2025	11:37	LB136778
	Beryllium	0.56	+/-3	U	6.00	P	08/11/2025	11:37	LB136778
	Cadmium	0.50	+/-3	U	6.00	P	08/11/2025	11:37	LB136778
	Calcium	234	+/-1000	U	2000	P	08/11/2025	11:37	LB136778
	Chromium	2.12	+/-5	U	10.0	P	08/11/2025	11:37	LB136778
	Cobalt	2.26	+/-15	U	30.0	P	08/11/2025	11:37	LB136778
	Copper	4.60	+/-10	U	20.0	P	08/11/2025	11:37	LB136778
	Iron	23.4	+/-50	U	100	P	08/11/2025	11:37	LB136778
	Lead	2.30	+/-6	U	12.0	P	08/11/2025	11:37	LB136778
	Magnesium	244	+/-1000	U	2000	P	08/11/2025	11:37	LB136778
	Manganese	5.94	+/-10	U	20.0	P	08/11/2025	11:37	LB136778
	Nickel	3.06	+/-20	U	40.0	P	08/11/2025	11:37	LB136778
	Potassium	918	+/-1000	U	2000	P	08/11/2025	11:37	LB136778
	Selenium	9.64	+/-10	U	20.0	P	08/11/2025	11:37	LB136778
	Silver	1.62	+/-5	U	10.0	P	08/11/2025	11:37	LB136778
	Sodium	868	+/-1000	U	2000	P	08/11/2025	11:37	LB136778
	Thallium	4.38	+/-20	U	40.0	P	08/11/2025	11:37	LB136778
	Vanadium	6.26	+/-20	U	40.0	P	08/11/2025	11:37	LB136778
	Zinc	16.7	+/-20	U	40.0	P	08/11/2025	11:37	LB136778

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q2795

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	23.2	+/-50	J	100	P	08/11/2025	12:09	LB136778
	Antimony	6.76	+/-25	U	50.0	P	08/11/2025	12:09	LB136778
	Arsenic	5.12	+/-10	U	20.0	P	08/11/2025	12:09	LB136778
	Barium	14.6	+/-50	U	100	P	08/11/2025	12:09	LB136778
	Beryllium	0.56	+/-3	U	6.00	P	08/11/2025	12:09	LB136778
	Cadmium	0.90	+/-3	J	6.00	P	08/11/2025	12:09	LB136778
	Calcium	234	+/-1000	U	2000	P	08/11/2025	12:09	LB136778
	Chromium	2.12	+/-5	U	10.0	P	08/11/2025	12:09	LB136778
	Cobalt	2.26	+/-15	U	30.0	P	08/11/2025	12:09	LB136778
	Copper	4.60	+/-10	U	20.0	P	08/11/2025	12:09	LB136778
	Iron	23.4	+/-50	U	100	P	08/11/2025	12:09	LB136778
	Lead	2.66	+/-6	J	12.0	P	08/11/2025	12:09	LB136778
	Magnesium	244	+/-1000	U	2000	P	08/11/2025	12:09	LB136778
	Manganese	5.94	+/-10	U	20.0	P	08/11/2025	12:09	LB136778
	Nickel	3.06	+/-20	U	40.0	P	08/11/2025	12:09	LB136778
	Potassium	918	+/-1000	U	2000	P	08/11/2025	12:09	LB136778
	Selenium	9.64	+/-10	U	20.0	P	08/11/2025	12:09	LB136778
	Silver	1.62	+/-5	U	10.0	P	08/11/2025	12:09	LB136778
	Sodium	868	+/-1000	U	2000	P	08/11/2025	12:09	LB136778
	Thallium	4.38	+/-20	U	40.0	P	08/11/2025	12:09	LB136778
	Vanadium	6.26	+/-20	U	40.0	P	08/11/2025	12:09	LB136778
	Zinc	16.7	+/-20	U	40.0	P	08/11/2025	12:09	LB136778
CCB02	Aluminum	11.8	+/-50	J	100	P	08/11/2025	13:04	LB136778
	Antimony	6.76	+/-25	U	50.0	P	08/11/2025	13:04	LB136778
	Arsenic	5.12	+/-10	U	20.0	P	08/11/2025	13:04	LB136778
	Barium	14.6	+/-50	U	100	P	08/11/2025	13:04	LB136778
	Beryllium	0.56	+/-3	U	6.00	P	08/11/2025	13:04	LB136778
	Cadmium	0.89	+/-3	J	6.00	P	08/11/2025	13:04	LB136778
	Calcium	234	+/-1000	U	2000	P	08/11/2025	13:04	LB136778
	Chromium	2.12	+/-5	U	10.0	P	08/11/2025	13:04	LB136778
	Cobalt	2.26	+/-15	U	30.0	P	08/11/2025	13:04	LB136778
	Copper	4.60	+/-10	U	20.0	P	08/11/2025	13:04	LB136778
	Iron	27.7	+/-50	J	100	P	08/11/2025	13:04	LB136778
	Lead	2.35	+/-6	J	12.0	P	08/11/2025	13:04	LB136778
	Magnesium	244	+/-1000	U	2000	P	08/11/2025	13:04	LB136778
	Manganese	5.94	+/-10	U	20.0	P	08/11/2025	13:04	LB136778
	Nickel	3.06	+/-20	U	40.0	P	08/11/2025	13:04	LB136778
	Potassium	918	+/-1000	U	2000	P	08/11/2025	13:04	LB136778
	Selenium	9.64	+/-10	U	20.0	P	08/11/2025	13:04	LB136778

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q2795

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	08/11/2025	13:04	LB136778
	Sodium	901	+/-1000	J	2000	P	08/11/2025	13:04	LB136778
	Thallium	4.38	+/-20	U	40.0	P	08/11/2025	13:04	LB136778
	Vanadium	6.26	+/-20	U	40.0	P	08/11/2025	13:04	LB136778
	Zinc	16.7	+/-20	U	40.0	P	08/11/2025	13:04	LB136778
CCB03	Aluminum	11.3	+/-50	U	100	P	08/11/2025	14:58	LB136778
	Antimony	6.76	+/-25	U	50.0	P	08/11/2025	14:58	LB136778
	Arsenic	5.12	+/-10	U	20.0	P	08/11/2025	14:58	LB136778
	Barium	14.6	+/-50	U	100	P	08/11/2025	14:58	LB136778
	Beryllium	0.56	+/-3	U	6.00	P	08/11/2025	14:58	LB136778
	Cadmium	0.50	+/-3	U	6.00	P	08/11/2025	14:58	LB136778
	Calcium	234	+/-1000	U	2000	P	08/11/2025	14:58	LB136778
	Chromium	2.12	+/-5	U	10.0	P	08/11/2025	14:58	LB136778
	Cobalt	2.26	+/-15	U	30.0	P	08/11/2025	14:58	LB136778
	Copper	4.60	+/-10	U	20.0	P	08/11/2025	14:58	LB136778
	Iron	23.4	+/-50	U	100	P	08/11/2025	14:58	LB136778
	Lead	2.30	+/-6	U	12.0	P	08/11/2025	14:58	LB136778
	Magnesium	244	+/-1000	U	2000	P	08/11/2025	14:58	LB136778
	Manganese	5.94	+/-10	U	20.0	P	08/11/2025	14:58	LB136778
	Nickel	3.06	+/-20	U	40.0	P	08/11/2025	14:58	LB136778
	Potassium	918	+/-1000	U	2000	P	08/11/2025	14:58	LB136778
	Selenium	9.64	+/-10	U	20.0	P	08/11/2025	14:58	LB136778
	Silver	1.62	+/-5	U	10.0	P	08/11/2025	14:58	LB136778
	Sodium	868	+/-1000	U	2000	P	08/11/2025	14:58	LB136778
	Thallium	4.38	+/-20	U	40.0	P	08/11/2025	14:58	LB136778
	Vanadium	6.26	+/-20	U	40.0	P	08/11/2025	14:58	LB136778
	Zinc	16.7	+/-20	U	40.0	P	08/11/2025	14:58	LB136778
CCB04	Aluminum	11.3	+/-50	U	100	P	08/11/2025	16:06	LB136778
	Antimony	6.76	+/-25	U	50.0	P	08/11/2025	16:06	LB136778
	Arsenic	5.12	+/-10	U	20.0	P	08/11/2025	16:06	LB136778
	Barium	14.6	+/-50	U	100	P	08/11/2025	16:06	LB136778
	Beryllium	0.56	+/-3	U	6.00	P	08/11/2025	16:06	LB136778
	Cadmium	0.50	+/-3	U	6.00	P	08/11/2025	16:06	LB136778
	Calcium	234	+/-1000	U	2000	P	08/11/2025	16:06	LB136778
	Chromium	2.12	+/-5	U	10.0	P	08/11/2025	16:06	LB136778
	Cobalt	2.26	+/-15	U	30.0	P	08/11/2025	16:06	LB136778
	Copper	4.60	+/-10	U	20.0	P	08/11/2025	16:06	LB136778
	Iron	23.4	+/-50	U	100	P	08/11/2025	16:06	LB136778
	Lead	2.30	+/-6	U	12.0	P	08/11/2025	16:06	LB136778

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q2795

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	08/11/2025	16:06	LB136778
	Manganese	5.94	+/-10	U	20.0	P	08/11/2025	16:06	LB136778
	Nickel	3.06	+/-20	U	40.0	P	08/11/2025	16:06	LB136778
	Potassium	918	+/-1000	U	2000	P	08/11/2025	16:06	LB136778
	Selenium	9.64	+/-10	U	20.0	P	08/11/2025	16:06	LB136778
	Silver	1.62	+/-5	U	10.0	P	08/11/2025	16:06	LB136778
	Sodium	868	+/-1000	U	2000	P	08/11/2025	16:06	LB136778
	Thallium	4.38	+/-20	U	40.0	P	08/11/2025	16:06	LB136778
	Vanadium	6.26	+/-20	U	40.0	P	08/11/2025	16:06	LB136778
	Zinc	16.7	+/-20	U	40.0	P	08/11/2025	16:06	LB136778

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q2795

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	08/15/2025	12:00	LB136851
	Antimony	6.76	+/-25	U	50.0	P	08/15/2025	12:00	LB136851
	Arsenic	5.12	+/-10	U	20.0	P	08/15/2025	12:00	LB136851
	Barium	14.6	+/-50	U	100	P	08/15/2025	12:00	LB136851
	Beryllium	0.56	+/-3	U	6.00	P	08/15/2025	12:00	LB136851
	Cadmium	0.50	+/-3	U	6.00	P	08/15/2025	12:00	LB136851
	Calcium	234	+/-1000	U	2000	P	08/15/2025	12:00	LB136851
	Chromium	2.12	+/-5	U	10.0	P	08/15/2025	12:00	LB136851
	Cobalt	2.26	+/-15	U	30.0	P	08/15/2025	12:00	LB136851
	Copper	4.60	+/-10	U	20.0	P	08/15/2025	12:00	LB136851
	Iron	23.4	+/-50	U	100	P	08/15/2025	12:00	LB136851
	Lead	2.30	+/-6	U	12.0	P	08/15/2025	12:00	LB136851
	Magnesium	244	+/-1000	U	2000	P	08/15/2025	12:00	LB136851
	Manganese	5.94	+/-10	U	20.0	P	08/15/2025	12:00	LB136851
	Nickel	3.06	+/-20	U	40.0	P	08/15/2025	12:00	LB136851
	Potassium	918	+/-1000	U	2000	P	08/15/2025	12:00	LB136851
	Selenium	9.64	+/-10	U	20.0	P	08/15/2025	12:00	LB136851
	Silver	1.62	+/-5	U	10.0	P	08/15/2025	12:00	LB136851
	Sodium	868	+/-1000	U	2000	P	08/15/2025	12:00	LB136851
	Thallium	4.38	+/-20	U	40.0	P	08/15/2025	12:00	LB136851
	Vanadium	6.26	+/-20	U	40.0	P	08/15/2025	12:00	LB136851
	Zinc	16.7	+/-20	U	40.0	P	08/15/2025	12:00	LB136851

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q2795

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	08/15/2025	12:30	LB136851
	Antimony	6.76	+/-25	U	50.0	P	08/15/2025	12:30	LB136851
	Arsenic	5.12	+/-10	U	20.0	P	08/15/2025	12:30	LB136851
	Barium	14.6	+/-50	U	100	P	08/15/2025	12:30	LB136851
	Beryllium	0.56	+/-3	U	6.00	P	08/15/2025	12:30	LB136851
	Cadmium	0.50	+/-3	U	6.00	P	08/15/2025	12:30	LB136851
	Calcium	234	+/-1000	U	2000	P	08/15/2025	12:30	LB136851
	Chromium	2.12	+/-5	U	10.0	P	08/15/2025	12:30	LB136851
	Cobalt	2.26	+/-15	U	30.0	P	08/15/2025	12:30	LB136851
	Copper	4.60	+/-10	U	20.0	P	08/15/2025	12:30	LB136851
	Iron	23.4	+/-50	U	100	P	08/15/2025	12:30	LB136851
	Lead	2.30	+/-6	U	12.0	P	08/15/2025	12:30	LB136851
	Magnesium	244	+/-1000	U	2000	P	08/15/2025	12:30	LB136851
	Manganese	5.94	+/-10	U	20.0	P	08/15/2025	12:30	LB136851
	Nickel	3.06	+/-20	U	40.0	P	08/15/2025	12:30	LB136851
	Potassium	918	+/-1000	U	2000	P	08/15/2025	12:30	LB136851
	Selenium	9.64	+/-10	U	20.0	P	08/15/2025	12:30	LB136851
	Silver	1.62	+/-5	U	10.0	P	08/15/2025	12:30	LB136851
	Sodium	868	+/-1000	U	2000	P	08/15/2025	12:30	LB136851
	Thallium	4.38	+/-20	U	40.0	P	08/15/2025	12:30	LB136851
	Vanadium	6.26	+/-20	U	40.0	P	08/15/2025	12:30	LB136851
	Zinc	16.7	+/-20	U	40.0	P	08/15/2025	12:30	LB136851
CCB02	Aluminum	11.3	+/-50	U	100	P	08/15/2025	13:45	LB136851
	Antimony	6.76	+/-25	U	50.0	P	08/15/2025	13:45	LB136851
	Arsenic	5.12	+/-10	U	20.0	P	08/15/2025	13:45	LB136851
	Barium	14.6	+/-50	U	100	P	08/15/2025	13:45	LB136851
	Beryllium	0.56	+/-3	U	6.00	P	08/15/2025	13:45	LB136851
	Cadmium	0.50	+/-3	U	6.00	P	08/15/2025	13:45	LB136851
	Calcium	234	+/-1000	U	2000	P	08/15/2025	13:45	LB136851
	Chromium	2.12	+/-5	U	10.0	P	08/15/2025	13:45	LB136851
	Cobalt	2.26	+/-15	U	30.0	P	08/15/2025	13:45	LB136851
	Copper	4.60	+/-10	U	20.0	P	08/15/2025	13:45	LB136851
	Iron	23.4	+/-50	U	100	P	08/15/2025	13:45	LB136851
	Lead	2.30	+/-6	U	12.0	P	08/15/2025	13:45	LB136851
	Magnesium	244	+/-1000	U	2000	P	08/15/2025	13:45	LB136851
	Manganese	5.94	+/-10	U	20.0	P	08/15/2025	13:45	LB136851
	Nickel	3.06	+/-20	U	40.0	P	08/15/2025	13:45	LB136851
	Potassium	918	+/-1000	U	2000	P	08/15/2025	13:45	LB136851
	Selenium	9.64	+/-10	U	20.0	P	08/15/2025	13:45	LB136851

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q2795

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	08/15/2025	13:45	LB136851
	Sodium	868	+/-1000	U	2000	P	08/15/2025	13:45	LB136851
	Thallium	4.38	+/-20	U	40.0	P	08/15/2025	13:45	LB136851
	Vanadium	6.26	+/-20	U	40.0	P	08/15/2025	13:45	LB136851
	Zinc	16.7	+/-20	U	40.0	P	08/15/2025	13:45	LB136851
CCB03	Aluminum	11.3	+/-50	U	100	P	08/15/2025	15:10	LB136851
	Antimony	6.76	+/-25	U	50.0	P	08/15/2025	15:10	LB136851
	Arsenic	5.12	+/-10	U	20.0	P	08/15/2025	15:10	LB136851
	Barium	14.6	+/-50	U	100	P	08/15/2025	15:10	LB136851
	Beryllium	0.56	+/-3	U	6.00	P	08/15/2025	15:10	LB136851
	Cadmium	0.50	+/-3	U	6.00	P	08/15/2025	15:10	LB136851
	Calcium	234	+/-1000	U	2000	P	08/15/2025	15:10	LB136851
	Chromium	2.12	+/-5	U	10.0	P	08/15/2025	15:10	LB136851
	Cobalt	2.26	+/-15	U	30.0	P	08/15/2025	15:10	LB136851
	Copper	4.60	+/-10	U	20.0	P	08/15/2025	15:10	LB136851
	Iron	23.4	+/-50	U	100	P	08/15/2025	15:10	LB136851
	Lead	2.30	+/-6	U	12.0	P	08/15/2025	15:10	LB136851
	Magnesium	244	+/-1000	U	2000	P	08/15/2025	15:10	LB136851
	Manganese	5.94	+/-10	U	20.0	P	08/15/2025	15:10	LB136851
	Nickel	3.06	+/-20	U	40.0	P	08/15/2025	15:10	LB136851
	Potassium	918	+/-1000	U	2000	P	08/15/2025	15:10	LB136851
	Selenium	9.64	+/-10	U	20.0	P	08/15/2025	15:10	LB136851
	Silver	1.62	+/-5	U	10.0	P	08/15/2025	15:10	LB136851
	Sodium	868	+/-1000	U	2000	P	08/15/2025	15:10	LB136851
	Thallium	4.38	+/-20	U	40.0	P	08/15/2025	15:10	LB136851
	Vanadium	6.26	+/-20	U	40.0	P	08/15/2025	15:10	LB136851
	Zinc	16.7	+/-20	U	40.0	P	08/15/2025	15:10	LB136851
CCB04	Aluminum	11.3	+/-50	U	100	P	08/15/2025	16:01	LB136851
	Antimony	6.76	+/-25	U	50.0	P	08/15/2025	16:01	LB136851
	Arsenic	5.12	+/-10	U	20.0	P	08/15/2025	16:01	LB136851
	Barium	14.6	+/-50	U	100	P	08/15/2025	16:01	LB136851
	Beryllium	0.56	+/-3	U	6.00	P	08/15/2025	16:01	LB136851
	Cadmium	0.50	+/-3	U	6.00	P	08/15/2025	16:01	LB136851
	Calcium	234	+/-1000	U	2000	P	08/15/2025	16:01	LB136851
	Chromium	2.12	+/-5	U	10.0	P	08/15/2025	16:01	LB136851
	Cobalt	2.26	+/-15	U	30.0	P	08/15/2025	16:01	LB136851
	Copper	4.60	+/-10	U	20.0	P	08/15/2025	16:01	LB136851
	Iron	23.4	+/-50	U	100	P	08/15/2025	16:01	LB136851
	Lead	2.30	+/-6	U	12.0	P	08/15/2025	16:01	LB136851

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q2795

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	08/15/2025	16:01	LB136851
	Manganese	5.94	+/-10	U	20.0	P	08/15/2025	16:01	LB136851
	Nickel	3.06	+/-20	U	40.0	P	08/15/2025	16:01	LB136851
	Potassium	918	+/-1000	U	2000	P	08/15/2025	16:01	LB136851
	Selenium	9.64	+/-10	U	20.0	P	08/15/2025	16:01	LB136851
	Silver	1.62	+/-5	U	10.0	P	08/15/2025	16:01	LB136851
	Sodium	868	+/-1000	U	2000	P	08/15/2025	16:01	LB136851
	Thallium	4.38	+/-20	U	40.0	P	08/15/2025	16:01	LB136851
	Vanadium	6.26	+/-20	U	40.0	P	08/15/2025	16:01	LB136851
	Zinc	16.7	+/-20	U	40.0	P	08/15/2025	16:01	LB136851
CCB05	Aluminum	21.2	+/-50	J	100	P	08/15/2025	16:45	LB136851
	Antimony	6.76	+/-25	U	50.0	P	08/15/2025	16:45	LB136851
	Arsenic	5.12	+/-10	U	20.0	P	08/15/2025	16:45	LB136851
	Barium	14.6	+/-50	U	100	P	08/15/2025	16:45	LB136851
	Beryllium	0.56	+/-3	U	6.00	P	08/15/2025	16:45	LB136851
	Cadmium	0.50	+/-3	U	6.00	P	08/15/2025	16:45	LB136851
	Calcium	234	+/-1000	U	2000	P	08/15/2025	16:45	LB136851
	Chromium	2.12	+/-5	U	10.0	P	08/15/2025	16:45	LB136851
	Cobalt	2.26	+/-15	U	30.0	P	08/15/2025	16:45	LB136851
	Copper	4.60	+/-10	U	20.0	P	08/15/2025	16:45	LB136851
	Iron	23.4	+/-50	U	100	P	08/15/2025	16:45	LB136851
	Lead	2.30	+/-6	U	12.0	P	08/15/2025	16:45	LB136851
	Magnesium	244	+/-1000	U	2000	P	08/15/2025	16:45	LB136851
	Manganese	5.94	+/-10	U	20.0	P	08/15/2025	16:45	LB136851
	Nickel	3.06	+/-20	U	40.0	P	08/15/2025	16:45	LB136851
	Potassium	918	+/-1000	U	2000	P	08/15/2025	16:45	LB136851
	Selenium	9.64	+/-10	U	20.0	P	08/15/2025	16:45	LB136851
	Silver	1.62	+/-5	U	10.0	P	08/15/2025	16:45	LB136851
	Sodium	868	+/-1000	U	2000	P	08/15/2025	16:45	LB136851
	Thallium	4.38	+/-20	U	40.0	P	08/15/2025	16:45	LB136851
	Vanadium	6.26	+/-20	U	40.0	P	08/15/2025	16:45	LB136851
	Zinc	16.7	+/-20	U	40.0	P	08/15/2025	16:45	LB136851