

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q2795

Continuing Calib Date: 08/08/2025

Initial Calibration Date(s): 08/01/2025

08/01/2025

Continuing Calib Time: 10:26

Initial Calibration Time(s): 12:05

20:28

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.81	5.81	5.71	5.91	0.01
Aroclor-1016-2	(2)	5.83	5.83	5.73	5.93	0.00
Aroclor-1016-3	(3)	5.89	5.90	5.80	6.00	0.01
Aroclor-1016-4	(4)	5.99	5.99	5.89	6.09	0.00
Aroclor-1016-5	(5)	6.28	6.29	6.19	6.39	0.01
Aroclor-1260-1	(1)	7.40	7.40	7.30	7.50	0.00
Aroclor-1260-2	(2)	7.65	7.66	7.56	7.76	0.01
Aroclor-1260-3	(3)	8.01	8.01	7.91	8.11	0.00
Aroclor-1260-4	(4)	8.24	8.24	8.14	8.34	0.01
Aroclor-1260-5	(5)	8.56	8.57	8.47	8.67	0.01
Tetrachloro-m-xylene		4.65	4.66	4.56	4.76	0.01
Decachlorobiphenyl		10.43	10.44	10.34	10.54	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q2795

Continuing Calib Date: 08/08/2025

Initial Calibration Date(s): 08/01/2025

08/01/2025

Continuing Calib Time: 10:26

Initial Calibration Time(s): 12:05

20:28

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.90	4.91	4.81	5.01	0.01
Aroclor-1016-2	(2)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-3	(3)	5.08	5.08	4.98	5.18	0.00
Aroclor-1016-4	(4)	5.12	5.12	5.02	5.22	0.00
Aroclor-1016-5	(5)	5.33	5.34	5.24	5.44	0.01
Aroclor-1260-1	(1)	6.55	6.56	6.46	6.66	0.01
Aroclor-1260-2	(2)	6.70	6.71	6.61	6.81	0.01
Aroclor-1260-3	(3)	6.91	6.92	6.82	7.02	0.01
Aroclor-1260-4	(4)	7.17	7.18	7.08	7.28	0.01
Aroclor-1260-5	(5)	7.41	7.42	7.32	7.52	0.01
Tetrachloro-m-xylene		3.80	3.81	3.71	3.91	0.01
Decachlorobiphenyl		8.82	8.83	8.73	8.93	0.01