

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2795
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074282.D **Time Analyzed:** 10:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.805	5.712	5.912	488.460	500.000	-2.3
Aroclor-1016-2	5.826	5.733	5.933	494.150	500.000	-1.2
Aroclor-1016-3	5.889	5.796	5.996	491.650	500.000	-1.7
Aroclor-1016-4	5.986	5.893	6.093	493.790	500.000	-1.2
Aroclor-1016-5	6.278	6.186	6.386	490.000	500.000	-2.0
Aroclor-1260-1	7.396	7.303	7.503	467.880	500.000	-6.4
Aroclor-1260-2	7.649	7.555	7.755	452.900	500.000	-9.4
Aroclor-1260-3	8.007	7.913	8.113	457.240	500.000	-8.6
Aroclor-1260-4	8.235	8.142	8.342	457.850	500.000	-8.4
Aroclor-1260-5	8.560	8.468	8.668	449.390	500.000	-10.1
Decachlorobiphenyl	10.429	10.338	10.538	46.480	50.000	-7.0
Tetrachloro-m-xylene	4.653	4.560	4.760	51.030	50.000	2.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2795
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074282.D **Time Analyzed:** 10:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.899	4.806	5.006	536.440	500.000	7.3
Aroclor-1016-2	4.956	4.863	5.063	537.320	500.000	7.5
Aroclor-1016-3	5.077	4.983	5.183	519.520	500.000	3.9
Aroclor-1016-4	5.117	5.023	5.223	538.260	500.000	7.7
Aroclor-1016-5	5.333	5.239	5.439	517.620	500.000	3.5
Aroclor-1260-1	6.549	6.455	6.655	526.220	500.000	5.2
Aroclor-1260-2	6.704	6.609	6.809	481.820	500.000	-3.6
Aroclor-1260-3	6.913	6.819	7.019	506.360	500.000	1.3
Aroclor-1260-4	7.174	7.079	7.279	494.450	500.000	-1.1
Aroclor-1260-5	7.412	7.318	7.518	487.110	500.000	-2.6
Decachlorobiphenyl	8.818	8.726	8.926	46.210	50.000	-7.6
Tetrachloro-m-xylene	3.799	3.705	3.905	55.090	50.000	10.2