

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2795
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074297.D **Time Analyzed:** 15:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.808	5.712	5.912	497.830	500.000	-0.4
Aroclor-1016-2	5.830	5.733	5.933	492.990	500.000	-1.4
Aroclor-1016-3	5.893	5.796	5.996	503.740	500.000	0.7
Aroclor-1016-4	5.990	5.893	6.093	505.980	500.000	1.2
Aroclor-1016-5	6.283	6.186	6.386	496.650	500.000	-0.7
Aroclor-1260-1	7.400	7.303	7.503	487.930	500.000	-2.4
Aroclor-1260-2	7.652	7.555	7.755	466.300	500.000	-6.7
Aroclor-1260-3	8.010	7.913	8.113	470.970	500.000	-5.8
Aroclor-1260-4	8.239	8.142	8.342	471.930	500.000	-5.6
Aroclor-1260-5	8.564	8.468	8.668	456.120	500.000	-8.8
Decachlorobiphenyl	10.433	10.338	10.538	47.660	50.000	-4.7
Tetrachloro-m-xylene	4.657	4.560	4.760	51.350	50.000	2.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2795
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074297.D **Time Analyzed:** 15:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.902	4.806	5.006	550.770	500.000	10.2
Aroclor-1016-2	4.959	4.863	5.063	557.310	500.000	11.5
Aroclor-1016-3	5.079	4.983	5.183	561.910	500.000	12.4
Aroclor-1016-4	5.121	5.023	5.223	532.110	500.000	6.4
Aroclor-1016-5	5.335	5.239	5.439	533.450	500.000	6.7
Aroclor-1260-1	6.551	6.455	6.655	556.370	500.000	11.3
Aroclor-1260-2	6.705	6.609	6.809	516.930	500.000	3.4
Aroclor-1260-3	6.916	6.819	7.019	539.130	500.000	7.8
Aroclor-1260-4	7.175	7.079	7.279	534.080	500.000	6.8
Aroclor-1260-5	7.414	7.318	7.518	535.550	500.000	7.1
Decachlorobiphenyl	8.820	8.726	8.926	48.450	50.000	-3.1
Tetrachloro-m-xylene	3.802	3.705	3.905	55.980	50.000	12.0