



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2795</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Kleinfelder</u>	Datafile :	<u>PP074288.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169172BS (Column 1)	AR1016	166.5	173	ug/kg	104				71	120	
	AR1260	166.5	155	ug/kg	93				65	130	
PB169172BS (Column 2)	AR1016	166.5	175	ug/kg	105				71	120	
	AR1260	166.5	166	ug/kg	100				65	130	