

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2795
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/08/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089810.D **Time Analyzed:** 13:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.701	6.604	6.804	54.010	50.000	8.0
4,4'-DDE	6.192	6.094	6.294	51.610	50.000	3.2
4,4'-DDT	7.017	6.919	7.119	47.220	50.000	-5.6
Aldrin	5.267	5.169	5.369	53.980	50.000	8.0
Decachlorobiphenyl	9.067	8.971	9.171	46.030	50.000	-7.9
Dieldrin	6.343	6.245	6.445	51.730	50.000	3.5
Tetrachloro-m-xylene	3.547	3.448	3.648	53.480	50.000	7.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2795
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/08/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089810.D **Time Analyzed:** 13:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.924	5.825	6.025	59.510	50.000	19.0
4,4'-DDE	5.369	5.271	5.471	58.910	50.000	17.8
4,4'-DDT	6.178	6.079	6.279	56.780	50.000	13.6
Aldrin	4.364	4.265	4.465	59.050	50.000	18.1
Decachlorobiphenyl	8.065	7.968	8.168	48.520	50.000	-3.0
Dieldrin	5.507	5.409	5.609	58.500	50.000	17.0
Tetrachloro-m-xylene	2.877	2.778	2.978	58.790	50.000	17.6