

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2795
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/08/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089821.D **Time Analyzed:** 17:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.701	6.604	6.804	51.660	50.000	3.3
4,4'-DDE	6.192	6.094	6.294	48.680	50.000	-2.6
4,4'-DDT	7.017	6.919	7.119	42.180	50.000	-15.6
Aldrin	5.267	5.169	5.369	51.350	50.000	2.7
Decachlorobiphenyl	9.066	8.971	9.171	44.620	50.000	-10.8
Dieldrin	6.343	6.245	6.445	48.960	50.000	-2.1
Tetrachloro-m-xylene	3.547	3.448	3.648	51.380	50.000	2.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2795
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/08/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089821.D **Time Analyzed:** 17:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.924	5.825	6.025	57.790	50.000	15.6
4,4'-DDE	5.370	5.271	5.471	56.640	50.000	13.3
4,4'-DDT	6.178	6.079	6.279	51.020	50.000	2.0
Aldrin	4.364	4.265	4.465	57.180	50.000	14.4
Decachlorobiphenyl	8.066	7.968	8.168	47.200	50.000	-5.6
Dieldrin	5.508	5.409	5.609	56.430	50.000	12.9
Tetrachloro-m-xylene	2.878	2.778	2.978	56.850	50.000	13.7