

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2795
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/08/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089829.D **Time Analyzed:** 19:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.702	6.604	6.804	50.340	50.000	0.7
4,4'-DDE	6.192	6.094	6.294	47.750	50.000	-4.5
4,4'-DDT	7.016	6.919	7.119	40.290	50.000	-19.4
Aldrin	5.267	5.169	5.369	50.440	50.000	0.9
Decachlorobiphenyl	9.067	8.971	9.171	44.280	50.000	-11.4
Dieldrin	6.343	6.245	6.445	47.040	50.000	-5.9
Tetrachloro-m-xylene	3.547	3.448	3.648	50.490	50.000	1.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q2795
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/08/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089829.D **Time Analyzed:** 19:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.923	5.825	6.025	57.320	50.000	14.6
4,4'-DDE	5.370	5.271	5.471	56.120	50.000	12.2
4,4'-DDT	6.178	6.079	6.279	48.390	50.000	-3.2
Aldrin	4.365	4.265	4.465	56.940	50.000	13.9
Decachlorobiphenyl	8.066	7.968	8.168	45.750	50.000	-8.5
Dieldrin	5.507	5.409	5.609	55.510	50.000	11.0
Tetrachloro-m-xylene	2.878	2.778	2.978	57.130	50.000	14.3