

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2795</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Kleinfelder</u>	Datafile :	<u>PD089812.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB169173BS (Column 1)	Aldrin	16.66	16.0	ug/kg	96				82	124		
	Dieldrin	16.66	15.5	ug/kg	93				85	121		
	4,4'-DDE	16.66	15.6	ug/kg	94				81	123		
	4,4'-DDD	16.66	15.9	ug/kg	95				80	131		
	4,4'-DDT	16.66	13.8	ug/kg	83				70	129		
PB169173BS (Column 2)	Aldrin	16.66	17.7	ug/kg	106				82	124		
	Dieldrin	16.66	17.6	ug/kg	106				85	121		
	4,4'-DDE	16.66	17.7	ug/kg	106				81	123		
	4,4'-DDD	16.66	18.5	ug/kg	111				80	131		
	4,4'-DDT	16.66	16.4	ug/kg	98				70	129		